



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	CDM Smith				SDG No.:	Q1956				
Contract:	CAMP02	Lab Code:	CHEM		Case No.:	Q1956	SAS No.:	Q1956		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB01	Mercury	0.20	+/-0.20	U	0.20	CV	05/07/2025	09:25	LB135690	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	CDM Smith				SDG No.:	Q1956			
Contract:	CAMP02	Lab Code:	CHEM		Case No.:	Q1956		SAS No.:	Q1956
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB86	Mercury	0.20	+/-0.20	U	0.20	CV	05/07/2025	09:29	LB135690
CCB87	Mercury	0.20	+/-0.20	U	0.20	CV	05/07/2025	10:02	LB135690
CCB88	Mercury	0.20	+/-0.20	U	0.20	CV	05/07/2025	10:25	LB135690

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith **SDG No.:** Q1956
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1956 **SAS No.:** Q1956

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/07/2025	13:09	LB135699
	Barium	100	+/-100	U	100	P	05/07/2025	13:09	LB135699
	Cadmium	6.00	+/-6.00	U	6.00	P	05/07/2025	13:09	LB135699
	Chromium	10.0	+/-10.0	U	10.0	P	05/07/2025	13:09	LB135699
	Lead	12.0	+/-12.0	U	12.0	P	05/07/2025	13:09	LB135699
	Selenium	20.0	+/-20.0	U	20.0	P	05/07/2025	13:09	LB135699
	Silver	10.0	+/-10.0	U	10.0	P	05/07/2025	13:09	LB135699

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1956							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1956	SAS No.: Q1956						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/07/2025	13:43	LB135699
	Barium	100	+/-100	U	100	P	05/07/2025	13:43	LB135699
	Cadmium	6.00	+/-6.00	U	6.00	P	05/07/2025	13:43	LB135699
	Chromium	10.0	+/-10.0	U	10.0	P	05/07/2025	13:43	LB135699
	Lead	12.0	+/-12.0	U	12.0	P	05/07/2025	13:43	LB135699
	Selenium	20.0	+/-20.0	U	20.0	P	05/07/2025	13:43	LB135699
	Silver	10.0	+/-10.0	U	10.0	P	05/07/2025	13:43	LB135699
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	05/07/2025	14:39	LB135699
	Barium	100	+/-100	U	100	P	05/07/2025	14:39	LB135699
	Cadmium	6.00	+/-6.00	U	6.00	P	05/07/2025	14:39	LB135699
	Chromium	10.0	+/-10.0	U	10.0	P	05/07/2025	14:39	LB135699
	Lead	12.0	+/-12.0	U	12.0	P	05/07/2025	14:39	LB135699
	Selenium	20.0	+/-20.0	U	20.0	P	05/07/2025	14:39	LB135699
	Silver	10.0	+/-10.0	U	10.0	P	05/07/2025	14:39	LB135699
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	05/07/2025	15:27	LB135699
	Barium	100	+/-100	U	100	P	05/07/2025	15:27	LB135699
	Cadmium	6.00	+/-6.00	U	6.00	P	05/07/2025	15:27	LB135699
	Chromium	10.0	+/-10.0	U	10.0	P	05/07/2025	15:27	LB135699
	Lead	12.0	+/-12.0	U	12.0	P	05/07/2025	15:27	LB135699
	Selenium	20.0	+/-20.0	U	20.0	P	05/07/2025	15:27	LB135699
	Silver	10.0	+/-10.0	U	10.0	P	05/07/2025	15:27	LB135699
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	05/07/2025	16:09	LB135699
	Barium	100	+/-100	U	100	P	05/07/2025	16:09	LB135699
	Cadmium	6.00	+/-6.00	U	6.00	P	05/07/2025	16:09	LB135699
	Chromium	10.0	+/-10.0	U	10.0	P	05/07/2025	16:09	LB135699
	Lead	12.0	+/-12.0	U	12.0	P	05/07/2025	16:09	LB135699
	Selenium	20.0	+/-20.0	U	20.0	P	05/07/2025	16:09	LB135699
	Silver	10.0	+/-10.0	U	10.0	P	05/07/2025	16:09	LB135699
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	05/07/2025	16:52	LB135699
	Barium	100	+/-100	U	100	P	05/07/2025	16:52	LB135699
	Cadmium	6.00	+/-6.00	U	6.00	P	05/07/2025	16:52	LB135699
	Chromium	10.0	+/-10.0	U	10.0	P	05/07/2025	16:52	LB135699
	Lead	12.0	+/-12.0	U	12.0	P	05/07/2025	16:52	LB135699
	Selenium	20.0	+/-20.0	U	20.0	P	05/07/2025	16:52	LB135699
	Silver	10.0	+/-10.0	U	10.0	P	05/07/2025	16:52	LB135699