

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q1956

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167851TB		WATER		Batch Number:	PB167883		Prep Date:	05/06/2025	
	Mercury	2.00	<2.00	U	2.00	CV	05/07/2025	10:13	LB135690
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167883BL		WATER		Batch Number:	PB167883		Prep Date:	05/06/2025	
	Mercury	0.20	<0.20	U	0.20	CV	05/07/2025	10:16	LB135690

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q1956

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167851TB		WATER		Batch Number:	PB167882		Prep Date:	05/06/2025	
	Arsenic	100	<100	U	100	P	05/07/2025	14:26	LB135699
	Barium	500	<500	U	500	P	05/07/2025	14:26	LB135699
	Cadmium	30.0	<30.0	U	30.0	P	05/07/2025	14:26	LB135699
	Chromium	50.0	<50.0	U	50.0	P	05/07/2025	14:26	LB135699
	Lead	60.0	<60.0	U	60.0	P	05/07/2025	14:26	LB135699
	Selenium	100	<100	U	100	P	05/07/2025	14:26	LB135699
	Silver	50.0	<50.0	U	50.0	P	05/07/2025	14:26	LB135699
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167882BL		WATER		Batch Number:	PB167882		Prep Date:	05/06/2025	
	Arsenic	100	<100	U	100	P	05/07/2025	13:47	LB135699
	Barium	500	<500	U	500	P	05/07/2025	13:47	LB135699
	Cadmium	30.0	<30.0	U	30.0	P	05/07/2025	13:47	LB135699
	Chromium	50.0	<50.0	U	50.0	P	05/07/2025	13:47	LB135699
	Lead	60.0	<60.0	U	60.0	P	05/07/2025	13:47	LB135699
	Selenium	100	<100	U	100	P	05/07/2025	13:47	LB135699
	Silver	50.0	<50.0	U	50.0	P	05/07/2025	13:47	LB135699