

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: CDM Smith **Level:** LOW **SDG No.:** Q1956
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1956 **SAS No.:** Q1956
Matrix: Water **Sample ID:** Q1960-03 **Client ID:** 72-11933DUP
Percent Solids for Sample: NA **Duplicate ID** Q1960-03DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	1320		1140		15		P
Cadmium	ug/L	20	5.41	J	4.13	J	27		P
Chromium	ug/L	20	50.0	U	50.0	U			P
Lead	ug/L	20	601		535		12		P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: CDM Smith **Level:** LOW **SDG No.:** Q1956
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1956 **SAS No.:** Q1956
Matrix: Water **Sample ID:** Q1960-03MS **Client ID:** 72-11933MSD
Percent Solids for Sample: NA **Duplicate ID** Q1960-03MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	3970		3820		4		P
Barium	ug/L	20	2160		2150		0		P
Cadmium	ug/L	20	998		957		4		P
Chromium	ug/L	20	1950		1910		2		P
Lead	ug/L	20	5290		5090		4		P
Mercury	ug/L	20	41.3		38.1		8		CV
Selenium	ug/L	20	9840		9430		4		P
Silver	ug/L	20	364		358		2		P