

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CAMP02
 ProjectID: Bergen Point Fueling System
 Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG No.: Q1956

Calibration Sequence : FF042225			Test : Diesel Range Organics	
Concentration	(PPM)	Area Count	Reference Factor	File ID
1000		116059922	116060	FF015786.D
500		58079559	116159	FF015787.D
200		21235975	106180	FF015788.D
100		11342548	113425	FF015789.D
50		7274526	145491	FF015790.D
AVG RF : 119463			% RSD : 12.646	AVG RT : 15.02