

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CAMP02
 ProjectID: Bergen Point Fueling System
 Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG No.: Q1956

Calibration Sequence : FG042425		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	122641169	122641	FG015756.D
500	64139521	128279	FG015757.D
200	25210755	126054	FG015758.D
100	13317775	133178	FG015759.D
50	6223650	124473	FG015760.D
AVG RF : 126925		% RSD : 3.202	AVG RT : 15.0012