

SOIL DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RI

Lab Name: Chemtech **Client:** CDM Smith
Lab Code: CHEM **Cas No:** Q1956 **SAS No :** Q1956 **SDG No:** Q1956
Matrix Spike - EPA Sample No : PB167975BS **Datafile:** FG015821.D

COMPOUND	SPIKE ADDED ug/kg	CONCENTRATION ug/kg	LCS/LCSD CONCENTRATION ug/kg	% REC	QC LIMITS
DRO	6662	0	6321	95	68-131