

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL01 Date Analyzed: 05/08/2025

Lab Sample No.: HSTDCCC750 Data File : PS030073.D Time Analyzed: 09:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.154	9.044	9.244	696.620	712.500	-2.2
2,4,5-TP(Silvex)	8.872	8.762	8.962	700.370	712.500	-1.7
2,4-D	8.030	7.919	8.119	680.540	705.000	-3.5
2,4-DB	9.713	9.603	9.803	673.260	712.500	-5.5
2,4-DCAA	6.954	6.842	7.042	710.520	750.000	-5.3
DICAMBA	7.130	7.018	7.218	684.650	705.000	-2.9
DICHLORPROP	7.811	7.699	7.899	674.330	705.000	-4.4
Dinoseb	10.874	10.766	10.966	683.770	705.000	-3.0

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL01 Date Analyzed: 05/08/2025

Lab Sample No.: HSTDCCC750 Data File : PS030073.D Time Analyzed: 09:28

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.934	9.847	10.047	695.970	712.500	-2.3
2,4,5-TP(Silvex)	9.530	9.442	9.642	713.980	712.500	0.2
2,4-D	8.651	8.561	8.761	674.960	705.000	-4.3
2,4-DB	10.493	10.407	10.607	678.850	712.500	-4.7
2,4-DCAA	7.459	7.367	7.567	723.850	750.000	-3.5
DICAMBA	7.645	7.553	7.753	704.770	705.000	0.0
DICHLORPROP	8.338	8.248	8.448	689.710	705.000	-2.2
Dinoseb	10.866	10.781	10.981	680.570	705.000	-3.5