

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL02 Date Analyzed: 05/08/2025

Lab Sample No.: HSTDCCC750 Data File : PS030081.D Time Analyzed: 12:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.137	9.044	9.244	746.170	712.500	4.7
2,4,5-TP(Silvex)	8.855	8.762	8.962	739.080	712.500	3.7
2,4-D	8.013	7.919	8.119	718.450	705.000	1.9
2,4-DB	9.694	9.603	9.803	744.710	712.500	4.5
2,4-DCAA	6.937	6.842	7.042	763.730	750.000	1.8
DICAMBA	7.113	7.018	7.218	730.480	705.000	3.6
DICHLORPROP	7.793	7.699	7.899	709.570	705.000	0.6
Dinoseb	10.857	10.766	10.966	679.450	705.000	-3.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL02 Date Analyzed: 05/08/2025

Lab Sample No.: HSTDCCC750 Data File : PS030081.D Time Analyzed: 12:45

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.935	9.847	10.047	763.550	712.500	7.2
2,4,5-TP(Silvex)	9.530	9.442	9.642	779.490	712.500	9.4
2,4-D	8.651	8.561	8.761	733.060	705.000	4.0
2,4-DB	10.493	10.407	10.607	731.240	712.500	2.6
2,4-DCAA	7.460	7.367	7.567	779.400	750.000	3.9
DICAMBA	7.646	7.553	7.753	772.630	705.000	9.6
DICHLORPROP	8.339	8.248	8.448	744.450	705.000	5.6
Dinoseb	10.867	10.781	10.981	739.590	705.000	4.9