

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL03 Date Analyzed: 05/08/2025

Lab Sample No.: HSTDCCC750 Data File : PS030093.D Time Analyzed: 18:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.144	9.044	9.244	755.420	712.500	6.0
2,4,5-TP(Silvex)	8.862	8.762	8.962	754.240	712.500	5.9
2,4-D	8.020	7.919	8.119	728.010	705.000	3.3
2,4-DB	9.702	9.603	9.803	749.540	712.500	5.2
2,4-DCAA	6.944	6.842	7.042	767.130	750.000	2.3
DICAMBA	7.120	7.018	7.218	737.180	705.000	4.6
DICHLORPROP	7.801	7.699	7.899	720.830	705.000	2.2
Dinoseb	10.864	10.766	10.966	730.240	705.000	3.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL03 Date Analyzed: 05/08/2025

Lab Sample No.: HSTDCCC750 Data File : PS030093.D Time Analyzed: 18:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.934	9.847	10.047	769.150	712.500	8.0
2,4,5-TP(Silvex)	9.530	9.442	9.642	785.490	712.500	10.2
2,4-D	8.651	8.561	8.761	745.890	705.000	5.8
2,4-DB	10.492	10.407	10.607	715.160	712.500	0.4
2,4-DCAA	7.459	7.367	7.567	782.710	750.000	4.4
DICAMBA	7.645	7.553	7.753	776.270	705.000	10.1
DICHLORPROP	8.338	8.248	8.448	750.880	705.000	6.5
Dinoseb	10.866	10.781	10.981	744.820	705.000	5.6