

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Continuing Calib Date: 05/06/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 09:55 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.66	5.67	5.57	5.77	0.01
Aroclor-1248-2 (2)	5.93	5.94	5.84	6.04	0.01
Aroclor-1248-3 (3)	6.14	6.14	6.04	6.24	0.00
Aroclor-1248-4 (4)	6.53	6.54	6.44	6.64	0.01
Aroclor-1248-5 (5)	6.57	6.58	6.48	6.68	0.01
Tetrachloro-m-xylene	4.51	4.51	4.41	4.61	0.00
Decachlorobiphenyl	10.22	10.24	10.14	10.34	0.02

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Continuing Calib Date: 05/06/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 09:55 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.89	4.90	4.80	5.00	0.01
Aroclor-1248-2	(2)	5.13	5.14	5.04	5.24	0.01
Aroclor-1248-3	(3)	5.17	5.18	5.08	5.28	0.01
Aroclor-1248-4	(4)	5.34	5.35	5.25	5.45	0.01
Aroclor-1248-5	(5)	5.73	5.74	5.64	5.84	0.01
Tetrachloro-m-xylene		3.81	3.81	3.71	3.91	0.00
Decachlorobiphenyl		8.84	8.85	8.75	8.95	0.01