

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL01 Date Analyzed: 05/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP071800.D Time Analyzed: 09:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.673	5.571	5.771	439.490	500.000	-12.1
Aroclor-1016-2	5.694	5.593	5.793	454.470	500.000	-9.1
Aroclor-1016-3	5.756	5.654	5.854	444.880	500.000	-11.0
Aroclor-1016-4	5.853	5.752	5.952	443.130	500.000	-11.4
Aroclor-1016-5	6.146	6.045	6.245	456.880	500.000	-8.6
Aroclor-1260-1	7.264	7.165	7.365	480.350	500.000	-3.9
Aroclor-1260-2	7.518	7.418	7.618	472.150	500.000	-5.6
Aroclor-1260-3	7.876	7.777	7.977	456.570	500.000	-8.7
Aroclor-1260-4	8.100	8.002	8.202	439.400	500.000	-12.1
Aroclor-1260-5	8.420	8.322	8.522	461.880	500.000	-7.6
Decachlorobiphenyl	10.236	10.137	10.337	47.570	50.000	-4.9
Tetrachloro-m-xylene	4.518	4.417	4.617	45.670	50.000	-8.7

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL01 Date Analyzed: 05/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP071800.D Time Analyzed: 09:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.893	4.798	4.998	477.790	500.000	-4.4
Aroclor-1016-2	4.912	4.816	5.016	488.710	500.000	-2.3
Aroclor-1016-3	5.089	4.993	5.193	489.730	500.000	-2.1
Aroclor-1016-4	5.131	5.036	5.236	480.630	500.000	-3.9
Aroclor-1016-5	5.345	5.250	5.450	535.400	500.000	7.1
Aroclor-1260-1	6.378	6.285	6.485	495.310	500.000	-0.9
Aroclor-1260-2	6.567	6.473	6.673	486.650	500.000	-2.7
Aroclor-1260-3	6.720	6.627	6.827	527.060	500.000	5.4
Aroclor-1260-4	7.191	7.098	7.298	498.400	500.000	-0.3
Aroclor-1260-5	7.432	7.339	7.539	484.340	500.000	-3.1
Decachlorobiphenyl	8.839	8.749	8.949	49.450	50.000	-1.1
Tetrachloro-m-xylene	3.808	3.712	3.912	48.720	50.000	-2.6