

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL03 Date Analyzed: 05/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP071835.D Time Analyzed: 10:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.657	5.571	5.771	447.310	500.000	-10.5
Aroclor-1016-2	5.680	5.593	5.793	491.530	500.000	-1.7
Aroclor-1016-3	5.743	5.654	5.854	486.440	500.000	-2.7
Aroclor-1016-4	5.840	5.752	5.952	490.090	500.000	-2.0
Aroclor-1016-5	6.133	6.045	6.245	477.650	500.000	-4.5
Aroclor-1260-1	7.252	7.165	7.365	428.970	500.000	-14.2
Aroclor-1260-2	7.505	7.418	7.618	436.270	500.000	-12.7
Aroclor-1260-3	7.864	7.777	7.977	467.010	500.000	-6.6
Aroclor-1260-4	8.088	8.002	8.202	429.700	500.000	-14.1
Aroclor-1260-5	8.407	8.322	8.522	442.650	500.000	-11.5
Decachlorobiphenyl	10.220	10.137	10.337	46.940	50.000	-6.1
Tetrachloro-m-xylene	4.506	4.417	4.617	48.110	50.000	-3.8

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL03 Date Analyzed: 05/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP071835.D Time Analyzed: 10:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.887	4.798	4.998	500.100	500.000	0.0
Aroclor-1016-2	4.906	4.816	5.016	524.070	500.000	4.8
Aroclor-1016-3	5.083	4.993	5.193	514.310	500.000	2.9
Aroclor-1016-4	5.124	5.036	5.236	524.310	500.000	4.9
Aroclor-1016-5	5.339	5.250	5.450	532.050	500.000	6.4
Aroclor-1260-1	6.372	6.285	6.485	514.800	500.000	3.0
Aroclor-1260-2	6.561	6.473	6.673	513.620	500.000	2.7
Aroclor-1260-3	6.713	6.627	6.827	507.560	500.000	1.5
Aroclor-1260-4	7.185	7.098	7.298	500.210	500.000	0.0
Aroclor-1260-5	7.426	7.339	7.539	490.350	500.000	-1.9
Decachlorobiphenyl	8.832	8.749	8.949	50.800	50.000	1.6
Tetrachloro-m-xylene	3.803	3.712	3.912	52.820	50.000	5.6