

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL04 Date Analyzed: 05/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP071847.D Time Analyzed: 17:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.664	5.571	5.771	490.910	500.000	-1.8
Aroclor-1016-2	5.686	5.593	5.793	486.990	500.000	-2.6
Aroclor-1016-3	5.748	5.654	5.854	484.260	500.000	-3.1
Aroclor-1016-4	5.845	5.752	5.952	490.690	500.000	-1.9
Aroclor-1016-5	6.138	6.045	6.245	482.090	500.000	-3.6
Aroclor-1260-1	7.257	7.165	7.365	470.440	500.000	-5.9
Aroclor-1260-2	7.510	7.418	7.618	478.000	500.000	-4.4
Aroclor-1260-3	7.869	7.777	7.977	472.760	500.000	-5.4
Aroclor-1260-4	8.093	8.002	8.202	436.510	500.000	-12.7
Aroclor-1260-5	8.413	8.322	8.522	460.610	500.000	-7.9
Decachlorobiphenyl	10.225	10.137	10.337	48.570	50.000	-2.9
Tetrachloro-m-xylene	4.510	4.417	4.617	48.870	50.000	-2.3

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL04 Date Analyzed: 05/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP071847.D Time Analyzed: 17:35

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.889	4.798	4.998	503.880	500.000	0.8
Aroclor-1016-2	4.908	4.816	5.016	533.120	500.000	6.6
Aroclor-1016-3	5.084	4.993	5.193	516.770	500.000	3.4
Aroclor-1016-4	5.126	5.036	5.236	522.060	500.000	4.4
Aroclor-1016-5	5.341	5.250	5.450	514.370	500.000	2.9
Aroclor-1260-1	6.374	6.285	6.485	506.580	500.000	1.3
Aroclor-1260-2	6.563	6.473	6.673	500.570	500.000	0.1
Aroclor-1260-3	6.716	6.627	6.827	510.520	500.000	2.1
Aroclor-1260-4	7.187	7.098	7.298	519.530	500.000	3.9
Aroclor-1260-5	7.429	7.339	7.539	511.410	500.000	2.3
Decachlorobiphenyl	8.834	8.749	8.949	51.640	50.000	3.3
Tetrachloro-m-xylene	3.805	3.712	3.912	52.680	50.000	5.4