

Cover Page

Order ID : Q1956

Project ID : Bergen Point Fueling System

Client : CDM Smith

Lab Sample Number

Q1956-01
Q1956-02
Q1956-03
Q1956-04
Q1956-05
Q1956-06
Q1956-07
Q1956-09

Client Sample Number

SB1-3-4
SB2-4-5
Q1956-02MS
Q1956-02MSD
COMP1
SB91-3-4
FB-05022025
TB

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 5/10/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following "Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1956

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: PRADIP PRAJAPATI

Date: 05/10/2025

LAB CHRONICLE

OrderID:	Q1956	OrderDate:	5/2/2025 3:14:35 PM
Client:	CDM Smith	Project:	Bergen Point Fueling System
Contact:	Marcie Ann Encinas	Location:	L31,VOA Ref. #2 Soil,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1956-01	SB1-3-4	SOIL			05/01/25			05/02/25
			Gasoline Range Organics	8015D			05/06/25	
			PCB	8082A		05/06/25	05/06/25	
Q1956-02	SB2-4-5	SOIL			05/02/25			05/02/25
			Gasoline Range Organics	8015D			05/06/25	
			PCB	8082A		05/06/25	05/06/25	
Q1956-06	SB91-3-4	SOIL			05/01/25			05/02/25
			Gasoline Range Organics	8015D			05/06/25	
			PCB	8082A		05/06/25	05/06/25	
Q1956-07	FB-05022025	Water			05/02/25			05/02/25
			Gasoline Range Organics	8015D			05/06/25	
			PCB	8082A		05/07/25	05/07/25	

Hit Summary Sheet
SW-846

SDG No.: Q1956

Order ID: Q1956

Client: CDM Smith

Project ID: Bergen Point Fueling System

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : Q1956-02	SB2-4-5 SB2-4-5	SOIL	Aroclor-1254	6.90	JP	3.40	18.1	ug/kg
Total Concentration:				6.900				



QC SUMMARY

Surrogate Summary

SDG No.: Q1956

Client: CDM Smith

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP071388.D	PIBLK-PP071388.D	Tetrachloro-m-xylene	1	20	17.1	85		60	140
		Decachlorobiphenyl	1	20	17.6	88		60	140
		Tetrachloro-m-xylene	2	20	17.2	86		60	140
		Decachlorobiphenyl	2	20	17.6	88		60	140
I.BLK-PP071804.D	PIBLK-PP071804.D	Tetrachloro-m-xylene	1	20	17.1	86		60	140
		Decachlorobiphenyl	1	20	19.0	95		60	140
		Tetrachloro-m-xylene	2	20	19.0	95		60	140
		Decachlorobiphenyl	2	20	20.2	101		60	140
PB167877BL	PB167877BL	Tetrachloro-m-xylene	1	20	20.6	103		32	144
		Decachlorobiphenyl	1	20	21.9	109		32	175
		Tetrachloro-m-xylene	2	20	22.9	115		32	144
		Decachlorobiphenyl	2	20	24.6	123		32	175
PB167877BS	PB167877BS	Tetrachloro-m-xylene	1	20	20.7	104		32	144
		Decachlorobiphenyl	1	20	22.2	111		32	175
		Tetrachloro-m-xylene	2	20	22.4	112		32	144
		Decachlorobiphenyl	2	20	24.1	121		32	175
Q1956-01	SB1-3-4	Tetrachloro-m-xylene	1	20	20.9	104		32	144
		Decachlorobiphenyl	1	20	16.1	81		32	175
		Tetrachloro-m-xylene	2	20	23.5	117		32	144
		Decachlorobiphenyl	2	20	18.3	91		32	175
Q1956-02	SB2-4-5	Tetrachloro-m-xylene	1	20	20.7	104		32	144
		Decachlorobiphenyl	1	20	17.2	86		32	175
		Tetrachloro-m-xylene	2	20	22.6	113		32	144
		Decachlorobiphenyl	2	20	18.0	90		32	175
Q1956-03MS	SB2-4-5MS	Tetrachloro-m-xylene	1	20	22.6	113		32	144
		Decachlorobiphenyl	1	20	21.1	105		32	175
		Tetrachloro-m-xylene	2	20	23.7	119		32	144
		Decachlorobiphenyl	2	20	23.3	117		32	175
Q1956-04MSD	SB2-4-5MSD	Tetrachloro-m-xylene	1	20	21.8	109		32	144
		Decachlorobiphenyl	1	20	21.0	105		32	175
		Tetrachloro-m-xylene	2	20	21.2	106		32	144
		Decachlorobiphenyl	2	20	22.7	114		32	175
Q1956-06	SB91-3-4	Tetrachloro-m-xylene	1	20	19.7	99		32	144
		Decachlorobiphenyl	1	20	14.9	74		32	175
		Tetrachloro-m-xylene	2	20	22.8	114		32	144
		Decachlorobiphenyl	2	20	16.2	81		32	175
I.BLK-PP071819.D	PIBLK-PP071819.D	Tetrachloro-m-xylene	1	20	17.0	85		60	140
		Decachlorobiphenyl	1	20	18.7	93		60	140
		Tetrachloro-m-xylene	2	20	19.6	98		60	140
		Decachlorobiphenyl	2	20	21.9	109		60	140
I.BLK-PP071839.D	PIBLK-PP071839.D	Tetrachloro-m-xylene	1	20	16.3	81		60	140

Surrogate Summary

SDG No.: Q1956

Client: CDM Smith

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP071839.D	PIBLK-PP071839.D	Decachlorobiphenyl	1	20	16.5	83		60	140
		Tetrachloro-m-xylene	2	20	18.1	91		60	140
PB167890BL	PB167890BL	Decachlorobiphenyl	2	20	19.2	96		60	140
		Tetrachloro-m-xylene	1	20	20.3	102		16	158
		Decachlorobiphenyl	1	20	20.3	101		10	173
		Tetrachloro-m-xylene	2	20	21.9	110		16	158
		Decachlorobiphenyl	2	20	23.4	117		10	173
		Tetrachloro-m-xylene	1	20	20.4	102		16	158
PB167890BS	PB167890BS	Decachlorobiphenyl	1	20	20.6	103		10	173
		Tetrachloro-m-xylene	2	20	20.5	102		16	158
		Decachlorobiphenyl	2	20	22.1	110		10	173
		Tetrachloro-m-xylene	1	20	19.7	99		16	158
PB167890BSD	PB167890BSD	Decachlorobiphenyl	1	20	19.8	99		10	173
		Tetrachloro-m-xylene	2	20	20.7	104		16	158
		Decachlorobiphenyl	2	20	22.1	111		10	173
		Tetrachloro-m-xylene	1	20	21.4	107		16	158
		Decachlorobiphenyl	1	20	12.3	62		10	173
Q1956-07	FB-05022025	Tetrachloro-m-xylene	2	20	23.5	117		16	158
		Decachlorobiphenyl	2	20	13.8	69		10	173
		Tetrachloro-m-xylene	1	20	16.2	81		60	140
		Decachlorobiphenyl	1	20	16.6	83		60	140
I.BLK-PP071851.D	PIBLK-PP071851.D	Tetrachloro-m-xylene	2	20	17.9	90		60	140
		Decachlorobiphenyl	2	20	19.1	96		60	140

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1956

Client: CDM Smith

Analytical Method: 8082A

DataFile : PP071811.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	RPD
			Result	Result			Qual	RPD	Qual	Low	High	
Client Sample ID: Q1956-03MS	SB2-4-5MS											
	AR1016	178.1	0	187	ug/kg	105				55	146	
	AR1260	178.1	0	182	ug/kg	102				31	146	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1956

Client: CDM Smith

Analytical Method: 8082A

DataFile : PP071812.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Client Sample ID: Q1956-04MSD	SB2-4-5MSD											
	AR1016	178.1	0	181	ug/kg	102		3		55	146	20
	AR1260	178.1	0	186	ug/kg	104		2		31	146	20



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1956

Client: CDM Smith

Analytical Method: **8082A** **Datafile :** PP071808.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167877BS	AR1016	166.6	152	ug/kg	91				71	120	
	AR1260	166.6	151	ug/kg	91				65	130	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1956

Client: CDM Smith

Analytical Method: **8082A** **Datafile :** PP071841.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167890BS	AR1016	5	4.60	ug/L	92				61	112	
	AR1260	5	4.70	ug/L	94				66	113	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1956

Client: CDM Smith

Analytical Method: **8082A** **Datafile :** PP071842.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167890BSD	AR1016	5	4.70	ug/L	94	2			61	112	20
	AR1260	5	4.70	ug/L	94	0			66	113	20

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167877BL

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956

SAS No.: Q1956 SDG NO.: Q1956

Lab Sample ID: PB167877BL

Lab File ID: PP071807.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 05/06/2025

Date Analyzed (1): 05/06/2025

Date Analyzed (2): 05/06/2025

Time Analyzed (1): 13:12

Time Analyzed (2): 13:12

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB167877BS	PB167877BS	PP071808.D	05/06/2025	05/06/2025
SB1-3-4	Q1956-01	PP071809.D	05/06/2025	05/06/2025
SB2-4-5	Q1956-02	PP071810.D	05/06/2025	05/06/2025
SB2-4-5MS	Q1956-03MS	PP071811.D	05/06/2025	05/06/2025
SB2-4-5MSD	Q1956-04MSD	PP071812.D	05/06/2025	05/06/2025
SB91-3-4	Q1956-06	PP071813.D	05/06/2025	05/06/2025

COMMENTS: _____

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167890BL

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956

SAS No.: Q1956 SDG NO.: Q1956

Lab Sample ID: PB167890BL

Lab File ID: PP071840.D

Matrix: (soil/water) WATER

Extraction: (Type) SEPF

Sulfur Cleanup: (Y/N) N

Date Extracted: 05/07/2025

Date Analyzed (1): 05/07/2025

Date Analyzed (2): 05/07/2025

Time Analyzed (1): 14:35

Time Analyzed (2): 14:35

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB167890BS	PB167890BS	PP071841.D	05/07/2025	05/07/2025
PB167890BSD	PB167890BSD	PP071842.D	05/07/2025	05/07/2025
FB-05022025	Q1956-07	PP071843.D	05/07/2025	05/07/2025

COMMENTS: _____



SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	05/01/25
Project:	Bergen Point Fueling System	Date Received:	05/02/25
Client Sample ID:	SB1-3-4	SDG No.:	Q1956
Lab Sample ID:	Q1956-01	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	89
Sample Wt/Vol:	30.07	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071809.D	1	05/06/25 08:55	05/06/25 13:45	PB167877

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	19.1	U	4.40	19.1	ug/kg
11104-28-2	Aroclor-1221	19.1	U	4.50	19.1	ug/kg
11141-16-5	Aroclor-1232	19.1	U	4.20	19.1	ug/kg
53469-21-9	Aroclor-1242	19.1	U	4.50	19.1	ug/kg
12672-29-6	Aroclor-1248	19.1	U	6.60	19.1	ug/kg
11097-69-1	Aroclor-1254	19.1	U	3.60	19.1	ug/kg
37324-23-5	Aroclor-1262	19.1	U	5.60	19.1	ug/kg
11100-14-4	Aroclor-1268	19.1	U	4.00	19.1	ug/kg
11096-82-5	Aroclor-1260	19.1	U	3.60	19.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.5		32 - 144	117%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.3		32 - 175	91%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
Data File : PP071809.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2025 13:45
Operator : YP\AJ
Sample : Q1956-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SB1-3-4

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 14:14:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.512	3.805	41389753	33079876	20.876	23.484
2) SA Decachlor...	10.223	8.832	23340776	16044971	16.119	18.294

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
Data File : PP071809.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2025 13:45
Operator : YP\AJ
Sample : Q1956-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SB1-3-4

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 14:14:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.512 min
Delta R.T.: -0.005 min
Response: 41389753
Conc: 20.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB1-3-4




#1 Tetrachloro-m-xylene

R.T.: 3.805 min
Delta R.T.: -0.007 min
Response: 33079876
Conc: 23.48 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.223 min
Delta R.T.: -0.014 min
Response: 23340776
Conc: 16.12 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.832 min
Delta R.T.: -0.017 min
Response: 16044971
Conc: 18.29 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/02/25
Project:	Bergen Point Fueling System	Date Received:	05/02/25
Client Sample ID:	SB2-4-5	SDG No.:	Q1956
Lab Sample ID:	Q1956-02	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	93.5
Sample Wt/Vol:	30.06	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071810.D	1	05/06/25 08:55	05/06/25 14:01	PB167877

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	18.1	U	4.20	18.1	ug/kg
11104-28-2	Aroclor-1221	18.1	U	4.30	18.1	ug/kg
11141-16-5	Aroclor-1232	18.1	U	4.00	18.1	ug/kg
53469-21-9	Aroclor-1242	18.1	U	4.30	18.1	ug/kg
12672-29-6	Aroclor-1248	18.1	U	6.30	18.1	ug/kg
11097-69-1	Aroclor-1254	6.90	JP	3.40	18.1	ug/kg
37324-23-5	Aroclor-1262	18.1	U	5.40	18.1	ug/kg
11100-14-4	Aroclor-1268	18.1	U	3.80	18.1	ug/kg
11096-82-5	Aroclor-1260	18.1	U	3.40	18.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.6		32 - 144	113%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.0		32 - 175	90%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
 Data File : PP071810.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2025 14:01
 Operator : YP\AJ
 Sample : Q1956-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SB2-4-5

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 14:14:48 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.508	3.805	41123156	31851556	20.742	22.612
2) SA Decachlor...	10.220	8.833	24913632	15794767	17.205	18.009
Target Compounds						
26) L6 AR-1254-1	6.512	5.692	1458108	1577220	17.792	17.256
27) L6 AR-1254-2	6.727	5.839	1615407	1315110	12.712	16.708 #
28) L6 AR-1254-3	7.090	6.244	1835241	2211627	14.363	18.662 #
29) L6 AR-1254-4	7.374	6.472	1733629	1856763	14.753	24.062 #
30) L6 AR-1254-5	7.788	6.888	1669612	2019692	15.498	19.905 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
Data File : PP071810.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2025 14:01
Operator : YP\AJ
Sample : Q1956-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SB2-4-5

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 14:14:48 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.508 min
Delta R.T.: -0.009 min
Response: 41123156
Conc: 20.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB2-4-5




#1 Tetrachloro-m-xylene

R.T.: 3.805 min
Delta R.T.: -0.008 min
Response: 31851556
Conc: 22.61 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.220 min
Delta R.T.: -0.017 min
Response: 24913632
Conc: 17.21 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.833 min
Delta R.T.: -0.016 min
Response: 15794767
Conc: 18.01 ng/ml



#26 AR-1254-1


R.T.: 6.512 min
Delta R.T.: -0.007 min
Response: 1458108
Conc: 17.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB2-4-5




#26 AR-1254-1

R.T.: 5.692 min
Delta R.T.: -0.010 min
Response: 1577220
Conc: 17.26 ng/ml




#27 AR-1254-2

R.T.: 6.727 min
Delta R.T.: -0.009 min
Response: 1615407
Conc: 12.71 ng/ml



#27 AR-1254-2


R.T.: 5.839 min
Delta R.T.: -0.012 min
Response: 1315110
Conc: 16.71 ng/ml



#28 AR-1254-3


R.T.: 7.090 min
Delta R.T.: -0.008 min
Response: 1835241
Conc: 14.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB2-4-5




#28 AR-1254-3

R.T.: 6.244 min
Delta R.T.: -0.011 min
Response: 2211627
Conc: 18.66 ng/ml




#29 AR-1254-4

R.T.: 7.374 min
Delta R.T.: -0.007 min
Response: 1733629
Conc: 14.75 ng/ml



#29 AR-1254-4


R.T.: 6.472 min
Delta R.T.: -0.011 min
Response: 1856763
Conc: 24.06 ng/ml



#30 AR-1254-5

R.T.: 7.788 min
Delta R.T.: -0.009 min
Response: 1669612
Conc: 15.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB2-4-5



#30 AR-1254-5

R.T.: 6.888 min
Delta R.T.: -0.013 min
Response: 2019692
Conc: 19.90 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/01/25
Project:	Bergen Point Fueling System	Date Received:	05/02/25
Client Sample ID:	SB91-3-4	SDG No.:	Q1956
Lab Sample ID:	Q1956-06	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	90.2
Sample Wt/Vol:	30.08	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071813.D	1	05/06/25 08:55	05/06/25 14:51	PB167877

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	18.8	U	4.40	18.8	ug/kg
11104-28-2	Aroclor-1221	18.8	U	4.50	18.8	ug/kg
11141-16-5	Aroclor-1232	18.8	U	4.10	18.8	ug/kg
53469-21-9	Aroclor-1242	18.8	U	4.40	18.8	ug/kg
12672-29-6	Aroclor-1248	18.8	U	6.50	18.8	ug/kg
11097-69-1	Aroclor-1254	18.8	U	3.50	18.8	ug/kg
37324-23-5	Aroclor-1262	18.8	U	5.60	18.8	ug/kg
11100-14-4	Aroclor-1268	18.8	U	4.00	18.8	ug/kg
11096-82-5	Aroclor-1260	18.8	U	3.60	18.8	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.8		32 - 144	114%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.2		32 - 175	81%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
 Data File : PP071813.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2025 14:51
 Operator : YP\AJ
 Sample : Q1956-06
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SB91-3-4

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 15:34:09 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.508	3.806	39137584	32102800	19.740	22.790
2) SA Decachlor...	10.219	8.833	21556000	14223608	14.887	16.218

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
Data File : PP071813.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2025 14:51
Operator : YP\AJ
Sample : Q1956-06
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SB91-3-4

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 15:34:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.508 min
Delta R.T.: -0.009 min
Response: 39137584
Conc: 19.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB91-3-4




#1 Tetrachloro-m-xylene

R.T.: 3.806 min
Delta R.T.: -0.007 min
Response: 32102800
Conc: 22.79 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.219 min
Delta R.T.: -0.018 min
Response: 21556000
Conc: 14.89 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.833 min
Delta R.T.: -0.016 min
Response: 14223608
Conc: 16.22 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/02/25
Project:	Bergen Point Fueling System	Date Received:	05/02/25
Client Sample ID:	FB-05022025	SDG No.:	Q1956
Lab Sample ID:	Q1956-07	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	430	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071843.D	1	05/07/25 09:15	05/07/25 15:24	PB167890

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.58	U	0.11	0.58	ug/L
11104-28-2	Aroclor-1221	0.58	U	0.15	0.58	ug/L
11141-16-5	Aroclor-1232	0.58	U	0.11	0.58	ug/L
53469-21-9	Aroclor-1242	0.58	U	0.14	0.58	ug/L
12672-29-6	Aroclor-1248	0.58	U	0.083	0.58	ug/L
11097-69-1	Aroclor-1254	0.58	U	0.11	0.58	ug/L
37324-23-5	Aroclor-1262	0.58	U	0.16	0.58	ug/L
11100-14-4	Aroclor-1268	0.58	U	0.13	0.58	ug/L
11096-82-5	Aroclor-1260	0.58	U	0.094	0.58	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.5		16 - 158	117%	SPK: 20
2051-24-3	Decachlorobiphenyl	13.8		10 - 173	69%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050725\
Data File : PP071843.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 May 2025 15:24
Operator : YP\AJ
Sample : Q1956-07
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
FB-05022025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 08 01:24:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.509	3.806	42454478	33089316	21.413	23.490
2) SA Decachlor...	10.223	8.833	17819934	12109804	12.306	13.807

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050725\
Data File : PP071843.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 May 2025 15:24
Operator : YP\AJ
Sample : Q1956-07
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
FB-05022025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 08 01:24:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.509 min
Delta R.T.: -0.008 min
Response: 42454478
Conc: 21.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
FB-05022025




#1 Tetrachloro-m-xylene

R.T.: 3.806 min
Delta R.T.: -0.006 min
Response: 33089316
Conc: 23.49 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.223 min
Delta R.T.: -0.015 min
Response: 17819934
Conc: 12.31 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.833 min
Delta R.T.: -0.016 min
Response: 12109804
Conc: 13.81 ng/ml



CALIBRATION SUMMARY

RETENTION TIMES OF INITIAL CALIBRATION

Contract: CAMP02Lab Code: CHEMCase No.: Q1956SAS No.: Q1956SDG NO.: Q1956Instrument ID: ECD_P

Calibration Date(s):

04/22/202504/22/2025

Calibration Times:

10:2917:49GC Column: ZB-MR1ID: 0.32 (mm)

LAB FILE ID:

RT 1000 = PP071389.DRT 750 = PP071390.DRT 500 = PP071391.DRT 250 = PP071392.DRT 050 = PP071393.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Aroclor-1016-1	(1)	5.66	5.67	5.67	5.67	5.67	5.67	5.57	5.77
Aroclor-1016-2	(2)	5.69	5.69	5.69	5.69	5.69	5.69	5.59	5.79
Aroclor-1016-3	(3)	5.75	5.75	5.75	5.75	5.75	5.75	5.65	5.85
Aroclor-1016-4	(4)	5.85	5.85	5.85	5.85	5.85	5.85	5.75	5.95
Aroclor-1016-5	(5)	6.14	6.14	6.15	6.14	6.14	6.14	6.04	6.24
Aroclor-1260-1	(1)	7.26	7.26	7.27	7.26	7.26	7.26	7.16	7.36
Aroclor-1260-2	(2)	7.51	7.51	7.52	7.52	7.52	7.52	7.42	7.62
Aroclor-1260-3	(3)	7.87	7.87	7.88	7.87	7.88	7.87	7.77	7.97
Aroclor-1260-4	(4)	8.10	8.10	8.10	8.10	8.10	8.10	8.00	8.20
Aroclor-1260-5	(5)	8.42	8.42	8.42	8.42	8.42	8.42	8.32	8.52
Decachlorobiphenyl		10.23	10.23	10.24	10.23	10.24	10.23	10.13	10.33
Tetrachloro-m-xylene		4.51	4.51	4.52	4.51	4.52	4.51	4.41	4.61
Aroclor-1242-1	(1)	5.67	5.67	5.67	5.67	5.67	5.67	5.57	5.77
Aroclor-1242-2	(2)	5.69	5.69	5.69	5.69	5.69	5.69	5.59	5.79
Aroclor-1242-3	(3)	5.75	5.75	5.75	5.75	5.75	5.75	5.65	5.85
Aroclor-1242-4	(4)	5.85	5.85	5.85	5.85	5.85	5.85	5.75	5.95
Aroclor-1242-5	(5)	6.58	6.58	6.58	6.58	6.58	6.58	6.48	6.68
Decachlorobiphenyl		10.23	10.24	10.24	10.23	10.24	10.24	10.14	10.34
Tetrachloro-m-xylene		4.51	4.52	4.51	4.51	4.52	4.51	4.41	4.61
Aroclor-1248-1	(1)	5.67	5.67	5.67	5.67	5.67	5.67	5.57	5.77
Aroclor-1248-2	(2)	5.94	5.94	5.94	5.94	5.94	5.94	5.84	6.04
Aroclor-1248-3	(3)	6.14	6.14	6.14	6.14	6.14	6.14	6.04	6.24
Aroclor-1248-4	(4)	6.54	6.54	6.54	6.54	6.54	6.54	6.44	6.64
Aroclor-1248-5	(5)	6.58	6.58	6.58	6.58	6.58	6.58	6.48	6.68
Decachlorobiphenyl		10.24	10.24	10.24	10.24	10.24	10.24	10.14	10.34
Tetrachloro-m-xylene		4.52	4.52	4.51	4.51	4.51	4.51	4.41	4.61
Aroclor-1254-1	(1)	6.52	6.52	6.52	6.52	6.52	6.52	6.42	6.62
Aroclor-1254-2	(2)	6.74	6.74	6.74	6.73	6.74	6.74	6.64	6.84
Aroclor-1254-3	(3)	7.10	7.10	7.10	7.10	7.10	7.10	7.00	7.20
Aroclor-1254-4	(4)	7.38	7.38	7.38	7.38	7.38	7.38	7.28	7.48
Aroclor-1254-5	(5)	7.80	7.80	7.80	7.80	7.80	7.80	7.70	7.90
Decachlorobiphenyl		10.24	10.24	10.24	10.23	10.24	10.24	10.14	10.34
Tetrachloro-m-xylene		4.52	4.52	4.52	4.51	4.51	4.51	4.41	4.61
Aroclor-1268-1	(1)	8.73	8.73	8.73	8.74	8.73	8.73	8.63	8.83
Aroclor-1268-2	(2)	8.83	8.83	8.83	8.83	8.83	8.83	8.73	8.93
Aroclor-1268-3	(3)	9.06	9.06	9.06	9.06	9.06	9.06	8.96	9.16
Aroclor-1268-4	(4)	9.48	9.48	9.48	9.48	9.48	9.48	9.38	9.58
Aroclor-1268-5	(5)	9.89	9.89	9.89	9.90	9.89	9.90	9.80	10.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	10.24	10.23	10.24	10.24	10.24	10.24	10.14	10.34
Tetrachloro-m-xylene	4.51	4.52	4.52	4.52	4.51	4.52	4.42	4.62

RETENTION TIMES OF INITIAL CALIBRATION

Contract: **CAMP02**

Lab Code: CHEM

Case No.: Q1956

SAS No.: Q1956

SDG NO.: Q1956

Instrument ID: ECD_P

Calibration Date(s):

04/22/2025

04/22/2025

Calibration Times:

10:29

17:49

GC Column: ZB-MR2

ID: 0.32 (mm)

LAB FILE ID:

RT 1000 = PP071389.D

RT 750 = PP071390.D

RT 500 = PP071391.D

RT 250 = PP071392.D

RT 050 = PP071393.D

[illegible]

[illegible]

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CAMP02
Lab Code: CHEM **Case No.:** Q1956 **SAS No.:** Q1956 **SDG NO.:** Q1956
Instrument ID: ECD_P **Calibration Date(s):** 04/22/2025 04/22/2025
Calibration Times: 10:29 17:49

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP071389.D</u>		CF 750 = <u>PP071390.D</u>			
CF 500 = <u>PP071391.D</u>		CF 250 = <u>PP071392.D</u>		CF 050 = <u>PP071393.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	64709894	67652695	70083532	71128656	61674340	67049823
Aroclor-1016-2	(2)	98977911	102364897	107369074	108813892	95615840	102628323
Aroclor-1016-3	(3)	59905708	62127143	64296308	65849928	56056440	61647105
Aroclor-1016-4	(4)	50563747	52419789	53619076	54941428	44946700	51298148
Aroclor-1016-5	(5)	47772963	49344036	50670668	50888564	42515680	48238382
Aroclor-1260-1	(1)	90879847	94857525	98814986	99659744	95910940	96024608
Aroclor-1260-2	(2)	137163459	142386411	147498372	149887700	138450980	143077384
Aroclor-1260-3	(3)	110350751	114445481	117702804	126515524	91505860	112104084
Aroclor-1260-4	(4)	105570732	110780085	111211068	131260612	105501660	112864831
Aroclor-1260-5	(5)	220536830	229070229	228919678	255108180	199231240	226573231
Decachlorobiphenyl		1409816860	1467002307	1522583160	1539948480	1300700600	1448010281
Tetrachloro-m-xylene		1947183160	2005072840	2066603680	2100352680	1793966200	1982635712
Aroclor-1242-1	(1)	53870259	55684409	58494980	58676784	47130380	54771362
Aroclor-1242-2	(2)	81881116	85641493	90236470	89744200	77399340	84980524
Aroclor-1242-3	(3)	49423643	51294363	53309060	53836868	47752740	51123335
Aroclor-1242-4	(4)	41638031	42855355	45034468	44656444	36858500	42208560
Aroclor-1242-5	(5)	44955701	46392219	48330700	47956460	43817460	46290508
Decachlorobiphenyl		1368547670	1421241013	1488838040	1503303480	1179092600	1392204561
Tetrachloro-m-xylene		1868945350	1926494133	2017272340	2030390160	1613146400	1891249677
Aroclor-1248-1	(1)	42447397	43370463	44490672	45736348	38649560	42938888
Aroclor-1248-2	(2)	57932834	60215229	62023042	63285912	58705480	60432499
Aroclor-1248-3	(3)	65030901	68053485	69372090	70318056	61147600	66784426
Aroclor-1248-4	(4)	80777360	84441085	86910244	89142916	81782860	84610893
Aroclor-1248-5	(5)	76029464	78502729	80776474	83552224	78380620	79448302
Decachlorobiphenyl		1379130770	1440268240	1485619680	1521197800	1358343200	1436911938
Tetrachloro-m-xylene		1879191440	1925518227	1989640420	2037603040	1811949400	1928780505
Aroclor-1254-1	(1)	78574372	81246615	85577334	88598188	75759400	81951182
Aroclor-1254-2	(2)	121624435	125816099	132146328	137115708	118678260	127076166
Aroclor-1254-3	(3)	123432630	127799839	132981654	137623252	117036460	127774767
Aroclor-1254-4	(4)	112512019	116259037	121737436	124392856	112665620	117513394
Aroclor-1254-5	(5)	106399089	105922557	113785120	116153044	96391680	107730298
Decachlorobiphenyl		1410101810	1469202520	1526389740	1549706840	1328725400	1456825262
Tetrachloro-m-xylene		1888241120	1936990333	2012368600	2071565920	1722515200	1926336235
Aroclor-1268-1	(1)	299594266	309097848	319744804	331826088	282063800	308465361



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	251109850	259047004	267960354	275856908	228123960	256419615	7
Aroclor-1268-3	(3)	211915080	218354113	225284280	231505696	216470880	220706010	4
Aroclor-1268-4	(4)	93231632	96431025	99007040	98059984	82736300	93893196	7
Aroclor-1268-5	(5)	590082071	603740247	621062448	631335032	536707900	596585540	6
Decachlorobiphenyl		2486623660	2556709627	2669414960	2694667240	2263272000	2534137497	7
Tetrachloro-m-xylene		1883977950	1931567613	1982008260	2006991440	1590966000	1879102253	9

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CAMP02
Lab Code: CHEM **Case No.:** Q1956 **SAS No.:** Q1956 **SDG NO.:** Q1956
Instrument ID: ECD_P **Calibration Date(s):** 04/22/2025 04/22/2025
Calibration Times: 10:29 17:49

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP071389.D</u>		CF 750 = <u>PP071390.D</u>			
CF 500 = <u>PP071391.D</u>		CF 250 = <u>PP071392.D</u>		CF 050 = <u>PP071393.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	48121828	49414308	53806128	56180596	60145020	53533576
Aroclor-1016-2	(2)	69736709	72096436	77213744	78520924	84250740	76363711
Aroclor-1016-3	(3)	38517048	40455577	43831736	45082616	43165080	42210411
Aroclor-1016-4	(4)	31206636	32814844	36123020	37095092	35190100	34485938
Aroclor-1016-5	(5)	39458687	41985771	45850608	46345748	48595640	44447291
Aroclor-1260-1	(1)	63219917	66123699	73727666	76350368	82074520	72299234
Aroclor-1260-2	(2)	77338708	77548395	89921718	91492956	101005300	87461415
Aroclor-1260-3	(3)	72062678	70944365	81698234	79344004	87888600	78387576
Aroclor-1260-4	(4)	57333188	56779915	65331750	68083052	68025140	63110609
Aroclor-1260-5	(5)	143809415	139161576	155842678	157512280	157560640	150777318
Decachlorobiphenyl		831147600	875337973	902319620	934998560	841460400	877052831
Tetrachloro-m-xylene		1380433570	1378340613	1415112020	1448341640	1420974200	1408640409
Aroclor-1242-1	(1)	42717527	44660211	47919806	49576672	46343800	46243603
Aroclor-1242-2	(2)	62346958	64901092	68295906	70279172	64290540	66022734
Aroclor-1242-3	(3)	34614031	35703872	38469574	40739776	32714120	36448275
Aroclor-1242-4	(4)	33518120	34659665	37548334	39683016	33143760	35710579
Aroclor-1242-5	(5)	41214500	42688541	44971456	47258656	41407120	43508055
Decachlorobiphenyl		815615080	851726400	897002660	917481280	766770800	849719244
Tetrachloro-m-xylene		1295159620	1325412680	1503876340	1432488040	1389045200	1389196376
Aroclor-1248-1	(1)	30996932	35945815	36538014	39256084	37796080	36106585
Aroclor-1248-2	(2)	43624561	47309389	48616518	53505512	51756020	48962400
Aroclor-1248-3	(3)	45791659	49899927	50774174	56002176	52959580	51085503
Aroclor-1248-4	(4)	53871444	58562581	58680784	64889280	67282380	60657294
Aroclor-1248-5	(5)	52150607	56290108	57608714	65271236	60133720	58290877
Decachlorobiphenyl		826375100	860756013	900400040	940375840	921771600	889935719
Tetrachloro-m-xylene		1324712230	1514156880	1548203640	1434858960	1416930600	1447772462
Aroclor-1254-1	(1)	81690839	87361264	89755404	97285536	100908960	91400401
Aroclor-1254-2	(2)	70045476	75469849	77317914	83154376	87573280	78712179
Aroclor-1254-3	(3)	109725744	114103376	120096818	124380164	124235440	118508308
Aroclor-1254-4	(4)	70169900	74068720	77810162	81877428	81901880	77165618
Aroclor-1254-5	(5)	92229111	96109599	104690794	107384400	106924600	101467701
Decachlorobiphenyl		846307500	875335867	916104900	955261880	868223800	892246789
Tetrachloro-m-xylene		1390553300	1501013893	1450901500	1518608160	1468503000	1465915971
Aroclor-1268-1	(1)	181860271	188460193	203483780	205414332	201904640	196224643



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	154801411	159518377	172755008	173272408	170561360	166181713	5
Aroclor-1268-3	(3)	128491160	133089205	143083692	144773184	135469120	136981272	5
Aroclor-1268-4	(4)	56964280	59695839	63521398	64177792	55793320	60030526	6
Aroclor-1268-5	(5)	367021219	377695863	385175068	387479012	342861840	372046600	5
Decachlorobiphenyl		1499051200	1546040093	1602790460	1652739760	1464953600	1553115023	5
Tetrachloro-m-xylene		1476470990	1445952813	1464214380	1457737600	1300882000	1429051557	5



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Instrument ID: ECD_P Date(s) Analyzed: 04/22/2025 04/22/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.72	4.62	4.82	24116800
		2	4.80	4.70	4.90	18016900
		3	4.88	4.78	4.98	56611200
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.88	4.78	4.98	44912000
		2	5.40	5.30	5.50	23013000
		3	5.69	5.59	5.79	47217600
		4	5.85	5.75	5.95	23179400
		5	5.94	5.84	6.04	16145700
Aroclor-1262	500	1	8.10	8.00	8.20	144491000
		2	8.42	8.32	8.52	274186000
		3	8.74	8.64	8.84	182878000
		4	8.82	8.72	8.92	134211000
		5	9.48	9.38	9.58	88162000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Instrument ID: ECD_P Date(s) Analyzed: 04/22/2025 04/22/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.02	3.92	4.12	21995000
		2	4.11	4.01	4.21	16561700
		3	4.19	4.09	4.29	48477000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.19	4.09	4.29	38317800
		2	4.92	4.82	5.02	38187400
		3	5.09	4.99	5.19	21175200
		4	5.18	5.08	5.28	18540400
		5	5.35	5.25	5.45	19901900
Aroclor-1262	500	1	6.94	6.84	7.04	113612000
		2	7.20	7.10	7.30	92339400
		3	7.72	7.62	7.82	75694400
		4	7.79	7.69	7.89	124826000
		5	8.29	8.19	8.39	55493200

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071389.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 10:29
 Operator : YP\AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:18:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.511	3.811	194.7E6	138.0E6	97.025	98.760
2) SA Decachlor...	10.230	8.849	141.0E6	83114760	96.154	95.894
Target Compounds						
3) L1 AR-1016-1	5.664	4.897	64709894	48121828	960.134	944.232
4) L1 AR-1016-2	5.686	4.915	98977911	69736709	959.335	949.119
5) L1 AR-1016-3	5.749	5.092	59905708	38517048	964.650	935.461
6) L1 AR-1016-4	5.846	5.134	50563747	31206636	970.673	926.980
7) L1 AR-1016-5	6.139	5.349	47772963	39458687	970.565	925.074
31) L7 AR-1260-1	7.259	6.384	90879847	63219917	958.169	923.272
32) L7 AR-1260-2	7.512	6.572	137.2E6	77338708	963.694	924.770
33) L7 AR-1260-3	7.872	6.726	110.4E6	72062678	967.762	937.334
34) L7 AR-1260-4	8.095	7.197	105.6E6	57333188	973.982	934.793
35) L7 AR-1260-5	8.416	7.439	220.5E6	143.8E6	981.349	959.843



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071389.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 10:29
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:18:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071390.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 10:45
 Operator : YP\AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:20:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.811	150.4E6	103.4E6	74.955	74.302
2) SA Decachlor...	10.233	8.848	110.0E6	65650348	75.027	75.495
Target Compounds						
3) L1 AR-1016-1	5.668	4.897	50739521	37060731	751.897	734.641
4) L1 AR-1016-2	5.690	4.915	76773673	54072327	746.071	740.558
5) L1 AR-1016-3	5.752	5.092	46595357	30341683	750.210	741.220
6) L1 AR-1016-4	5.849	5.134	39314842	24611133	753.145	737.269
7) L1 AR-1016-5	6.142	5.349	37008027	31489328	751.240	742.118
31) L7 AR-1260-1	7.261	6.383	71143144	49592774	750.053	732.641
32) L7 AR-1260-2	7.514	6.572	106.8E6	58161296	750.195	712.735
33) L7 AR-1260-3	7.873	6.725	85834111	53208274	751.834	710.374
34) L7 AR-1260-4	8.097	7.197	83085064	42584936	760.941	711.945
35) L7 AR-1260-5	8.418	7.438	171.8E6	104.4E6	759.599	713.546



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071390.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 10:45
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:20:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071391.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 11:02
 Operator : YP\AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:15:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.517	3.812	103.3E6	70755601	50.000	50.000
2) SA Decachlor...	10.237	8.849	76129158	45115981	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.671	4.898	35041766	26903064	500.000	500.000
4) L1 AR-1016-2	5.693	4.916	53684537	38606872	500.000	500.000
5) L1 AR-1016-3	5.754	5.093	32148154	21915868	500.000	500.000
6) L1 AR-1016-4	5.852	5.136	26809538	18061510	500.000	500.000
7) L1 AR-1016-5	6.145	5.350	25335334	22925304	500.000	500.000
31) L7 AR-1260-1	7.265	6.385	49407493	36863833	500.000	500.000
32) L7 AR-1260-2	7.518	6.573	73749186	44960859	500.000	500.000
33) L7 AR-1260-3	7.877	6.727	58851402	40849117	500.000	500.000
34) L7 AR-1260-4	8.102	7.198	55605534	32665875	500.000	500.000
35) L7 AR-1260-5	8.422	7.439	114.5E6	77921339	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071391.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 11:02
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:15:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071392.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 11:18
 Operator : YP\AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:30:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:29:52 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	52508817	36208541	25.869	25.761
2) SA Decachlor...	10.233	8.850	38498712	23374964	25.928	26.384
Target Compounds						
3) L1 AR-1016-1	5.668	4.899	17782164	14045149	259.997	270.720
4) L1 AR-1016-2	5.689	4.917	27203473	19630231	260.616	263.876
5) L1 AR-1016-3	5.752	5.094	16462482	11270654	261.124	268.530
6) L1 AR-1016-4	5.849	5.136	13735357	9273773	259.716	270.294
7) L1 AR-1016-5	6.142	5.351	12722141	11586437	256.138	266.906
31) L7 AR-1260-1	7.261	6.386	24914936	19087592	259.387	273.244
32) L7 AR-1260-2	7.515	6.574	37471925	22873239	259.800	272.056
33) L7 AR-1260-3	7.873	6.727	31628881	19836001	269.748	260.958
34) L7 AR-1260-4	8.098	7.199	32815153	17020763	286.081	275.052
35) L7 AR-1260-5	8.417	7.441	63777045	39378070	273.242	264.138



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071392.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 11:18
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:30:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:29:52 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071393.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 11:34
 Operator : YP\AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:47:04 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:46:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	8969831	7104871	4.524	5.044
2) SA Decachlor...	10.235	8.850	6503503	4207302	4.491	4.797
Target Compounds						
3) L1 AR-1016-1	5.669	4.898	3083717	3007251	45.991	56.175
4) L1 AR-1016-2	5.691	4.916	4780792	4212537	46.584	55.164
5) L1 AR-1016-3	5.753	5.094	2802822	2158254	45.466	51.131
6) L1 AR-1016-4	5.851	5.135	2247335	1759505	43.809	51.021
7) L1 AR-1016-5	6.144	5.349	2125784	2429782	44.068	54.667
31) L7 AR-1260-1	7.262	6.384	4795547	4103726	49.941	56.760
32) L7 AR-1260-2	7.516	6.573	6922549	5050265	48.383	57.743
33) L7 AR-1260-3	7.875	6.726	4575293	4394430	40.810m	56.060 #
34) L7 AR-1260-4	8.098	7.197	5275083	3401257	46.807m	53.894
35) L7 AR-1260-5	8.417	7.439	9961562	7878032	43.897m	52.249

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071393.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 11:34
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:47:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:46:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071394.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 11:51
 Operator : YP\AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 12:26:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 12:26:37 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.812	102.0E6	70886063	50.000	50.000
2) SA Decachlor...	10.233	8.849	73129811	43684216	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	4.715	4.022	12058427	10997465	500.000	500.000
9) L2 AR-1221-2	4.800	4.109	9008471	8280856	500.000	500.000
10) L2 AR-1221-3	4.876	4.185	28305615	24238496	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071394.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 11:51
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 12:26:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 12:26:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071395.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 12:07
 Operator : YP\AJ
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 12:29:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 12:26:37 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	97587735	67524433	50.000	50.000
2) SA Decachlor...	10.233	8.851	71248479	42941843	50.000	50.000
Target Compounds						
11) L3 AR-1232-1	4.878	4.186	22456038	19158895	500.000	500.000
12) L3 AR-1232-2	5.403	4.917	11506547	19093695	500.000	500.000
13) L3 AR-1232-3	5.690	5.094	23608778	10587643	500.000	500.000
14) L3 AR-1232-4	5.850	5.178	11589674	9270206	500.000	500.000
15) L3 AR-1232-5	5.940	5.351	8072861	9950926	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071395.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 12:07
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 12:29:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 12:26:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071396.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 12:23
 Operator : YP\AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:20:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:13:15 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.812	186.9E6	129.5E6	96.183	92.543
2) SA Decachlor...	10.233	8.849	136.9E6	81561508	95.790	95.248
Target Compounds						
16) L4 AR-1242-1	5.668	4.897	53870259	42717527	958.842	942.603
17) L4 AR-1242-2	5.689	4.916	81881116	62346958	951.456	954.464
18) L4 AR-1242-3	5.751	5.093	49423643	34614031	962.179	947.245
19) L4 AR-1242-4	5.848	5.177	41638031	33518120	960.813	943.289
20) L4 AR-1242-5	6.580	5.701	44955701	41214500	963.821	956.409



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071396.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 12:23
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:20:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:13:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071397.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 12:39
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:23:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:13:15 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	144.5E6	99405951	74.571	72.305
2) SA Decachlor...	10.236	8.850	106.6E6	63879480	74.739	74.732
Target Compounds						
16) L4 AR-1242-1	5.669	4.899	41763307	33495158	745.553	742.700
17) L4 AR-1242-2	5.691	4.917	64231120	48675819	747.572	746.776
18) L4 AR-1242-3	5.753	5.094	38470772	26777904	749.299	738.446
19) L4 AR-1242-4	5.850	5.179	32141516	25994749	744.431	737.606
20) L4 AR-1242-5	6.581	5.703	34794164	32016406	747.305	745.293



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071397.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 12:39
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:23:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:13:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071398.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 12:56
 Operator : YP\AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:13:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:13:15 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	100.9E6	75193817	50.000	50.000
2) SA Decachlor...	10.235	8.850	74441902	44850133	50.000	50.000
Target Compounds						
16) L4 AR-1242-1	5.669	4.898	29247490	23959903	500.000	500.000
17) L4 AR-1242-2	5.690	4.916	45118235	34147953	500.000	500.000
18) L4 AR-1242-3	5.752	5.093	26654530	19234787	500.000	500.000
19) L4 AR-1242-4	5.849	5.178	22517234	18774167	500.000	500.000
20) L4 AR-1242-5	6.581	5.701	24165350	22485728	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071398.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 12:56
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:13:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:13:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071399.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 13:12
 Operator : YP\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:34:40 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:34:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	50759754	35812201	25.888	25.778
2) SA Decachlor...	10.233	8.850	37582587	22937032	26.000	26.351
Target Compounds						
16) L4 AR-1242-1	5.668	4.899	14669196	12394168	258.800	268.164
17) L4 AR-1242-2	5.689	4.917	22436050	17569793	258.254	264.383
18) L4 AR-1242-3	5.751	5.094	13459217	10184944	259.001	272.457
19) L4 AR-1242-4	5.849	5.178	11164111	9920754	256.375	272.906
20) L4 AR-1242-5	6.580	5.702	11989115	11814664	255.584	268.312



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071399.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 13:12
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:34:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:34:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071400.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 13:28
 Operator : YP\AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:44:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:44:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	8065732	6945226	4.265	4.999
2) SA Decachlor...	10.237	8.850	5895463	3833854	4.235	4.512
Target Compounds						
16) L4 AR-1242-1	5.670	4.899	2356519	2317190	43.025	50.108
17) L4 AR-1242-2	5.691	4.917	3869967	3214527	45.539	48.688
18) L4 AR-1242-3	5.754	5.094	2387637	1635706	46.703	44.877
19) L4 AR-1242-4	5.851	5.179	1842925	1657188	43.662	46.406
20) L4 AR-1242-5	6.582	5.702	2190873	2070356	47.329	47.586



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071400.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 13:28
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:44:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:44:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071401.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 13:45
 Operator : YP\AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 14:30:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 14:27:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	187.9E6	132.5E6	97.145	92.221
2) SA Decachlor...	10.237	8.850	137.9E6	82637510	96.283	95.713
Target Compounds						
21) L5 AR-1248-1	5.669	4.898	42447397	30996932	976.497	917.952
22) L5 AR-1248-2	5.941	5.136	57932834	43624561	965.902	945.881
23) L5 AR-1248-3	6.144	5.178	65030901	45791659	967.700	948.403
24) L5 AR-1248-4	6.543	5.350	80777360	53871444	963.427	957.270
25) L5 AR-1248-5	6.582	5.743	76029464	52150607	969.727	950.272



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071401.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 13:45
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 14:30:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 14:27:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071402.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 14:01
 Operator : YP\AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 14:33:31 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 14:27:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.516	3.813	144.4E6	113.6E6	74.770	77.657
2) SA Decachlor...	10.238	8.850	108.0E6	64556701	75.275	74.847
Target Compounds						
21) L5 AR-1248-1	5.670	4.898	32527847	26959361	748.865	781.576
22) L5 AR-1248-2	5.941	5.135	45161422	35482042	751.976	762.779
23) L5 AR-1248-3	6.144	5.177	51040114	37424945	756.312	766.560
24) L5 AR-1248-4	6.543	5.350	63330814	43921936	753.553	770.043
25) L5 AR-1248-5	6.583	5.743	58877047	42217581	750.636	762.741



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071402.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 14:01
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 14:33:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 14:27:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071403.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 14:17
 Operator : YP\AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 14:28:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 14:27:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	99482021	77410182	50.000	50.000
2) SA Decachlor...	10.237	8.850	74280984	45020002	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	5.668	4.898	22245336	18269007	500.000	500.000
22) L5 AR-1248-2	5.940	5.136	31011521	24308259	500.000	500.000
23) L5 AR-1248-3	6.142	5.178	34686045	25387087	500.000	500.000
24) L5 AR-1248-4	6.542	5.350	43455122	29340392	500.000	500.000
25) L5 AR-1248-5	6.581	5.743	40388237	28804357	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071403.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 14:17
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 14:28:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 14:27:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071404.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 14:33
 Operator : YP\AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 14:46:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 14:46:22 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	50940076	35871474	26.017	24.646
2) SA Decachlor...	10.236	8.850	38029945	23509396	26.110	26.655
Target Compounds						
21) L5 AR-1248-1	5.668	4.898	11434087	9814021	259.799	275.024
22) L5 AR-1248-2	5.940	5.135	15821478	13376378	259.947	277.150
23) L5 AR-1248-3	6.143	5.178	17579514	14000544	257.788	276.598
24) L5 AR-1248-4	6.542	5.350	22285729	16222320	261.208	274.950
25) L5 AR-1248-5	6.581	5.743	20888056	16317809	262.033	282.168



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071404.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 14:33
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 14:46:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 14:46:22 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071405.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 14:50
 Operator : YP\AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 15:04:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 15:04:44 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.813	9059747	7084653	4.697	4.893
2) SA Decachlor...	10.236	8.850	6791716	4608858	4.727	5.179
Target Compounds						
21) L5 AR-1248-1	5.667	4.899	1932478	1889804	45.005	52.340
22) L5 AR-1248-2	5.939	5.136	2935274	2587801	48.571	52.853
23) L5 AR-1248-3	6.142	5.178	3057380	2647979	45.780	51.834
24) L5 AR-1248-4	6.541	5.350	4089143	3364119	48.329	55.461
25) L5 AR-1248-5	6.581	5.744	3919031	3006686	49.328	51.581



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071405.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 14:50
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 15:04:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 15:04:44 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071406.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 15:06
 Operator : YP\AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 15:51:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 15:48:40 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	188.8E6	139.1E6	96.818	97.876
2) SA Decachlor...	10.236	8.850	141.0E6	84630750	96.040	96.040
Target Compounds						
26) L6 AR-1254-1	6.519	5.702	78574372	81690839	957.338	952.962
27) L6 AR-1254-2	6.735	5.851	121.6E6	70045476	958.538	950.650
28) L6 AR-1254-3	7.099	6.254	123.4E6	109.7E6	962.759	954.874
29) L6 AR-1254-4	7.381	6.483	112.5E6	70169900	960.617	948.370
30) L6 AR-1254-5	7.798	6.901	106.4E6	92229111	966.455	936.717



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071406.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 15:06
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 15:51:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 15:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071407.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 15:22
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 15:54:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 15:48:40 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	145.3E6	112.6E6	74.658	77.773
2) SA Decachlor...	10.236	8.850	110.2E6	65650190	75.033	74.666
Target Compounds						
26) L6 AR-1254-1	6.519	5.703	60934961	65520948	744.931	759.494
27) L6 AR-1254-2	6.735	5.851	94362074	56602387	745.775	762.037
28) L6 AR-1254-3	7.098	6.255	95849879	85577532	748.410	746.476
29) L6 AR-1254-4	7.381	6.483	87194278	55551540	746.295	750.532
30) L6 AR-1254-5	7.797	6.901	79441918	72082199	721.000m	737.969

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071407.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 15:22
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 15:54:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 15:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071408.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 15:38
 Operator : YP\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 15:48:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 15:48:40 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	100.6E6	72545075	50.000	50.000
2) SA Decachlor...	10.237	8.849	76319487	45805245	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.519	5.702	42788667	44877702	500.000	500.000
27) L6 AR-1254-2	6.735	5.851	66073164	38658957	500.000	500.000
28) L6 AR-1254-3	7.098	6.254	66490827	60048409	500.000	500.000
29) L6 AR-1254-4	7.381	6.483	60868718	38905081	500.000	500.000
30) L6 AR-1254-5	7.797	6.901	56892560	52345397	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071408.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 15:38
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 15:48:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 15:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071409.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 15:55
 Operator : YP\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 16:07:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 16:07:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.813	51789148	37965204	26.192	25.910
2) SA Decachlor...	10.234	8.850	38742671	23881547	26.022	26.587
Target Compounds						
26) L6 AR-1254-1	6.517	5.703	22149547	24321384	265.267	273.203
27) L6 AR-1254-2	6.733	5.851	34278927	20788594	265.367	271.757
28) L6 AR-1254-3	7.096	6.255	34405813	31095041	263.728	265.596
29) L6 AR-1254-4	7.379	6.483	31098214	20469357	261.934	269.399
30) L6 AR-1254-5	7.796	6.901	29038261	26846100	262.635	268.183



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071409.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 15:55
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 16:07:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 16:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071410.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 16:11
 Operator : YP\AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 16:54:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 16:54:36 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	8612576	7342515	4.471m	5.009
2) SA Decachlor...	10.236	8.850	6643627	4341119	4.560	4.865
Target Compounds						
26) L6 AR-1254-1	6.520	5.703	3787970	5045448	46.222	55.202
27) L6 AR-1254-2	6.735	5.852	5933913	4378664	46.696	55.629
28) L6 AR-1254-3	7.099	6.255	5851823	6211772	45.798	52.416
29) L6 AR-1254-4	7.382	6.484	5633281	4095094	47.937	53.069
30) L6 AR-1254-5	7.798	6.901	4819584	5346230	44.737	52.689

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071410.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 16:11
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 16:54:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 16:54:36 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071411.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 16:27
 Operator : YP\AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 17:08:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 17:08:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	98220466	71170259	50.000	50.000
2) SA Decachlor...	10.235	8.850	74643123	44870081	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	8.100	6.939	72245642	56806132	500.000	500.000
37) L8 AR-1262-2	8.420	7.198	137.1E6	46169678	500.000	500.000
38) L8 AR-1262-3	8.737	7.721	91438860	37847231	500.000	500.000
39) L8 AR-1262-4	8.822	7.786	67105487	62412938	500.000	500.000
40) L8 AR-1262-5	9.478	8.286	44080983	27746588	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071411.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 16:27
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 17:08:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 17:08:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071412.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 16:44
 Operator : YP\AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:44:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	188.4E6	147.6E6	95.054	100.837
2) SA Decachlor...	10.235	8.850	248.7E6	149.9E6	93.152	93.528
Target Compounds						
41) L9 AR-1268-1	8.733	7.722	299.6E6	181.9E6	936.979	895.922m
42) L9 AR-1268-2	8.826	7.788	251.1E6	154.8E6	937.116	896.075
43) L9 AR-1268-3	9.058	7.993	211.9E6	128.5E6	940.656	898.014
44) L9 AR-1268-4	9.477	8.287	93231632	56964280	941.667	896.773
45) L9 AR-1268-5	9.894	8.588	590.1E6	367.0E6	950.117	952.869

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071412.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 16:44
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:44:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071413.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 17:00
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:45:40 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	144.9E6	108.4E6	73.091	74.065
2) SA Decachlor...	10.234	8.849	191.8E6	116.0E6	71.833	72.344
Target Compounds						
41) L9 AR-1268-1	8.733	7.721	231.8E6	141.3E6	725.026	696.327m
42) L9 AR-1268-2	8.826	7.787	194.3E6	119.6E6	725.052	692.534
43) L9 AR-1268-3	9.058	7.991	163.8E6	99816904	726.929	697.612
44) L9 AR-1268-4	9.476	8.285	72323269	44771879	730.486	704.831
45) L9 AR-1268-5	9.894	8.587	452.8E6	283.3E6	729.082	735.437

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071413.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 17:00
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

AR1268ICC750

Manual Integrations



APPROVED

Reviewed By :Yogesh Patel 04/23/2025

Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:45:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071414.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 17:16
 Operator : YP\AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:46:05 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.814	99100413	73210719	50.000	50.000
2) SA Decachlor...	10.236	8.851	133.5E6	80139523	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	8.733	7.722	159.9E6	101.7E6	500.000	501.224m
42) L9 AR-1268-2	8.826	7.788	134.0E6	86377504	500.000	500.000
43) L9 AR-1268-3	9.059	7.993	112.6E6	71541846	500.000	500.000
44) L9 AR-1268-4	9.478	8.287	49503520	31760699	500.000	500.000
45) L9 AR-1268-5	9.894	8.589	310.5E6	192.6E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071414.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 17:16
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:46:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071415.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 17:33
 Operator : YP\AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:47:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.517	3.813	50174786	36443440	25.315	24.889
2) SA Decachlor...	10.236	8.850	67366681	41318494	25.236	25.779
Target Compounds						
41) L9 AR-1268-1	8.735	7.721	82956522	51353583	259.446	252.990m
42) L9 AR-1268-2	8.828	7.788	68964227	43318102	257.367	250.749
43) L9 AR-1268-3	9.061	7.991	57876424	36193296	256.904	252.952
44) L9 AR-1268-4	9.480	8.286	24514996	16044448	247.609	252.583
45) L9 AR-1268-5	9.897	8.587	157.8E6	96869753	254.135	251.495

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071415.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 17:33
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:47:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071416.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 17:49
 Operator : YP\AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:47:36 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	7954830	6504410	4.014	4.442
2) SA Decachlor...	10.235	8.850	11316360	7324768	4.239	4.570
Target Compounds						
41) L9 AR-1268-1	8.731	7.721	14103190	10095232	44.108m	49.733m
42) L9 AR-1268-2	8.825	7.787	11406198	8528068	42.567m	49.365
43) L9 AR-1268-3	9.058	7.992	10823544	6773456	48.044	47.339
44) L9 AR-1268-4	9.476	8.287	4136815	2789666	41.783	43.917
45) L9 AR-1268-5	9.894	8.588	26835395	17143092	43.209	44.507

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071416.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 17:49
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:47:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071417.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 18:05
 Operator : YP\AJ
 Sample : PP042225ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 19:27:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 17:08:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	98003244	70310326	49.431	49.914
2) SA Decachlor...	10.233	8.849	73424396	44681766	50.707	50.945
Target Compounds						
3) L1 AR-1016-1	5.668	4.898	33546404	27255669	500.321	509.132
4) L1 AR-1016-2	5.690	4.916	50377176	38783246	490.870	507.875
5) L1 AR-1016-3	5.752	5.093	30221252	21921068	490.230	519.328
6) L1 AR-1016-4	5.850	5.135	25444869	18087794	496.019	524.498
7) L1 AR-1016-5	6.142	5.349	24049539	22819712	498.556	513.411
31) L7 AR-1260-1	7.262	6.384	46974832	35744650	489.196	494.399
32) L7 AR-1260-2	7.515	6.573	70731391	43416292	494.358	496.405
33) L7 AR-1260-3	7.875	6.726	58007061	39241127	517.439	500.604
34) L7 AR-1260-4	8.098	7.198	56235072	31343464	498.252	496.643
35) L7 AR-1260-5	8.418	7.439	114.7E6	75416996	506.086	500.188



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071417.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 18:05
Operator : YP\AJ
Sample : PP042225ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP042225

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 19:27:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 17:08:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071418.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 18:38
 Operator : YP\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 02:23:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 02:23:12 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	98322267	73653264	48.740	48.976
2) SA Decachlor...	10.234	8.849	73973652	44775920	49.685	49.917
Target Compounds						
16) L4 AR-1242-1	5.669	4.898	28608484	23663233	489.076	493.809
17) L4 AR-1242-2	5.690	4.917	43367430	33449196	480.598	489.769
18) L4 AR-1242-3	5.753	5.094	25764272	18726327	483.300	486.783
19) L4 AR-1242-4	5.850	5.178	21798483	18331184	484.040	488.202
20) L4 AR-1242-5	6.581	5.702	23472650	21942094	485.668	487.912



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071418.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 18:38
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 02:23:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 02:23:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071419.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 19:27
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 01 07:42:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	98804591	74071948	49.835	52.584
2) SA Decachlor...	10.234	8.848	74699675	45534064	51.588	51.917
Target Compounds						
21) L5 AR-1248-1	5.667	4.897	22002964	19168563	512.425	530.888
22) L5 AR-1248-2	5.938	5.135	30310191	26184570	501.555	534.789
23) L5 AR-1248-3	6.141	5.177	34304891	27346403	513.666	535.307
24) L5 AR-1248-4	6.541	5.349	43159659	31642146	510.096	521.654
25) L5 AR-1248-5	6.580	5.742	40310718	30029077	507.383	515.159



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071419.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 19:27
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 01 07:42:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071420.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 19:43
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 01:28:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 19:27:22 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.813	98111190	74345146	50.931	50.716
2) SA Decachlor...	10.232	8.849	73988546	45066273	50.788	50.509
Target Compounds						
26) L6 AR-1254-1	6.517	5.702	41728109	42206013	509.183	461.771
27) L6 AR-1254-2	6.733	5.850	64122746	35938664	504.601	456.583
28) L6 AR-1254-3	7.097	6.254	64574948	56055056	505.381	473.005
29) L6 AR-1254-4	7.379	6.482	58986526	36616929	501.956	474.524
30) L6 AR-1254-5	7.796	6.900	55244441	47442452	512.803	467.562



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071420.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 19:43
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 01:28:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 19:27:22 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071421.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 20:16
 Operator : YP\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225AR1268

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:47:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.516	3.814	98428428	75098106	49.661	51.289
2) SA Decachlor...	10.234	8.850	131.8E6	79464849	49.365	49.579
Target Compounds						
41) L9 AR-1268-1	8.733	7.722	157.9E6	98709054	493.783	486.283m
42) L9 AR-1268-2	8.827	7.788	132.3E6	84218105	493.704	487.500
43) L9 AR-1268-3	9.058	7.992	112.3E6	69254815	498.379	484.016
44) L9 AR-1268-4	9.477	8.286	48904008	30724632	493.945	483.689
45) L9 AR-1268-5	9.894	8.588	309.0E6	189.8E6	497.455	492.818

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071421.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 20:16
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1268

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:47:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Continuing Calib Date: 05/06/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 09:22 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.67	5.67	5.57	5.77	0.00
Aroclor-1016-2	(2)	5.69	5.69	5.59	5.79	0.00
Aroclor-1016-3	(3)	5.76	5.75	5.65	5.85	-0.01
Aroclor-1016-4	(4)	5.85	5.85	5.75	5.95	0.00
Aroclor-1016-5	(5)	6.15	6.15	6.05	6.25	0.00
Aroclor-1260-1	(1)	7.26	7.27	7.17	7.37	0.01
Aroclor-1260-2	(2)	7.52	7.52	7.42	7.62	0.00
Aroclor-1260-3	(3)	7.88	7.88	7.78	7.98	0.00
Aroclor-1260-4	(4)	8.10	8.10	8.00	8.20	0.00
Aroclor-1260-5	(5)	8.42	8.42	8.32	8.52	0.00
Tetrachloro-m-xylene		4.52	4.52	4.42	4.62	0.00
Decachlorobiphenyl		10.24	10.24	10.14	10.34	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Continuing Calib Date: 05/06/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 09:22 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.89	4.90	4.80	5.00	0.01
Aroclor-1016-2	(2)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-3	(3)	5.09	5.09	4.99	5.19	0.00
Aroclor-1016-4	(4)	5.13	5.14	5.04	5.24	0.01
Aroclor-1016-5	(5)	5.35	5.35	5.25	5.45	0.01
Aroclor-1260-1	(1)	6.38	6.39	6.29	6.49	0.01
Aroclor-1260-2	(2)	6.57	6.57	6.47	6.67	0.00
Aroclor-1260-3	(3)	6.72	6.73	6.63	6.83	0.01
Aroclor-1260-4	(4)	7.19	7.20	7.10	7.30	0.01
Aroclor-1260-5	(5)	7.43	7.44	7.34	7.54	0.01
Tetrachloro-m-xylene		3.81	3.81	3.71	3.91	0.00
Decachlorobiphenyl		8.84	8.85	8.75	8.95	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL01 Date Analyzed: 05/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP071800.D Time Analyzed: 09:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.673	5.571	5.771	439.490	500.000	-12.1
Aroclor-1016-2	5.694	5.593	5.793	454.470	500.000	-9.1
Aroclor-1016-3	5.756	5.654	5.854	444.880	500.000	-11.0
Aroclor-1016-4	5.853	5.752	5.952	443.130	500.000	-11.4
Aroclor-1016-5	6.146	6.045	6.245	456.880	500.000	-8.6
Aroclor-1260-1	7.264	7.165	7.365	480.350	500.000	-3.9
Aroclor-1260-2	7.518	7.418	7.618	472.150	500.000	-5.6
Aroclor-1260-3	7.876	7.777	7.977	456.570	500.000	-8.7
Aroclor-1260-4	8.100	8.002	8.202	439.400	500.000	-12.1
Aroclor-1260-5	8.420	8.322	8.522	461.880	500.000	-7.6
Decachlorobiphenyl	10.236	10.137	10.337	47.570	50.000	-4.9
Tetrachloro-m-xylene	4.518	4.417	4.617	45.670	50.000	-8.7



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL01 Date Analyzed: 05/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP071800.D Time Analyzed: 09:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.893	4.798	4.998	477.790	500.000	-4.4
Aroclor-1016-2	4.912	4.816	5.016	488.710	500.000	-2.3
Aroclor-1016-3	5.089	4.993	5.193	489.730	500.000	-2.1
Aroclor-1016-4	5.131	5.036	5.236	480.630	500.000	-3.9
Aroclor-1016-5	5.345	5.250	5.450	535.400	500.000	7.1
Aroclor-1260-1	6.378	6.285	6.485	495.310	500.000	-0.9
Aroclor-1260-2	6.567	6.473	6.673	486.650	500.000	-2.7
Aroclor-1260-3	6.720	6.627	6.827	527.060	500.000	5.4
Aroclor-1260-4	7.191	7.098	7.298	498.400	500.000	-0.3
Aroclor-1260-5	7.432	7.339	7.539	484.340	500.000	-3.1
Decachlorobiphenyl	8.839	8.749	8.949	49.450	50.000	-1.1
Tetrachloro-m-xylene	3.808	3.712	3.912	48.720	50.000	-2.6

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
 Data File : PP071800.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2025 09:22
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 13:56:09 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.518	3.808	90556452	68635763	45.675	48.725
2) SA Decachlor...	10.236	8.839	68885759	43366937	47.573	49.446
Target Compounds						
3) L1 AR-1016-1	5.673	4.893	29467541	25577920	439.487	477.792
4) L1 AR-1016-2	5.694	4.912	46641449	37319334	454.470	488.705
5) L1 AR-1016-3	5.756	5.089	27425846	20671824	444.885	489.733
6) L1 AR-1016-4	5.853	5.131	22731614	16574922	443.127	480.628
7) L1 AR-1016-5	6.146	5.345	22039352	23796863	456.884	535.395
31) L7 AR-1260-1	7.264	6.378	46125009	35810635	480.346	495.311
32) L7 AR-1260-2	7.518	6.567	67554463	42563479	472.153	486.654
33) L7 AR-1260-3	7.876	6.720	51182938	41315326	456.566	527.065
34) L7 AR-1260-4	8.100	7.191	49592585	31454516	439.398	498.403
35) L7 AR-1260-5	8.420	7.432	104.7E6	73026872	461.885	484.336



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
Data File : PP071800.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2025 09:22
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 13:56:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Continuing Calib Date: 05/06/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 16:30 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.66	5.67	5.57	5.77	0.01
Aroclor-1016-2	(2)	5.68	5.69	5.59	5.79	0.01
Aroclor-1016-3	(3)	5.75	5.75	5.65	5.85	0.00
Aroclor-1016-4	(4)	5.84	5.85	5.75	5.95	0.01
Aroclor-1016-5	(5)	6.14	6.15	6.05	6.25	0.01
Aroclor-1260-1	(1)	7.26	7.27	7.17	7.37	0.01
Aroclor-1260-2	(2)	7.51	7.52	7.42	7.62	0.01
Aroclor-1260-3	(3)	7.87	7.88	7.78	7.98	0.01
Aroclor-1260-4	(4)	8.09	8.10	8.00	8.20	0.01
Aroclor-1260-5	(5)	8.41	8.42	8.32	8.52	0.01
Tetrachloro-m-xylene		4.51	4.52	4.42	4.62	0.01
Decachlorobiphenyl		10.22	10.24	10.14	10.34	0.02



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Continuing Calib Date: 05/06/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 16:30 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.89	4.90	4.80	5.00	0.01
Aroclor-1016-2	(2)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-3	(3)	5.09	5.09	4.99	5.19	0.00
Aroclor-1016-4	(4)	5.13	5.14	5.04	5.24	0.01
Aroclor-1016-5	(5)	5.34	5.35	5.25	5.45	0.01
Aroclor-1260-1	(1)	6.38	6.39	6.29	6.49	0.01
Aroclor-1260-2	(2)	6.56	6.57	6.47	6.67	0.01
Aroclor-1260-3	(3)	6.72	6.73	6.63	6.83	0.01
Aroclor-1260-4	(4)	7.19	7.20	7.10	7.30	0.01
Aroclor-1260-5	(5)	7.43	7.44	7.34	7.54	0.01
Tetrachloro-m-xylene		3.81	3.81	3.71	3.91	0.00
Decachlorobiphenyl		8.84	8.85	8.75	8.95	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL02 Date Analyzed: 05/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP071815.D Time Analyzed: 16:30

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.662	5.571	5.771	449.440	500.000	-10.1
Aroclor-1016-2	5.683	5.593	5.793	446.130	500.000	-10.8
Aroclor-1016-3	5.746	5.654	5.854	445.440	500.000	-10.9
Aroclor-1016-4	5.843	5.752	5.952	447.340	500.000	-10.5
Aroclor-1016-5	6.135	6.045	6.245	449.910	500.000	-10.0
Aroclor-1260-1	7.255	7.165	7.365	476.360	500.000	-4.7
Aroclor-1260-2	7.508	7.418	7.618	470.540	500.000	-5.9
Aroclor-1260-3	7.866	7.777	7.977	459.710	500.000	-8.1
Aroclor-1260-4	8.091	8.002	8.202	441.090	500.000	-11.8
Aroclor-1260-5	8.411	8.322	8.522	459.970	500.000	-8.0
Decachlorobiphenyl	10.224	10.137	10.337	47.460	50.000	-5.1
Tetrachloro-m-xylene	4.508	4.417	4.617	45.250	50.000	-9.5



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL02 Date Analyzed: 05/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP071815.D Time Analyzed: 16:30

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.890	4.798	4.998	490.230	500.000	-2.0
Aroclor-1016-2	4.907	4.816	5.016	469.630	500.000	-6.1
Aroclor-1016-3	5.085	4.993	5.193	482.800	500.000	-3.4
Aroclor-1016-4	5.127	5.036	5.236	472.050	500.000	-5.6
Aroclor-1016-5	5.341	5.250	5.450	513.160	500.000	2.6
Aroclor-1260-1	6.375	6.285	6.485	513.750	500.000	2.8
Aroclor-1260-2	6.563	6.473	6.673	505.100	500.000	1.0
Aroclor-1260-3	6.716	6.627	6.827	532.400	500.000	6.5
Aroclor-1260-4	7.187	7.098	7.298	498.560	500.000	-0.3
Aroclor-1260-5	7.429	7.339	7.539	489.620	500.000	-2.1
Decachlorobiphenyl	8.835	8.749	8.949	50.790	50.000	1.6
Tetrachloro-m-xylene	3.806	3.712	3.912	47.950	50.000	-4.1

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
 Data File : PP071815.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2025 16:30
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/07/2025
 Supervised By :mohammad ahmed 05/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 07 01:40:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.508	3.806	89710355	67551123	45.248	47.955
2) SA Decachlor...	10.224	8.835	68726979	44543238	47.463	50.787
Target Compounds						
3) L1 AR-1016-1	5.662	4.890	30134830	26243550	449.439	490.226m
4) L1 AR-1016-2	5.683	4.907	45785537	35862949	446.130	469.633m
5) L1 AR-1016-3	5.746	5.085	27460141	20379267	445.441	482.802
6) L1 AR-1016-4	5.843	5.127	22947688	16279072	447.339	472.050
7) L1 AR-1016-5	6.135	5.341	21702875	22808461	449.909	513.158
31) L7 AR-1260-1	7.255	6.375	45742505	37144060	476.362	513.755
32) L7 AR-1260-2	7.508	6.563	67323308	44176811	470.538	505.101
33) L7 AR-1260-3	7.866	6.716	51535610	41733369	459.712	532.398
34) L7 AR-1260-4	8.091	7.187	49783747	31464383	441.092	498.559
35) L7 AR-1260-5	8.411	7.429	104.2E6	73824138	459.967	489.624

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
Data File : PP071815.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2025 16:30
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/07/2025
Supervised By :mohammad ahmed 05/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 07 01:40:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Continuing Calib Date: 05/07/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 10:04 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.66	5.67	5.57	5.77	0.01
Aroclor-1016-2	(2)	5.68	5.69	5.59	5.79	0.01
Aroclor-1016-3	(3)	5.74	5.75	5.65	5.85	0.01
Aroclor-1016-4	(4)	5.84	5.85	5.75	5.95	0.01
Aroclor-1016-5	(5)	6.13	6.15	6.05	6.25	0.02
Aroclor-1260-1	(1)	7.25	7.27	7.17	7.37	0.02
Aroclor-1260-2	(2)	7.51	7.52	7.42	7.62	0.01
Aroclor-1260-3	(3)	7.86	7.88	7.78	7.98	0.02
Aroclor-1260-4	(4)	8.09	8.10	8.00	8.20	0.01
Aroclor-1260-5	(5)	8.41	8.42	8.32	8.52	0.01
Tetrachloro-m-xylene		4.51	4.52	4.42	4.62	0.01
Decachlorobiphenyl		10.22	10.24	10.14	10.34	0.02



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Continuing Calib Date: 05/07/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 10:04 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.89	4.90	4.80	5.00	0.01
Aroclor-1016-2	(2)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-3	(3)	5.08	5.09	4.99	5.19	0.01
Aroclor-1016-4	(4)	5.12	5.14	5.04	5.24	0.02
Aroclor-1016-5	(5)	5.34	5.35	5.25	5.45	0.01
Aroclor-1260-1	(1)	6.37	6.39	6.29	6.49	0.02
Aroclor-1260-2	(2)	6.56	6.57	6.47	6.67	0.01
Aroclor-1260-3	(3)	6.71	6.73	6.63	6.83	0.02
Aroclor-1260-4	(4)	7.19	7.20	7.10	7.30	0.02
Aroclor-1260-5	(5)	7.43	7.44	7.34	7.54	0.01
Tetrachloro-m-xylene		3.80	3.81	3.71	3.91	0.01
Decachlorobiphenyl		8.83	8.85	8.75	8.95	0.02



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL03 Date Analyzed: 05/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP071835.D Time Analyzed: 10:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.657	5.571	5.771	447.310	500.000	-10.5
Aroclor-1016-2	5.680	5.593	5.793	491.530	500.000	-1.7
Aroclor-1016-3	5.743	5.654	5.854	486.440	500.000	-2.7
Aroclor-1016-4	5.840	5.752	5.952	490.090	500.000	-2.0
Aroclor-1016-5	6.133	6.045	6.245	477.650	500.000	-4.5
Aroclor-1260-1	7.252	7.165	7.365	428.970	500.000	-14.2
Aroclor-1260-2	7.505	7.418	7.618	436.270	500.000	-12.7
Aroclor-1260-3	7.864	7.777	7.977	467.010	500.000	-6.6
Aroclor-1260-4	8.088	8.002	8.202	429.700	500.000	-14.1
Aroclor-1260-5	8.407	8.322	8.522	442.650	500.000	-11.5
Decachlorobiphenyl	10.220	10.137	10.337	46.940	50.000	-6.1
Tetrachloro-m-xylene	4.506	4.417	4.617	48.110	50.000	-3.8



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL03 Date Analyzed: 05/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP071835.D Time Analyzed: 10:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.887	4.798	4.998	500.100	500.000	0.0
Aroclor-1016-2	4.906	4.816	5.016	524.070	500.000	4.8
Aroclor-1016-3	5.083	4.993	5.193	514.310	500.000	2.9
Aroclor-1016-4	5.124	5.036	5.236	524.310	500.000	4.9
Aroclor-1016-5	5.339	5.250	5.450	532.050	500.000	6.4
Aroclor-1260-1	6.372	6.285	6.485	514.800	500.000	3.0
Aroclor-1260-2	6.561	6.473	6.673	513.620	500.000	2.7
Aroclor-1260-3	6.713	6.627	6.827	507.560	500.000	1.5
Aroclor-1260-4	7.185	7.098	7.298	500.210	500.000	0.0
Aroclor-1260-5	7.426	7.339	7.539	490.350	500.000	-1.9
Decachlorobiphenyl	8.832	8.749	8.949	50.800	50.000	1.6
Tetrachloro-m-xylene	3.803	3.712	3.912	52.820	50.000	5.6

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050725\
 Data File : PP071835.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 May 2025 10:04
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/08/2025
 Supervised By :mohammad ahmed 05/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 08 01:20:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.506	3.803	95392678	74401169	48.114	52.818m
2) SA Decachlor...	10.220	8.832	67967006	44550927	46.938	50.796m
Target Compounds						
3) L1 AR-1016-1	5.657	4.887	29991796	26772190	447.306m	500.101m
4) L1 AR-1016-2	5.680	4.906	50444895	40019634	491.530m	524.066m
5) L1 AR-1016-3	5.743	5.083	29987470	21709194	486.438	514.309
6) L1 AR-1016-4	5.840	5.124	25140695	18081177	490.090	524.306
7) L1 AR-1016-5	6.133	5.339	23041288	23648111	477.655	532.048
31) L7 AR-1260-1	7.252	6.372	41191867	37219315	428.972	514.795
32) L7 AR-1260-2	7.505	6.561	62420029	44921514	436.268	513.615
33) L7 AR-1260-3	7.864	6.713	52353199	39786131	467.005	507.557
34) L7 AR-1260-4	8.088	7.185	48498149	31568294	429.701	500.206
35) L7 AR-1260-5	8.407	7.426	100.3E6	73934114	442.652	490.353

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050725\
Data File : PP071835.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 May 2025 10:04
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/08/2025
Supervised By :mohammad ahmed 05/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 08 01:20:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Continuing Calib Date: 05/07/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 17:35 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.66	5.67	5.57	5.77	0.01
Aroclor-1016-2	(2)	5.69	5.69	5.59	5.79	0.00
Aroclor-1016-3	(3)	5.75	5.75	5.65	5.85	0.00
Aroclor-1016-4	(4)	5.85	5.85	5.75	5.95	0.01
Aroclor-1016-5	(5)	6.14	6.15	6.05	6.25	0.01
Aroclor-1260-1	(1)	7.26	7.27	7.17	7.37	0.01
Aroclor-1260-2	(2)	7.51	7.52	7.42	7.62	0.01
Aroclor-1260-3	(3)	7.87	7.88	7.78	7.98	0.01
Aroclor-1260-4	(4)	8.09	8.10	8.00	8.20	0.01
Aroclor-1260-5	(5)	8.41	8.42	8.32	8.52	0.01
Tetrachloro-m-xylene		4.51	4.52	4.42	4.62	0.01
Decachlorobiphenyl		10.23	10.24	10.14	10.34	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Continuing Calib Date: 05/07/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 17:35 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.89	4.90	4.80	5.00	0.01
Aroclor-1016-2	(2)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-3	(3)	5.08	5.09	4.99	5.19	0.01
Aroclor-1016-4	(4)	5.13	5.14	5.04	5.24	0.01
Aroclor-1016-5	(5)	5.34	5.35	5.25	5.45	0.01
Aroclor-1260-1	(1)	6.37	6.39	6.29	6.49	0.02
Aroclor-1260-2	(2)	6.56	6.57	6.47	6.67	0.01
Aroclor-1260-3	(3)	6.72	6.73	6.63	6.83	0.01
Aroclor-1260-4	(4)	7.19	7.20	7.10	7.30	0.01
Aroclor-1260-5	(5)	7.43	7.44	7.34	7.54	0.01
Tetrachloro-m-xylene		3.81	3.81	3.71	3.91	0.00
Decachlorobiphenyl		8.83	8.85	8.75	8.95	0.02



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL04 Date Analyzed: 05/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP071847.D Time Analyzed: 17:35

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.664	5.571	5.771	490.910	500.000	-1.8
Aroclor-1016-2	5.686	5.593	5.793	486.990	500.000	-2.6
Aroclor-1016-3	5.748	5.654	5.854	484.260	500.000	-3.1
Aroclor-1016-4	5.845	5.752	5.952	490.690	500.000	-1.9
Aroclor-1016-5	6.138	6.045	6.245	482.090	500.000	-3.6
Aroclor-1260-1	7.257	7.165	7.365	470.440	500.000	-5.9
Aroclor-1260-2	7.510	7.418	7.618	478.000	500.000	-4.4
Aroclor-1260-3	7.869	7.777	7.977	472.760	500.000	-5.4
Aroclor-1260-4	8.093	8.002	8.202	436.510	500.000	-12.7
Aroclor-1260-5	8.413	8.322	8.522	460.610	500.000	-7.9
Decachlorobiphenyl	10.225	10.137	10.337	48.570	50.000	-2.9
Tetrachloro-m-xylene	4.510	4.417	4.617	48.870	50.000	-2.3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL04 Date Analyzed: 05/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP071847.D Time Analyzed: 17:35

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.889	4.798	4.998	503.880	500.000	0.8
Aroclor-1016-2	4.908	4.816	5.016	533.120	500.000	6.6
Aroclor-1016-3	5.084	4.993	5.193	516.770	500.000	3.4
Aroclor-1016-4	5.126	5.036	5.236	522.060	500.000	4.4
Aroclor-1016-5	5.341	5.250	5.450	514.370	500.000	2.9
Aroclor-1260-1	6.374	6.285	6.485	506.580	500.000	1.3
Aroclor-1260-2	6.563	6.473	6.673	500.570	500.000	0.1
Aroclor-1260-3	6.716	6.627	6.827	510.520	500.000	2.1
Aroclor-1260-4	7.187	7.098	7.298	519.530	500.000	3.9
Aroclor-1260-5	7.429	7.339	7.539	511.410	500.000	2.3
Decachlorobiphenyl	8.834	8.749	8.949	51.640	50.000	3.3
Tetrachloro-m-xylene	3.805	3.712	3.912	52.680	50.000	5.4

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050725\
 Data File : PP071847.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 May 2025 17:35
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/08/2025
 Supervised By :mohammad ahmed 05/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 08 01:26:05 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.510	3.805	96896447	74207672	48.873	52.680
2) SA Decachlor...	10.225	8.834	70335457	45290832	48.574	51.640
Target Compounds						
3) L1 AR-1016-1	5.664	4.889	32915180	26974495	490.906	503.880m
4) L1 AR-1016-2	5.686	4.908	49978493	40710671	486.985	533.115m
5) L1 AR-1016-3	5.748	5.084	29853329	21812898	484.262	516.766
6) L1 AR-1016-4	5.845	5.126	25171387	18003684	490.688	522.059
7) L1 AR-1016-5	6.138	5.341	23255169	22862171	482.088	514.366
31) L7 AR-1260-1	7.257	6.374	45174019	36625258	470.442	506.579
32) L7 AR-1260-2	7.510	6.563	68391036	43780476	478.000	500.569
33) L7 AR-1260-3	7.869	6.716	52998766	40018387	472.764	510.520
34) L7 AR-1260-4	8.093	7.187	49267095	32787902	436.514	519.531
35) L7 AR-1260-5	8.413	7.429	104.4E6	77108760	460.613	511.408

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050725\
Data File : PP071847.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 May 2025 17:35
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/08/2025
Supervised By :mohammad ahmed 05/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 08 01:26:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Analytical Sequence

Client: CDM Smith

SDG No.: Q1956

Project: Bergen Point Fueling System

Instrument ID: ECD_P

GC Column: ZB-MR1

ID: 0.32 (mm)

Inst. Calib. Date(s): 04/22/2025 04/22/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
1.BLK	1.BLK	04/22/2025	10:13	PP071388.D	10.23	4.51
AR1660ICC1000	AR1660ICC1000	04/22/2025	10:29	PP071389.D	10.23	4.51
AR1660ICC750	AR1660ICC750	04/22/2025	10:45	PP071390.D	10.23	4.51
AR1660ICC500	AR1660ICC500	04/22/2025	11:02	PP071391.D	10.24	4.52
AR1660ICC250	AR1660ICC250	04/22/2025	11:18	PP071392.D	10.23	4.51
AR1660ICC050	AR1660ICC050	04/22/2025	11:34	PP071393.D	10.24	4.52
AR1221ICC500	AR1221ICC500	04/22/2025	11:51	PP071394.D	10.23	4.51
AR1232ICC500	AR1232ICC500	04/22/2025	12:07	PP071395.D	10.23	4.52
AR1242ICC1000	AR1242ICC1000	04/22/2025	12:23	PP071396.D	10.23	4.51
AR1242ICC750	AR1242ICC750	04/22/2025	12:39	PP071397.D	10.24	4.52
AR1242ICC500	AR1242ICC500	04/22/2025	12:56	PP071398.D	10.24	4.51
AR1242ICC250	AR1242ICC250	04/22/2025	13:12	PP071399.D	10.23	4.51
AR1242ICC050	AR1242ICC050	04/22/2025	13:28	PP071400.D	10.24	4.52
AR1248ICC1000	AR1248ICC1000	04/22/2025	13:45	PP071401.D	10.24	4.52
AR1248ICC750	AR1248ICC750	04/22/2025	14:01	PP071402.D	10.24	4.52
AR1248ICC500	AR1248ICC500	04/22/2025	14:17	PP071403.D	10.24	4.51
AR1248ICC250	AR1248ICC250	04/22/2025	14:33	PP071404.D	10.24	4.51
AR1248ICC050	AR1248ICC050	04/22/2025	14:50	PP071405.D	10.24	4.51
AR1254ICC1000	AR1254ICC1000	04/22/2025	15:06	PP071406.D	10.24	4.52
AR1254ICC750	AR1254ICC750	04/22/2025	15:22	PP071407.D	10.24	4.52
AR1254ICC500	AR1254ICC500	04/22/2025	15:38	PP071408.D	10.24	4.52
AR1254ICC250	AR1254ICC250	04/22/2025	15:55	PP071409.D	10.23	4.51
AR1254ICC050	AR1254ICC050	04/22/2025	16:11	PP071410.D	10.24	4.51
AR1262ICC500	AR1262ICC500	04/22/2025	16:27	PP071411.D	10.24	4.52
AR1268ICC1000	AR1268ICC1000	04/22/2025	16:44	PP071412.D	10.24	4.51
AR1268ICC750	AR1268ICC750	04/22/2025	17:00	PP071413.D	10.23	4.52
AR1268ICC500	AR1268ICC500	04/22/2025	17:16	PP071414.D	10.24	4.52
AR1268ICC250	AR1268ICC250	04/22/2025	17:33	PP071415.D	10.24	4.52
AR1268ICC050	AR1268ICC050	04/22/2025	17:49	PP071416.D	10.24	4.51
AR1660CCC500	AR1660CCC500	05/06/2025	09:22	PP071800.D	10.24	4.52
1.BLK	1.BLK	05/06/2025	10:27	PP071804.D	10.22	4.51
PB167877BL	PB167877BL	05/06/2025	13:12	PP071807.D	10.22	4.51
PB167877BS	PB167877BS	05/06/2025	13:28	PP071808.D	10.22	4.51
SB1-3-4	Q1956-01	05/06/2025	13:45	PP071809.D	10.22	4.51
SB2-4-5	Q1956-02	05/06/2025	14:01	PP071810.D	10.22	4.51
SB2-4-5MS	Q1956-03MS	05/06/2025	14:18	PP071811.D	10.22	4.51
SB2-4-5MSD	Q1956-04MSD	05/06/2025	14:35	PP071812.D	10.22	4.51
SB91-3-4	Q1956-06	05/06/2025	14:51	PP071813.D	10.22	4.51
AR1660CCC500	AR1660CCC500	05/06/2025	16:30	PP071815.D	10.22	4.51
1.BLK	1.BLK	05/06/2025	17:51	PP071819.D	10.22	4.51
AR1660CCC500	AR1660CCC500	05/07/2025	10:04	PP071835.D	10.22	4.51
1.BLK	1.BLK	05/07/2025	11:30	PP071839.D	10.24	4.52

Analytical Sequence

PB167890BL	PB167890BL	05/07/2025	14:35	PP071840.D	10.22	4.51
PB167890BS	PB167890BS	05/07/2025	14:52	PP071841.D	10.22	4.51
PB167890BSD	PB167890BSD	05/07/2025	15:08	PP071842.D	10.22	4.51
FB-05022025	Q1956-07	05/07/2025	15:24	PP071843.D	10.22	4.51
AR1660CCC500	AR1660CCC500	05/07/2025	17:35	PP071847.D	10.23	4.51
LBLK	LBLK	05/07/2025	18:40	PP071851.D	10.22	4.51

Analytical Sequence

Client: CDM Smith

SDG No.: Q1956

Project: Bergen Point Fueling System

Instrument ID: ECD_P

GC Column: ZB-MR2

ID: 0.32 (mm)

Inst. Calib. Date(s): 04/22/2025

04/22/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
1.BLK	1.BLK	04/22/2025	10:13	PP071388.D	8.85	3.81
AR1660ICC1000	AR1660ICC1000	04/22/2025	10:29	PP071389.D	8.85	3.81
AR1660ICC750	AR1660ICC750	04/22/2025	10:45	PP071390.D	8.85	3.81
AR1660ICC500	AR1660ICC500	04/22/2025	11:02	PP071391.D	8.85	3.81
AR1660ICC250	AR1660ICC250	04/22/2025	11:18	PP071392.D	8.85	3.81
AR1660ICC050	AR1660ICC050	04/22/2025	11:34	PP071393.D	8.85	3.81
AR1221ICC500	AR1221ICC500	04/22/2025	11:51	PP071394.D	8.85	3.81
AR1232ICC500	AR1232ICC500	04/22/2025	12:07	PP071395.D	8.85	3.81
AR1242ICC1000	AR1242ICC1000	04/22/2025	12:23	PP071396.D	8.85	3.81
AR1242ICC750	AR1242ICC750	04/22/2025	12:39	PP071397.D	8.85	3.81
AR1242ICC500	AR1242ICC500	04/22/2025	12:56	PP071398.D	8.85	3.81
AR1242ICC250	AR1242ICC250	04/22/2025	13:12	PP071399.D	8.85	3.81
AR1242ICC050	AR1242ICC050	04/22/2025	13:28	PP071400.D	8.85	3.81
AR1248ICC1000	AR1248ICC1000	04/22/2025	13:45	PP071401.D	8.85	3.81
AR1248ICC750	AR1248ICC750	04/22/2025	14:01	PP071402.D	8.85	3.81
AR1248ICC500	AR1248ICC500	04/22/2025	14:17	PP071403.D	8.85	3.81
AR1248ICC250	AR1248ICC250	04/22/2025	14:33	PP071404.D	8.85	3.81
AR1248ICC050	AR1248ICC050	04/22/2025	14:50	PP071405.D	8.85	3.81
AR1254ICC1000	AR1254ICC1000	04/22/2025	15:06	PP071406.D	8.85	3.81
AR1254ICC750	AR1254ICC750	04/22/2025	15:22	PP071407.D	8.85	3.81
AR1254ICC500	AR1254ICC500	04/22/2025	15:38	PP071408.D	8.85	3.81
AR1254ICC250	AR1254ICC250	04/22/2025	15:55	PP071409.D	8.85	3.81
AR1254ICC050	AR1254ICC050	04/22/2025	16:11	PP071410.D	8.85	3.81
AR1262ICC500	AR1262ICC500	04/22/2025	16:27	PP071411.D	8.85	3.81
AR1268ICC1000	AR1268ICC1000	04/22/2025	16:44	PP071412.D	8.85	3.81
AR1268ICC750	AR1268ICC750	04/22/2025	17:00	PP071413.D	8.85	3.81
AR1268ICC500	AR1268ICC500	04/22/2025	17:16	PP071414.D	8.85	3.81
AR1268ICC250	AR1268ICC250	04/22/2025	17:33	PP071415.D	8.85	3.81
AR1268ICC050	AR1268ICC050	04/22/2025	17:49	PP071416.D	8.85	3.81
AR1660CCC500	AR1660CCC500	05/06/2025	09:22	PP071800.D	8.84	3.81
1.BLK	1.BLK	05/06/2025	10:27	PP071804.D	8.83	3.81
PB167877BL	PB167877BL	05/06/2025	13:12	PP071807.D	8.83	3.81
PB167877BS	PB167877BS	05/06/2025	13:28	PP071808.D	8.83	3.80
SB1-3-4	Q1956-01	05/06/2025	13:45	PP071809.D	8.83	3.81
SB2-4-5	Q1956-02	05/06/2025	14:01	PP071810.D	8.83	3.81
SB2-4-5MS	Q1956-03MS	05/06/2025	14:18	PP071811.D	8.83	3.81
SB2-4-5MSD	Q1956-04MSD	05/06/2025	14:35	PP071812.D	8.83	3.81
SB91-3-4	Q1956-06	05/06/2025	14:51	PP071813.D	8.83	3.81
AR1660CCC500	AR1660CCC500	05/06/2025	16:30	PP071815.D	8.84	3.81
1.BLK	1.BLK	05/06/2025	17:51	PP071819.D	8.83	3.81
AR1660CCC500	AR1660CCC500	05/07/2025	10:04	PP071835.D	8.83	3.80
1.BLK	1.BLK	05/07/2025	11:30	PP071839.D	8.84	3.81

Analytical Sequence

PB167890BL	PB167890BL	05/07/2025	14:35	PP071840.D	8.83	3.81
PB167890BS	PB167890BS	05/07/2025	14:52	PP071841.D	8.83	3.81
PB167890BSD	PB167890BSD	05/07/2025	15:08	PP071842.D	8.83	3.81
FB-05022025	Q1956-07	05/07/2025	15:24	PP071843.D	8.83	3.81
AR1660CCC500	AR1660CCC500	05/07/2025	17:35	PP071847.D	8.83	3.81
LBLK	LBLK	05/07/2025	18:40	PP071851.D	8.83	3.80

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB167877BS

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Lab Sample ID: PB167877BS Date(s) Analyzed: 05/06/2025 05/06/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP071808.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.659	5.609	5.709	138	145	4.71
COLUMN 1	2	5.681	5.631	5.731	149		
	3	5.744	5.694	5.794	149		
	4	5.842	5.792	5.892	150		
	5	6.134	6.084	6.184	139		
	1	4.887	4.837	4.937	155		
COLUMN 2	2	4.906	4.856	4.956	149	152	
	3	5.083	5.033	5.133	153		
	4	5.125	5.075	5.175	154		
	5	5.339	5.289	5.389	150		
Aroclor-1260	1	7.253	7.203	7.303	149	141	6.85
COLUMN 1	2	7.506	7.456	7.556	152		
	3	7.864	7.814	7.914	131		
	4	8.089	8.039	8.139	135		
	5	8.408	8.358	8.458	137		
	1	6.373	6.323	6.423	156		
COLUMN 2	2	6.561	6.511	6.611	156	151	
	3	6.714	6.664	6.764	161		
	4	7.185	7.135	7.235	143		
	5	7.426	7.376	7.476	141		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SB2-4-5

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Lab Sample ID: Q1956-02 Date(s) Analyzed: 05/06/2025 05/06/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP071810.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD	
			FROM	TO				
Aroclor-1254	1	6.512	6.462	6.562	6.33	5.30	26.23	
COLUMN 1	2	6.727	6.677	6.777	4.52			
	3	7.09	7.04	7.14	5.11			
	4	7.374	7.324	7.424	5.25			
	5	7.788	7.738	7.838	5.51			
	1	5.692	5.642	5.742	6.14	6.90		
2	5.839	5.789	5.889	5.95				
3	6.244	6.194	6.294	6.64				
4	6.472	6.422	6.522	8.56				
5	6.888	6.838	6.938	7.08				

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SB2-4-5MS

Contract: CAMP02

Lab Code: CHEM

Case No.: Q1956

SAS No.: Q1956

SDG NO.: Q1956

Lab Sample ID: Q1956-03MS

Date(s) Analyzed: 05/06/2025

05/06/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP071811.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.661	5.611	5.711	186	177	5.49
COLUMN 1	2	5.683	5.633	5.733	178		
	3	5.745	5.695	5.795	175		
	4	5.842	5.792	5.892	175		
	5	6.135	6.085	6.185	170		
	1	4.887	4.837	4.937	178		
COLUMN 2	2	4.906	4.856	4.956	201	187	
	3	5.083	5.033	5.133	187		
	4	5.125	5.075	5.175	184		
	5	5.339	5.289	5.389	182		
Aroclor-1254	1	6.508	6.458	6.558	84.1	89.5	3.64
COLUMN 1	2	6.725	6.675	6.775	62.3		
	3	7.089	7.039	7.139	30.0		
	4	7.372	7.322	7.422	24.1		
	5	7.787	7.737	7.837	247		
	1	5.691	5.641	5.741	71.0		
COLUMN 2	2	5.839	5.789	5.889	76.8	86.3	
	3	6.257	6.207	6.307	90.1		
	4	6.471	6.421	6.521	17.6		
	5	6.888	6.838	6.938	176		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

Aroclor-1260	1	7.253	7.203	7.303	178			
	2	7.507	7.457	7.557	178			
	3	7.865	7.815	7.915	155			
	4	8.088	8.038	8.138	164			
	5	8.408	8.358	8.458	161			
COLUMN 1						167		
	1	6.373	6.323	6.423	194			
	2	6.561	6.511	6.611	190			
	3	6.714	6.664	6.764	191			
	4	7.184	7.134	7.234	170			
COLUMN 2	5	7.426	7.376	7.476	166	182		8.6

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SB2-4-5MSD

Contract: CAMP02

Lab Code: CHEM

Case No.: Q1956

SAS No.: Q1956

SDG NO.: Q1956

Lab Sample ID: Q1956-04MSD

Date(s) Analyzed: 05/06/2025

05/06/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP071812.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.658	5.608	5.708	158	174	3.94
COLUMN 1	2	5.68	5.63	5.73	180		
	3	5.742	5.692	5.792	179		
	4	5.841	5.791	5.891	179		
	5	6.133	6.083	6.183	173		
	1	4.889	4.839	4.939	180		
COLUMN 2	2	4.906	4.856	4.956	190	181	
	3	5.084	5.034	5.134	178		
	4	5.126	5.076	5.176	176		
	5	5.34	5.29	5.39	180		
Aroclor-1254	1	6.506	6.456	6.556	87.2	91.5	4.93
COLUMN 1	2	6.724	6.674	6.774	63.1		
	3	7.088	7.038	7.138	30.6		
	4	7.371	7.321	7.421	25.5		
	5	7.786	7.736	7.836	251		
	1	5.692	5.642	5.742	69.9		
COLUMN 2	2	5.84	5.79	5.89	77.1	87.1	
	3	6.258	6.208	6.308	90.6		
	4	6.472	6.422	6.522	16.9		
	5	6.888	6.838	6.938	181		



Aroclor-1260	1	7.252	7.202	7.302	183		
	2	7.506	7.456	7.556	182		
	3	7.864	7.814	7.914	158		
	4	8.088	8.038	8.138	169		
	5	8.408	8.358	8.458	165		
COLUMN 1						172	
	1	6.374	6.324	6.424	196		
	2	6.562	6.512	6.612	192		
	3	6.715	6.665	6.765	199		
	4	7.185	7.135	7.235	174		
COLUMN 2	5	7.427	7.377	7.477	171		186

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB167890BS

Contract: CAMP02

Lab Code: CHEM

Case No.: Q1956

SAS No.: Q1956

SDG NO.: Q1956

Lab Sample ID: PB167890BS

Date(s) Analyzed: 05/07/2025

05/07/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP071841.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.663	5.613	5.713	4.50	4.50	2.2
COLUMN 1	2	5.685	5.635	5.735	4.62		
	3	5.747	5.697	5.797	4.55		
	4	5.844	5.794	5.894	4.61		
	5	6.137	6.087	6.187	4.46		
	1	4.889	4.839	4.939	4.63		
COLUMN 2	2	4.908	4.858	4.958	4.58	4.60	
	3	5.084	5.034	5.134	4.57		
	4	5.126	5.076	5.176	4.56		
	5	5.341	5.291	5.391	4.48		
Aroclor-1260	1	7.256	7.206	7.306	4.58	4.20	11.24
COLUMN 1	2	7.51	7.46	7.56	4.42		
	3	7.868	7.818	7.918	3.81		
	4	8.094	8.044	8.144	4.27		
	5	8.412	8.362	8.462	4.01		
	1	6.374	6.324	6.424	4.73		
COLUMN 2	2	6.562	6.512	6.612	4.82	4.70	
	3	6.715	6.665	6.765	5.06		
	4	7.186	7.136	7.236	4.60		
	5	7.428	7.378	7.478	4.41		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB167890BSD

Contract: CAMP02

Lab Code: CHEM

Case No.: Q1956

SAS No.: Q1956

SDG NO.: Q1956

Lab Sample ID: PB167890BSD

Date(s) Analyzed: 05/07/2025

05/07/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP071842.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.662	5.612	5.712	4.60	4.60	2.15
COLUMN 1	2	5.683	5.633	5.733	4.58		
	3	5.746	5.696	5.796	4.62		
	4	5.843	5.793	5.893	4.64		
	5	6.136	6.086	6.186	4.43		
	1	4.888	4.838	4.938	4.95		
COLUMN 2	2	4.906	4.856	4.956	4.70	4.70	
	3	5.084	5.034	5.134	4.63		
	4	5.125	5.075	5.175	4.64		
	5	5.34	5.29	5.39	4.55		
Aroclor-1260	1	7.255	7.205	7.305	4.60	4.10	13.64
COLUMN 1	2	7.509	7.459	7.559	4.47		
	3	7.867	7.817	7.917	3.82		
	4	8.091	8.041	8.141	3.88		
	5	8.411	8.361	8.461	3.93		
	1	6.374	6.324	6.424	5.04		
COLUMN 2	2	6.562	6.512	6.612	5.04	4.70	
	3	6.715	6.665	6.765	5.01		
	4	7.186	7.136	7.236	4.41		
	5	7.428	7.378	7.478	4.24		



QC SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Bergen Point Fueling System	Date Received:	
Client Sample ID:	PB167877BL	SDG No.:	Q1956
Lab Sample ID:	PB167877BL	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	100 Decanted:
Sample Wt/Vol:	30.02 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071807.D	1	05/06/25 08:55	05/06/25 13:12	PB167877

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	17.0	U	3.90	17.0	ug/kg
11104-28-2	Aroclor-1221	17.0	U	4.00	17.0	ug/kg
11141-16-5	Aroclor-1232	17.0	U	3.70	17.0	ug/kg
53469-21-9	Aroclor-1242	17.0	U	4.00	17.0	ug/kg
12672-29-6	Aroclor-1248	17.0	U	5.90	17.0	ug/kg
11097-69-1	Aroclor-1254	17.0	U	3.20	17.0	ug/kg
37324-23-5	Aroclor-1262	17.0	U	5.00	17.0	ug/kg
11100-14-4	Aroclor-1268	17.0	U	3.60	17.0	ug/kg
11096-82-5	Aroclor-1260	17.0	U	3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.9		32 - 144	115%	SPK: 20
2051-24-3	Decachlorobiphenyl	24.6		32 - 175	123%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
 Data File : PP071807.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2025 13:12
 Operator : YP\AJ
 Sample : PB167877BL
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB167877BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 14:00:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.509	3.806	40926288	32294022	20.642	22.926
2) SA Decachlor...	10.221	8.833	31663775	21582324	21.867	24.608

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
Data File : PP071807.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2025 13:12
Operator : YP\AJ
Sample : PB167877BL
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB167877BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 14:00:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.509 min
Delta R.T.: -0.008 min
Response: 40926288
Conc: 20.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB167877BL




#1 Tetrachloro-m-xylene

R.T.: 3.806 min
Delta R.T.: -0.006 min
Response: 32294022
Conc: 22.93 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.221 min
Delta R.T.: -0.017 min
Response: 31663775
Conc: 21.87 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.833 min
Delta R.T.: -0.016 min
Response: 21582324
Conc: 24.61 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Bergen Point Fueling System	Date Received:	
Client Sample ID:	PB167890BL	SDG No.:	Q1956
Lab Sample ID:	PB167890BL	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071840.D	1	05/07/25 09:15	05/07/25 14:35	PB167890

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.50	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.094	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.50	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.9		16 - 158	110%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.4		10 - 173	117%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050725\
Data File : PP071840.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 May 2025 14:35
Operator : YP\AJ
Sample : PB167890BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB167890BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 08 01:22:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.506	3.805	40310831	30910520	20.332	21.944
2) SA Decachlor...	10.220	8.833	29329511	20571032	20.255	23.455

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050725\
Data File : PP071840.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 May 2025 14:35
Operator : YP\AJ
Sample : PB167890BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB167890BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 08 01:22:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.506 min
Delta R.T.: -0.011 min
Response: 40310831
Conc: 20.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB167890BL




#1 Tetrachloro-m-xylene

R.T.: 3.805 min
Delta R.T.: -0.007 min
Response: 30910520
Conc: 21.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.220 min
Delta R.T.: -0.017 min
Response: 29329511
Conc: 20.26 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.833 min
Delta R.T.: -0.016 min
Response: 20571032
Conc: 23.45 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	04/22/25
Project:	Bergen Point Fueling System	Date Received:	04/22/25
Client Sample ID:	PIBLK-PP071388.D	SDG No.:	Q1956
Lab Sample ID:	I.BLK-PP071388.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071388.D	1		04/22/25	PP042225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.50	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.50	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.1		60 - 140	85%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.6		60 - 140	88%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071388.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 10:13
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 19:33:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 17:08:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.512	3.811	33865777	24238749	17.081	17.207
2) SA Decachlor...	10.233	8.847	25533987	15462668	17.634	17.630

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071388.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 10:13
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 19:33:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 17:08:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.512 min
Delta R.T.: -0.005 min
Response: 33865777
Conc: 17.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.811 min
Delta R.T.: -0.001 min
Response: 24238749
Conc: 17.21 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.233 min
Delta R.T.: -0.005 min
Response: 25533987
Conc: 17.63 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.847 min
Delta R.T.: -0.002 min
Response: 15462668
Conc: 17.63 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/06/25
Project:	Bergen Point Fueling System	Date Received:	05/06/25
Client Sample ID:	PIBLK-PP071804.D	SDG No.:	Q1956
Lab Sample ID:	I.BLK-PP071804.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071804.D	1		05/06/25	PP050625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.50	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.50	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.1		60 - 140	86%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.0		60 - 140	95%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
Data File : PP071804.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2025 10:27
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 13:58:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.508	3.806	33904961	26792082	17.101	19.020
2) SA Decachlor...	10.221	8.832	27546038	17705422	19.023	20.187

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
Data File : PP071804.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2025 10:27
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 13:58:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.508 min
Delta R.T.: -0.009 min
Response: 33904961
Conc: 17.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.806 min
Delta R.T.: -0.007 min
Response: 26792082
Conc: 19.02 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.221 min
Delta R.T.: -0.017 min
Response: 27546038
Conc: 19.02 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.832 min
Delta R.T.: -0.017 min
Response: 17705422
Conc: 20.19 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/06/25
Project:	Bergen Point Fueling System	Date Received:	05/06/25
Client Sample ID:	PIBLK-PP071819.D	SDG No.:	Q1956
Lab Sample ID:	I.BLK-PP071819.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071819.D	1		05/06/25	PP050625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.50	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.50	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.0		60 - 140	85%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.7		60 - 140	93%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
 Data File : PP071819.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2025 17:51
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 07 01:42:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.508	3.806	33669696	27573594	16.982	19.575
2) SA Decachlor...	10.221	8.834	27044346	19170944	18.677	21.858

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
Data File : PP071819.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2025 17:51
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 07 01:42:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.508 min
Delta R.T.: -0.009 min
Response: 33669696
Conc: 16.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.806 min
Delta R.T.: -0.006 min
Response: 27573594
Conc: 19.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.221 min
Delta R.T.: -0.017 min
Response: 27044346
Conc: 18.68 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.834 min
Delta R.T.: -0.015 min
Response: 19170944
Conc: 21.86 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/07/25
Project:	Bergen Point Fueling System	Date Received:	05/07/25
Client Sample ID:	PIBLK-PP071839.D	SDG No.:	Q1956
Lab Sample ID:	I.BLK-PP071839.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071839.D	1		05/07/25	pp050725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.50	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.50	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.3		60 - 140	81%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.5		60 - 140	83%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050725\
 Data File : PP071839.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 May 2025 11:30
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 08 01:22:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.520	3.808	32227816	25543475	16.255	18.133
2) SA Decachlor...	10.238	8.840	23956093	16807575	16.544	19.164

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050725\
Data File : PP071839.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 May 2025 11:30
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 08 01:22:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.520 min
Delta R.T.: 0.003 min
Response: 32227816
Conc: 16.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.808 min
Delta R.T.: -0.005 min
Response: 25543475
Conc: 18.13 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.238 min
Delta R.T.: 0.000 min
Response: 23956093
Conc: 16.54 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.840 min
Delta R.T.: -0.009 min
Response: 16807575
Conc: 19.16 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/07/25
Project:	Bergen Point Fueling System	Date Received:	05/07/25
Client Sample ID:	PIBLK-PP071851.D	SDG No.:	Q1956
Lab Sample ID:	I.BLK-PP071851.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071851.D	1		05/07/25	pp050725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.50	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.50	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.2		60 - 140	81%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.6		60 - 140	83%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050725\
Data File : PP071851.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 May 2025 18:40
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 08 01:28:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.507	3.804	32132630	25290401	16.207	17.954
2) SA Decachlor...	10.221	8.833	23986042	16784316	16.565	19.137

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050725\
Data File : PP071851.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 May 2025 18:40
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 08 01:28:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.507 min
Delta R.T.: -0.010 min
Response: 32132630
Conc: 16.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.804 min
Delta R.T.: -0.008 min
Response: 25290401
Conc: 17.95 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.221 min
Delta R.T.: -0.017 min
Response: 23986042
Conc: 16.56 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.833 min
Delta R.T.: -0.016 min
Response: 16784316
Conc: 19.14 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Bergen Point Fueling System	Date Received:	
Client Sample ID:	PB167877BS	SDG No.:	Q1956
Lab Sample ID:	PB167877BS	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	100 Decanted:
Sample Wt/Vol:	30.01 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071808.D	1	05/06/25 08:55	05/06/25 13:28	PB167877

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	152		3.90	17.0	ug/kg
11104-28-2	Aroclor-1221	17.0	U	4.00	17.0	ug/kg
11141-16-5	Aroclor-1232	17.0	U	3.70	17.0	ug/kg
53469-21-9	Aroclor-1242	17.0	U	4.00	17.0	ug/kg
12672-29-6	Aroclor-1248	17.0	U	5.90	17.0	ug/kg
11097-69-1	Aroclor-1254	17.0	U	3.20	17.0	ug/kg
37324-23-5	Aroclor-1262	17.0	U	5.00	17.0	ug/kg
11100-14-4	Aroclor-1268	17.0	U	3.60	17.0	ug/kg
11096-82-5	Aroclor-1260	151		3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.4		32 - 144	112%	SPK: 20
2051-24-3	Decachlorobiphenyl	24.1		32 - 175	121%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
 Data File : PP071808.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2025 13:28
 Operator : YP\AJ
 Sample : PB167877BS
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

PB167877BS

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 05/07/2025

Supervised By :mohammad ahmed 05/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 14:01:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.507	3.804	41083017	31478636	20.721	22.347
2) SA Decachlor...	10.217	8.832	32122920	21142530	22.184	24.106
Target Compounds						
3) L1 AR-1016-1	5.659	4.887	27743673	24977686	413.777m	466.580m
4) L1 AR-1016-2	5.681	4.906	45978834	34104582	448.013	446.607m
5) L1 AR-1016-3	5.744	5.083	27474734	19418113	445.678	460.031
6) L1 AR-1016-4	5.842	5.125	23115403	15903969	450.609	461.173
7) L1 AR-1016-5	6.134	5.339	20127859	20001437	417.258	450.004
31) L7 AR-1260-1	7.253	6.373	42992895	33918552	447.728	469.141
32) L7 AR-1260-2	7.506	6.561	65305853	40913071	456.437	467.784
33) L7 AR-1260-3	7.864	6.714	44095757	37907714	393.347	483.593
34) L7 AR-1260-4	8.089	7.185	45571430	27079804	403.770	429.085
35) L7 AR-1260-5	8.408	7.426	93279542	63793477	411.697	423.097

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
Data File : PP071808.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2025 13:28
Operator : YP\AJ
Sample : PB167877BS
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

PB167877BS

Manual Integrations



APPROVED

Reviewed By :Yogesh Patel 05/07/2025

Supervised By :mohammad ahmed 05/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 14:01:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Bergen Point Fueling System	Date Received:	
Client Sample ID:	PB167890BS	SDG No.:	Q1956
Lab Sample ID:	PB167890BS	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071841.D	1	05/07/25 09:15	05/07/25 14:52	PB167890

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	4.60		0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.094	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	4.70		0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.5		16 - 158	102%	SPK: 20
2051-24-3	Decachlorobiphenyl	22.1		10 - 173	110%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050725\
 Data File : PP071841.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 May 2025 14:52
 Operator : YP\AJ
 Sample : PB167890BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

PB167890BS

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 05/08/2025

Supervised By :mohammad ahmed 05/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 08 01:23:36 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.510	3.805	40504649	28847903	20.430	20.479
2)	SA Decachlor...	10.224	8.833	29788514	19376098	20.572	22.092
Target Compounds							
3)	L1 AR-1016-1	5.663	4.889	30167577	24811837	449.928	463.482m
4)	L1 AR-1016-2	5.685	4.908	47457725	34999303	462.423	458.324m
5)	L1 AR-1016-3	5.747	5.084	28040076	19281923	454.848	456.805
6)	L1 AR-1016-4	5.844	5.126	23662826	15722780	461.280	455.919
7)	L1 AR-1016-5	6.137	5.341	21492130	19920356	445.540	448.179
31)	L7 AR-1260-1	7.256	6.374	43954950	34162221	457.747	472.512
32)	L7 AR-1260-2	7.510	6.562	63270923	42176193	442.215	482.226
33)	L7 AR-1260-3	7.868	6.715	42658658	39656951	380.527	505.909 #
34)	L7 AR-1260-4	8.094	7.186	48216668	29036797	427.207	460.094
35)	L7 AR-1260-5	8.412	7.428	90877779	66511957	401.097	441.127

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050725\
Data File : PP071841.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 May 2025 14:52
Operator : YP\AJ
Sample : PB167890BS
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
PB167890BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/08/2025
Supervised By :mohammad ahmed 05/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 08 01:23:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	CDM Smith		Date Collected:		
Project:	Bergen Point Fueling System		Date Received:		
Client Sample ID:	PB167890BSD		SDG No.:	Q1956	
Lab Sample ID:	PB167890BSD		Matrix:	WATER	
Analytical Method:	SW8082A		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071842.D	1	05/07/25 09:15	05/07/25 15:08	PB167890

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	4.70		0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.094	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	4.70		0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.7		16 - 158	104%	SPK: 20
2051-24-3	Decachlorobiphenyl	22.1		10 - 173	111%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050725\
 Data File : PP071842.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 May 2025 15:08
 Operator : YP\AJ
 Sample : PB167890BSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

PB167890BSD

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 05/08/2025

Supervised By :mohammad ahmed 05/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 08 01:24:05 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.508	3.805	39127722	29213230	19.735	20.739
2) SA Decachlor...	10.224	8.834	28686821	19380408	19.811	22.097
Target Compounds						
3) L1 AR-1016-1	5.662	4.888	30858865	26504037	460.238	495.092m
4) L1 AR-1016-2	5.683	4.906	46982706	35855262	457.795	469.533m
5) L1 AR-1016-3	5.746	5.084	28454802	19557819	461.576	463.341
6) L1 AR-1016-4	5.843	5.125	23820985	16012833	464.363	464.329
7) L1 AR-1016-5	6.136	5.340	21373035	20240441	443.071	455.381
31) L7 AR-1260-1	7.255	6.374	44205655	36457055	460.358	504.252
32) L7 AR-1260-2	7.509	6.562	64011334	44041060	447.390	503.548
33) L7 AR-1260-3	7.867	6.715	42818768	39241773	381.955	500.612 #
34) L7 AR-1260-4	8.091	7.186	43783148	27808343	387.926	440.629
35) L7 AR-1260-5	8.411	7.428	89042379	63947717	392.996	424.120

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050725\
Data File : PP071842.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 May 2025 15:08
Operator : YP\AJ
Sample : PB167890BSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

PB167890BSD

Manual Integrations



APPROVED

Reviewed By :Yogesh Patel 05/08/2025

Supervised By :mohammad ahmed 05/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 08 01:24:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	CDM Smith	Date Collected:	05/02/25
Project:	Bergen Point Fueling System	Date Received:	05/02/25
Client Sample ID:	SB2-4-5MS	SDG No.:	Q1956
Lab Sample ID:	Q1956-03MS	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	93.5
Sample Wt/Vol:	30.02	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071811.D	1	05/06/25 08:55	05/06/25 14:18	PB167877

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	187		4.20	18.2	ug/kg
11104-28-2	Aroclor-1221	18.2	U	4.30	18.2	ug/kg
11141-16-5	Aroclor-1232	18.2	U	4.00	18.2	ug/kg
53469-21-9	Aroclor-1242	18.2	U	4.30	18.2	ug/kg
12672-29-6	Aroclor-1248	18.2	U	6.30	18.2	ug/kg
11097-69-1	Aroclor-1254	89.5		3.40	18.2	ug/kg
37324-23-5	Aroclor-1262	18.2	U	5.40	18.2	ug/kg
11100-14-4	Aroclor-1268	18.2	U	3.80	18.2	ug/kg
11096-82-5	Aroclor-1260	182		3.50	18.2	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.7		32 - 144	119%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.3		32 - 175	117%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
 Data File : PP071811.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2025 14:18
 Operator : YP\AJ
 Sample : Q1956-03MS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

SB2-4-5MS

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 05/07/2025

Supervised By :mohammad ahmed 05/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 15:32:49 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.508	3.805	44852510	33393400	22.623	23.706
2) SA Decachlor...	10.219	8.832	30480881	20464784	21.050	23.334
Target Compounds						
3) L1 AR-1016-1	5.661	4.887	34924488	26810529	520.874	500.817m
4) L1 AR-1016-2	5.683	4.906	51322587	43180752	500.082	565.462m
5) L1 AR-1016-3	5.745	5.083	30357315	22155971	492.437	524.894
6) L1 AR-1016-4	5.842	5.125	25240493	17817815	492.035	516.669
7) L1 AR-1016-5	6.135	5.339	23004153	22710417	476.885	510.952
26) L6 AR-1254-1	6.508	5.691	19351690	18207340	236.137m	199.204
27) L6 AR-1254-2	6.725	5.839	22231642	16964623	174.947	215.527
28) L6 AR-1254-3	7.089	6.257	10753327	29967520	84.158	252.873 #
29) L6 AR-1254-4	7.372	6.471	7965209	3802603	67.781	49.278 #
30) L6 AR-1254-5	7.787	6.888	74613029	50194325	692.591	494.683 #
31) L7 AR-1260-1	7.253	6.373	48061626	39340981	500.514	544.141
32) L7 AR-1260-2	7.507	6.561	71634746	46614540	500.671	532.973
33) L7 AR-1260-3	7.865	6.714	48743183	42006702	434.803	535.885
34) L7 AR-1260-4	8.088	7.184	52086624	30057831	461.496	476.272
35) L7 AR-1260-5	8.408	7.426	102.2E6	70108895	451.206	464.983

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
Data File : PP071811.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2025 14:18
Operator : YP\AJ
Sample : Q1956-03MS
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

SB2-4-5MS

Manual Integrations



APPROVED

Reviewed By :Yogesh Patel 05/07/2025

Supervised By :mohammad ahmed 05/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 15:32:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	CDM Smith	Date Collected:	05/02/25
Project:	Bergen Point Fueling System	Date Received:	05/02/25
Client Sample ID:	SB2-4-5MSD	SDG No.:	Q1956
Lab Sample ID:	Q1956-04MSD	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	93.5
Sample Wt/Vol:	30.03	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071812.D	1	05/06/25 08:55	05/06/25 14:35	PB167877

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	181		4.20	18.2	ug/kg
11104-28-2	Aroclor-1221	18.2	U	4.30	18.2	ug/kg
11141-16-5	Aroclor-1232	18.2	U	4.00	18.2	ug/kg
53469-21-9	Aroclor-1242	18.2	U	4.30	18.2	ug/kg
12672-29-6	Aroclor-1248	18.2	U	6.30	18.2	ug/kg
11097-69-1	Aroclor-1254	91.5		3.40	18.2	ug/kg
37324-23-5	Aroclor-1262	18.2	U	5.40	18.2	ug/kg
11100-14-4	Aroclor-1268	18.2	U	3.80	18.2	ug/kg
11096-82-5	Aroclor-1260	186		3.50	18.2	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.8		32 - 144	109%	SPK: 20
2051-24-3	Decachlorobiphenyl	22.7		32 - 175	114%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
 Data File : PP071812.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2025 14:35
 Operator : YP\AJ
 Sample : Q1956-04MSD
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

SB2-4-5MSD

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 05/07/2025

Supervised By :mohammad ahmed 05/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 07 06:56:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.507	3.805	43240440	29821647	21.810	21.171
2) SA Decachlor...	10.218	8.833	30392702	19928777	20.989	22.722
Target Compounds						
3) L1 AR-1016-1	5.658	4.889	29758829	27021469	443.832m	504.757m
4) L1 AR-1016-2	5.680	4.906	51878618	40829419	505.500m	534.670m
5) L1 AR-1016-3	5.742	5.084	31001988	21155623	502.894m	501.194
6) L1 AR-1016-4	5.841	5.126	25780735	17088766	502.567	495.529
7) L1 AR-1016-5	6.133	5.340	23448910	22466661	486.105	505.468
26) L6 AR-1254-1	6.506	5.692	20061334	17950229	244.796m	196.391
27) L6 AR-1254-2	6.724	5.840	22502096	17033634	177.076	216.404
28) L6 AR-1254-3	7.088	6.258	10967569	30142868	85.835	254.352 #
29) L6 AR-1254-4	7.371	6.472	8406983	3651990	71.541	47.327 #
30) L6 AR-1254-5	7.786	6.888	75952609	51559694	705.026	508.139 #
31) L7 AR-1260-1	7.252	6.374	49230273	39695561	512.684	549.045
32) L7 AR-1260-2	7.506	6.562	73294579	47129134	512.272	538.856
33) L7 AR-1260-3	7.864	6.715	49851490	43820834	444.689	559.028 #
34) L7 AR-1260-4	8.088	7.185	53629481	30784044	475.166	487.779
35) L7 AR-1260-5	8.408	7.427	105.2E6	72544832	464.165	481.139

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050625\
Data File : PP071812.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2025 14:35
Operator : YP\AJ
Sample : Q1956-04MSD
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

SB2-4-5MSD

Manual Integrations



APPROVED

Reviewed By :Yogesh Patel 05/07/2025

Supervised By :mohammad ahmed 05/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 07 06:56:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integration Report

Sequence:	PP042225	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC050	PP071393.D	AR-1260-3	yogesh	4/23/2025 7:32:38 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1660ICC050	PP071393.D	AR-1260-4	yogesh	4/23/2025 7:32:38 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1660ICC050	PP071393.D	AR-1260-5	yogesh	4/23/2025 7:32:38 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1254ICC750	PP071407.D	AR-1254-5	yogesh	4/23/2025 7:32:40 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1254ICC050	PP071410.D	Tetrachloro-m-xylene	yogesh	4/23/2025 7:32:41 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC1000	PP071412.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:43 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC750	PP071413.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:44 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC500	PP071414.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:46 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC250	PP071415.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:48 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC050	PP071416.D	AR-1268-1	yogesh	4/23/2025 7:32:50 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC050	PP071416.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:50 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC050	PP071416.D	AR-1268-2	yogesh	4/23/2025 7:32:50 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICV500	PP071421.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:52 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

PP042225

Instrument

ECD_p

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:	PP050625	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242CCC500	PP071801.D	AR-1242-1 #2	yogesh	5/7/2025 7:42:55 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
AR1242CCC500	PP071801.D	AR-1242-2 #2	yogesh	5/7/2025 7:42:55 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
AR1248CCC500	PP071802.D	AR-1248-1 #2	yogesh	5/7/2025 7:42:58 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
AR1254CCC500	PP071803.D	AR-1254-5	yogesh	5/7/2025 7:43:00 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
PB167877BS	PP071808.D	AR-1016-2 #2	yogesh	5/7/2025 7:43:02 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
Q1956-03MS	PP071811.D	AR-1016-1 #2	yogesh	5/7/2025 7:43:04 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
Q1956-03MS	PP071811.D	AR-1016-2 #2	yogesh	5/7/2025 7:43:04 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
Q1956-03MS	PP071811.D	AR-1254-1	yogesh	5/7/2025 7:43:04 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
Q1956-04MSD	PP071812.D	AR-1016-1	yogesh	5/7/2025 7:43:06 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
Q1956-04MSD	PP071812.D	AR-1016-1 #2	yogesh	5/7/2025 7:43:06 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
Q1956-04MSD	PP071812.D	AR-1016-2	yogesh	5/7/2025 7:43:06 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
Q1956-04MSD	PP071812.D	AR-1016-2 #2	yogesh	5/7/2025 7:43:06 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
Q1956-04MSD	PP071812.D	AR-1016-3	yogesh	5/7/2025 7:43:06 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software

Manual Integration Report

Sequence:

PP050625

Instrument

ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q1956-04MSD	PP071812.D	AR-1254-1	yogesh	5/7/2025 7:43:06 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
AR1660CCC500	PP071815.D	AR-1016-1 #2	yogesh	5/7/2025 7:43:09 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
AR1660CCC500	PP071815.D	AR-1016-2 #2	yogesh	5/7/2025 7:43:09 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
AR1660CCC500	PP071829.D	AR-1016-1 #2	yogesh	5/7/2025 7:43:16 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
AR1660CCC500	PP071829.D	AR-1016-2 #2	yogesh	5/7/2025 7:43:16 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
AR1242CCC500	PP071830.D	AR-1242-1 #2	yogesh	5/7/2025 7:43:18 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
AR1242CCC500	PP071830.D	AR-1242-2 #2	yogesh	5/7/2025 7:43:18 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
AR1248CCC500	PP071831.D	AR-1248-1 #2	yogesh	5/7/2025 7:43:20 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software
AR1254CCC500	PP071832.D	AR-1254-5	yogesh	5/7/2025 7:43:22 AM	mohammad	5/8/2025 2:15:08	Peak Integrated by Software

Manual Integration Report

Sequence:	pp050725	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PP071835.D	AR-1016-1	yogesh	5/8/2025 7:39:36 AM	 	 	Peak Integrated by Software
AR1660CCC500	PP071835.D	AR-1016-1 #2	yogesh	5/8/2025 7:39:36 AM	 	 	Peak Integrated by Software
AR1660CCC500	PP071835.D	AR-1016-2	yogesh	5/8/2025 7:39:36 AM	 	 	Peak Integrated by Software
AR1660CCC500	PP071835.D	AR-1016-2 #2	yogesh	5/8/2025 7:39:36 AM	 	 	Peak Integrated by Software
AR1660CCC500	PP071835.D	Decachlorobiphenyl #2	yogesh	5/8/2025 7:39:36 AM	 	 	Peak Integrated by Software
AR1660CCC500	PP071835.D	Tetrachloro-m-xylene #2	yogesh	5/8/2025 7:39:36 AM	 	 	Peak Integrated by Software
AR1242CCC500	PP071836.D	AR-1242-1 #2	yogesh	5/8/2025 7:39:38 AM	 	 	Peak Integrated by Software
AR1248CCC500	PP071837.D	AR-1248-1 #2	yogesh	5/8/2025 7:39:40 AM	 	 	Peak Integrated by Software
AR1248CCC500	PP071837.D	Decachlorobiphenyl #2	yogesh	5/8/2025 7:39:40 AM	 	 	Peak Integrated by Software
PB167890BS	PP071841.D	AR-1016-1 #2	yogesh	5/8/2025 7:39:41 AM	 	 	Peak Integrated by Software
PB167890BS	PP071841.D	AR-1016-2 #2	yogesh	5/8/2025 7:39:41 AM	 	 	Peak Integrated by Software
PB167890BSD	PP071842.D	AR-1016-1 #2	yogesh	5/8/2025 7:39:43 AM	 	 	Peak Integrated by Software
PB167890BSD	PP071842.D	AR-1016-2 #2	yogesh	5/8/2025 7:39:43 AM	 	 	Peak Integrated by Software

Manual Integration Report

Sequence:

pp050725

Instrument

ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PP071847.D	AR-1016-1 #2	yogesh	5/8/2025 7:39:46 AM	 	 	Peak Integrated by Software
AR1660CCC500	PP071847.D	AR-1016-2 #2	yogesh	5/8/2025 7:39:46 AM	 	 	Peak Integrated by Software
AR1242CCC500	PP071848.D	AR-1242-1 #2	yogesh	5/8/2025 7:39:48 AM	 	 	Peak Integrated by Software
AR1242CCC500	PP071848.D	AR-1242-2 #2	yogesh	5/8/2025 7:39:48 AM	 	 	Peak Integrated by Software
AR1248CCC500	PP071849.D	AR-1248-1 #2	yogesh	5/8/2025 7:39:49 AM	 	 	Peak Integrated by Software

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP071387.D	22 Apr 2025 09:56	YP\AJ	Ok
2	I.BLK	PP071388.D	22 Apr 2025 10:13	YP\AJ	Ok
3	AR1660ICC1000	PP071389.D	22 Apr 2025 10:29	YP\AJ	Ok
4	AR1660ICC750	PP071390.D	22 Apr 2025 10:45	YP\AJ	Ok
5	AR1660ICC500	PP071391.D	22 Apr 2025 11:02	YP\AJ	Ok
6	AR1660ICC250	PP071392.D	22 Apr 2025 11:18	YP\AJ	Ok
7	AR1660ICC050	PP071393.D	22 Apr 2025 11:34	YP\AJ	Ok,M
8	AR1221ICC500	PP071394.D	22 Apr 2025 11:51	YP\AJ	Ok
9	AR1232ICC500	PP071395.D	22 Apr 2025 12:07	YP\AJ	Ok
10	AR1242ICC1000	PP071396.D	22 Apr 2025 12:23	YP\AJ	Ok
11	AR1242ICC750	PP071397.D	22 Apr 2025 12:39	YP\AJ	Ok
12	AR1242ICC500	PP071398.D	22 Apr 2025 12:56	YP\AJ	Ok
13	AR1242ICC250	PP071399.D	22 Apr 2025 13:12	YP\AJ	Ok
14	AR1242ICC050	PP071400.D	22 Apr 2025 13:28	YP\AJ	Ok
15	AR1248ICC1000	PP071401.D	22 Apr 2025 13:45	YP\AJ	Ok
16	AR1248ICC750	PP071402.D	22 Apr 2025 14:01	YP\AJ	Ok
17	AR1248ICC500	PP071403.D	22 Apr 2025 14:17	YP\AJ	Ok
18	AR1248ICC250	PP071404.D	22 Apr 2025 14:33	YP\AJ	Ok
19	AR1248ICC050	PP071405.D	22 Apr 2025 14:50	YP\AJ	Ok
20	AR1254ICC1000	PP071406.D	22 Apr 2025 15:06	YP\AJ	Ok
21	AR1254ICC750	PP071407.D	22 Apr 2025 15:22	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PP071408.D	22 Apr 2025 15:38	YP\AJ	Ok
23	AR1254ICC250	PP071409.D	22 Apr 2025 15:55	YP\AJ	Ok
24	AR1254ICC050	PP071410.D	22 Apr 2025 16:11	YP\AJ	Ok,M
25	AR1262ICC500	PP071411.D	22 Apr 2025 16:27	YP\AJ	Ok
26	AR1268ICC1000	PP071412.D	22 Apr 2025 16:44	YP\AJ	Ok,M
27	AR1268ICC750	PP071413.D	22 Apr 2025 17:00	YP\AJ	Ok,M
28	AR1268ICC500	PP071414.D	22 Apr 2025 17:16	YP\AJ	Ok,M
29	AR1268ICC250	PP071415.D	22 Apr 2025 17:33	YP\AJ	Ok,M
30	AR1268ICC050	PP071416.D	22 Apr 2025 17:49	YP\AJ	Ok,M
31	PP042225ICV500	PP071417.D	22 Apr 2025 18:05	YP\AJ	Ok
32	AR1242ICV500	PP071418.D	22 Apr 2025 18:38	YP\AJ	Ok
33	AR1248ICV500	PP071419.D	22 Apr 2025 19:27	YP\AJ	Ok
34	AR1254ICV500	PP071420.D	22 Apr 2025 19:43	YP\AJ	Ok
35	AR1268ICV500	PP071421.D	22 Apr 2025 20:16	YP\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP050625

Review By	yogesh	Review On	5/6/2025 2:49:19 PM
Supervise By	mohammad	Supervise On	5/8/2025 2:15:08 AM
SubDirectory	PP050625	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP071799.D	06 May 2025 08:36	YP\AJ	Ok
2	AR1660CCC500	PP071800.D	06 May 2025 09:22	YP\AJ	Ok
3	AR1242CCC500	PP071801.D	06 May 2025 09:38	YP\AJ	Ok,M
4	AR1248CCC500	PP071802.D	06 May 2025 09:55	YP\AJ	Ok,M
5	AR1254CCC500	PP071803.D	06 May 2025 10:11	YP\AJ	Ok,M
6	I.BLK	PP071804.D	06 May 2025 10:27	YP\AJ	Ok
7	PB167853BL	PP071805.D	06 May 2025 10:44	YP\AJ	Not Ok
8	Q1938-03	PP071806.D	06 May 2025 11:00	YP\AJ	Not Ok
9	PB167877BL	PP071807.D	06 May 2025 13:12	YP\AJ	Ok
10	PB167877BS	PP071808.D	06 May 2025 13:28	YP\AJ	Ok,M
11	Q1956-01	PP071809.D	06 May 2025 13:45	YP\AJ	Ok
12	Q1956-02	PP071810.D	06 May 2025 14:01	YP\AJ	Ok
13	Q1956-03MS	PP071811.D	06 May 2025 14:18	YP\AJ	Ok,M
14	Q1956-04MSD	PP071812.D	06 May 2025 14:35	YP\AJ	Ok,M
15	Q1956-06	PP071813.D	06 May 2025 14:51	YP\AJ	Ok
16	Q1959-01	PP071814.D	06 May 2025 15:08	YP\AJ	Ok,M
17	AR1660CCC500	PP071815.D	06 May 2025 16:30	YP\AJ	Ok,M
18	AR1242CCC500	PP071816.D	06 May 2025 17:03	YP\AJ	Ok
19	AR1248CCC500	PP071817.D	06 May 2025 17:19	YP\AJ	Ok
20	AR1254CCC500	PP071818.D	06 May 2025 17:35	YP\AJ	Ok
21	I.BLK	PP071819.D	06 May 2025 17:51	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP050625

Review By	yogesh	Review On	5/6/2025 2:49:19 PM
Supervise By	mohammad	Supervise On	5/8/2025 2:15:08 AM
SubDirectory	PP050625	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PB167885BS	PP071820.D	06 May 2025 18:08	YP\AJ	Ok,M
23	PB167885BL	PP071821.D	06 May 2025 18:24	YP\AJ	Ok
24	Q1960-01	PP071822.D	06 May 2025 18:40	YP\AJ	Ok
25	Q1960-02	PP071823.D	06 May 2025 18:57	YP\AJ	Ok
26	Q1961-01	PP071824.D	06 May 2025 19:13	YP\AJ	Ok,M
27	Q1962-01	PP071825.D	06 May 2025 19:29	YP\AJ	Ok
28	Q1953-01	PP071826.D	06 May 2025 19:45	YP\AJ	Ok
29	Q1953-02	PP071827.D	06 May 2025 20:02	YP\AJ	Ok
30	Q1953-03	PP071828.D	06 May 2025 20:18	YP\AJ	Ok
31	AR1660CCC500	PP071829.D	06 May 2025 21:39	YP\AJ	Ok,M
32	AR1242CCC500	PP071830.D	06 May 2025 21:56	YP\AJ	Ok,M
33	AR1248CCC500	PP071831.D	06 May 2025 22:12	YP\AJ	Ok,M
34	AR1254CCC500	PP071832.D	06 May 2025 22:28	YP\AJ	Ok,M
35	I.BLK	PP071833.D	06 May 2025 22:45	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP050725

Review By	yogesh	Review On	5/7/2025 12:25:03 PM
Supervise By		Supervise On	
SubDirectory	PP050725	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP071834.D	07 May 2025 09:47	YP\AJ	Ok
2	AR1660CCC500	PP071835.D	07 May 2025 10:04	YP\AJ	Ok,NS
3	AR1242CCC500	PP071836.D	07 May 2025 10:24	YP\AJ	Ok,NS
4	AR1248CCC500	PP071837.D	07 May 2025 10:41	YP\AJ	Ok,NS
5	AR1254CCC500	PP071838.D	07 May 2025 10:57	YP\AJ	Ok
6	I.BLK	PP071839.D	07 May 2025 11:30	YP\AJ	Ok
7	PB167890BL	PP071840.D	07 May 2025 14:35	YP\AJ	Ok
8	PB167890BS	PP071841.D	07 May 2025 14:52	YP\AJ	Ok,NS
9	PB167890BSD	PP071842.D	07 May 2025 15:08	YP\AJ	Ok,NS
10	Q1956-07	PP071843.D	07 May 2025 15:24	YP\AJ	Ok
11	Q1965-03	PP071844.D	07 May 2025 15:41	YP\AJ	Ok
12	Q1965-04	PP071845.D	07 May 2025 15:57	YP\AJ	Ok
13	Q1931-01	PP071846.D	07 May 2025 16:13	YP\AJ	Ok,NS
14	AR1660CCC500	PP071847.D	07 May 2025 17:35	YP\AJ	Ok,NS
15	AR1242CCC500	PP071848.D	07 May 2025 17:51	YP\AJ	Ok,NS
16	AR1248CCC500	PP071849.D	07 May 2025 18:08	YP\AJ	Ok,NS
17	AR1254CCC500	PP071850.D	07 May 2025 18:24	YP\AJ	Ok
18	I.BLK	PP071851.D	07 May 2025 18:40	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP071387.D	22 Apr 2025 09:56		YPIAJ	Ok
2	I.BLK	I.BLK	PP071388.D	22 Apr 2025 10:13		YPIAJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP071389.D	22 Apr 2025 10:29		YPIAJ	Ok
4	AR1660ICC750	AR1660ICC750	PP071390.D	22 Apr 2025 10:45		YPIAJ	Ok
5	AR1660ICC500	AR1660ICC500	PP071391.D	22 Apr 2025 11:02		YPIAJ	Ok
6	AR1660ICC250	AR1660ICC250	PP071392.D	22 Apr 2025 11:18		YPIAJ	Ok
7	AR1660ICC050	AR1660ICC050	PP071393.D	22 Apr 2025 11:34		YPIAJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP071394.D	22 Apr 2025 11:51		YPIAJ	Ok
9	AR1232ICC500	AR1232ICC500	PP071395.D	22 Apr 2025 12:07		YPIAJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP071396.D	22 Apr 2025 12:23		YPIAJ	Ok
11	AR1242ICC750	AR1242ICC750	PP071397.D	22 Apr 2025 12:39		YPIAJ	Ok
12	AR1242ICC500	AR1242ICC500	PP071398.D	22 Apr 2025 12:56		YPIAJ	Ok
13	AR1242ICC250	AR1242ICC250	PP071399.D	22 Apr 2025 13:12		YPIAJ	Ok
14	AR1242ICC050	AR1242ICC050	PP071400.D	22 Apr 2025 13:28		YPIAJ	Ok
15	AR1248ICC1000	AR1248ICC1000	PP071401.D	22 Apr 2025 13:45		YPIAJ	Ok
16	AR1248ICC750	AR1248ICC750	PP071402.D	22 Apr 2025 14:01		YPIAJ	Ok
17	AR1248ICC500	AR1248ICC500	PP071403.D	22 Apr 2025 14:17		YPIAJ	Ok
18	AR1248ICC250	AR1248ICC250	PP071404.D	22 Apr 2025 14:33		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP071405.D	22 Apr 2025 14:50		YPIAJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PP071406.D	22 Apr 2025 15:06		YPIAJ	Ok
21	AR1254ICC750	AR1254ICC750	PP071407.D	22 Apr 2025 15:22		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP071408.D	22 Apr 2025 15:38		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PP071409.D	22 Apr 2025 15:55		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PP071410.D	22 Apr 2025 16:11		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP071411.D	22 Apr 2025 16:27		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP071412.D	22 Apr 2025 16:44		YPIAJ	Ok,M
27	AR1268ICC750	AR1268ICC750	PP071413.D	22 Apr 2025 17:00		YPIAJ	Ok,M
28	AR1268ICC500	AR1268ICC500	PP071414.D	22 Apr 2025 17:16		YPIAJ	Ok,M
29	AR1268ICC250	AR1268ICC250	PP071415.D	22 Apr 2025 17:33		YPIAJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PP071416.D	22 Apr 2025 17:49		YPIAJ	Ok,M
31	PP042225ICV500	ICVPP042225	PP071417.D	22 Apr 2025 18:05		YPIAJ	Ok
32	AR1242ICV500	ICVPP042225AR1242	PP071418.D	22 Apr 2025 18:38		YPIAJ	Ok
33	AR1248ICV500	ICVPP042225AR1248	PP071419.D	22 Apr 2025 19:27		YPIAJ	Ok
34	AR1254ICV500	ICVPP042225AR1254	PP071420.D	22 Apr 2025 19:43		YPIAJ	Ok
35	AR1268ICV500	ICVPP042225AR1268	PP071421.D	22 Apr 2025 20:16		YPIAJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP050625

Review By	yogesh	Review On	5/6/2025 2:49:19 PM
Supervise By	mohammad	Supervise On	5/8/2025 2:15:08 AM
SubDirectory	PP050625	HP Acquire Method	HP Processing Method PP042225

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP071799.D	06 May 2025 08:36		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP071800.D	06 May 2025 09:22		YP\AJ	Ok
3	AR1242CCC500	AR1242CCC500	PP071801.D	06 May 2025 09:38		YP\AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP071802.D	06 May 2025 09:55		YP\AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP071803.D	06 May 2025 10:11		YP\AJ	Ok,M
6	I.BLK	I.BLK	PP071804.D	06 May 2025 10:27		YP\AJ	Ok
7	PB167853BL	PB167853BL	PP071805.D	06 May 2025 10:44	Already run	YP\AJ	Not Ok
8	Q1938-03	LOWER-WALL-PILE-B	PP071806.D	06 May 2025 11:00	Already run	YP\AJ	Not Ok
9	PB167877BL	PB167877BL	PP071807.D	06 May 2025 13:12		YP\AJ	Ok
10	PB167877BS	PB167877BS	PP071808.D	06 May 2025 13:28		YP\AJ	Ok,M
11	Q1956-01	SB1-3-4	PP071809.D	06 May 2025 13:45		YP\AJ	Ok
12	Q1956-02	SB2-4-5	PP071810.D	06 May 2025 14:01	AR1254 Hit	YP\AJ	Ok
13	Q1956-03MS	SB2-4-5MS	PP071811.D	06 May 2025 14:18		YP\AJ	Ok,M
14	Q1956-04MSD	SB2-4-5MSD	PP071812.D	06 May 2025 14:35		YP\AJ	Ok,M
15	Q1956-06	SB91-3-4	PP071813.D	06 May 2025 14:51		YP\AJ	Ok
16	Q1959-01	SOIL-PILE	PP071814.D	06 May 2025 15:08	AR1260 Hit	YP\AJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PP071815.D	06 May 2025 16:30		YP\AJ	Ok,M
18	AR1242CCC500	AR1242CCC500	PP071816.D	06 May 2025 17:03		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP050625

Review By	yogesh	Review On	5/6/2025 2:49:19 PM
Supervise By	mohammad	Supervise On	5/8/2025 2:15:08 AM
SubDirectory	PP050625	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PP071817.D	06 May 2025 17:19		YPIAJ	Ok
20	AR1254CCC500	AR1254CCC500	PP071818.D	06 May 2025 17:35	Tcmx high in 2nd column	YPIAJ	Ok
21	I.BLK	I.BLK	PP071819.D	06 May 2025 17:51		YPIAJ	Ok
22	PB167885BS	PB167885BS	PP071820.D	06 May 2025 18:08		YPIAJ	Ok,M
23	PB167885BL	PB167885BL	PP071821.D	06 May 2025 18:24		YPIAJ	Ok
24	Q1960-01	344	PP071822.D	06 May 2025 18:40		YPIAJ	Ok
25	Q1960-02	72-11933	PP071823.D	06 May 2025 18:57		YPIAJ	Ok
26	Q1961-01	EO-01-050525	PP071824.D	06 May 2025 19:13		YPIAJ	Ok,M
27	Q1962-01	HD-01-050525	PP071825.D	06 May 2025 19:29		YPIAJ	Ok
28	Q1953-01	AUD-25-0075	PP071826.D	06 May 2025 19:45		YPIAJ	Ok
29	Q1953-02	AUD-25-0076	PP071827.D	06 May 2025 20:02		YPIAJ	Ok
30	Q1953-03	AUD-25-0077	PP071828.D	06 May 2025 20:18		YPIAJ	Ok
31	AR1660CCC500	AR1660CCC500	PP071829.D	06 May 2025 21:39		YPIAJ	Ok,M
32	AR1242CCC500	AR1242CCC500	PP071830.D	06 May 2025 21:56		YPIAJ	Ok,M
33	AR1248CCC500	AR1248CCC500	PP071831.D	06 May 2025 22:12		YPIAJ	Ok,M
34	AR1254CCC500	AR1254CCC500	PP071832.D	06 May 2025 22:28		YPIAJ	Ok,M
35	I.BLK	I.BLK	PP071833.D	06 May 2025 22:45		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP050725

Review By	yogesh	Review On	5/7/2025 12:25:03 PM
Supervise By		Supervise On	
SubDirectory	PP050725	HP Acquire Method	HP Processing Method PP042225

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP071834.D	07 May 2025 09:47		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP071835.D	07 May 2025 10:04		YPIAJ	Ok,NS
3	AR1242CCC500	AR1242CCC500	PP071836.D	07 May 2025 10:24		YPIAJ	Ok,NS
4	AR1248CCC500	AR1248CCC500	PP071837.D	07 May 2025 10:41		YPIAJ	Ok,NS
5	AR1254CCC500	AR1254CCC500	PP071838.D	07 May 2025 10:57		YPIAJ	Ok
6	I.BLK	I.BLK	PP071839.D	07 May 2025 11:30		YPIAJ	Ok
7	PB167890BL	PB167890BL	PP071840.D	07 May 2025 14:35		YPIAJ	Ok
8	PB167890BS	PB167890BS	PP071841.D	07 May 2025 14:52		YPIAJ	Ok,NS
9	PB167890BSD	PB167890BSD	PP071842.D	07 May 2025 15:08		YPIAJ	Ok,NS
10	Q1956-07	FB-05022025	PP071843.D	07 May 2025 15:24		YPIAJ	Ok
11	Q1965-03	0502	PP071844.D	07 May 2025 15:41		YPIAJ	Ok
12	Q1965-04	0505	PP071845.D	07 May 2025 15:57		YPIAJ	Ok
13	Q1931-01	NWB-2119	PP071846.D	07 May 2025 16:13		YPIAJ	Ok,NS
14	AR1660CCC500	AR1660CCC500	PP071847.D	07 May 2025 17:35		YPIAJ	Ok,NS
15	AR1242CCC500	AR1242CCC500	PP071848.D	07 May 2025 17:51		YPIAJ	Ok,NS
16	AR1248CCC500	AR1248CCC500	PP071849.D	07 May 2025 18:08		YPIAJ	Ok,NS
17	AR1254CCC500	AR1254CCC500	PP071850.D	07 May 2025 18:24		YPIAJ	Ok
18	I.BLK	I.BLK	PP071851.D	07 May 2025 18:40		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP050725

Review By	yogesh	Review On	5/7/2025 12:25:03 PM
Supervise By		Supervise On	
SubDirectory	PP050725	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

M : Manual Integration



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/6/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:05
In Date: 05/05/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 05/06/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135668

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q1956-01	SB1-3-4	1	1.17	9.98	11.15	10.05	89.0	
Q1956-02	SB2-4-5	2	1.18	10.62	11.8	11.11	93.5	
Q1956-03	Q1956-02MS	3	1.18	10.62	11.8	11.11	93.5	
Q1956-04	Q1956-02MSD	4	1.18	10.62	11.8	11.11	93.5	
Q1956-06	SB91-3-4	5	1.14	10.65	11.79	10.75	90.2	
Q1957-01	AT090P-SD05-050125-00	6	1.15	10.74	11.89	3.3	20.0	
Q1957-02	AT090P-SD03-050125-00	7	1.18	10.90	12.08	3.97	25.6	
Q1957-03	AT090P-SD04-050125-00	8	1.19	10.36	11.55	2.09	8.7	
Q1958-01	SS050P-SD28-043025-00	9	1.16	10.21	11.37	9.51	81.8	
Q1958-02	SS050P-SD29-043025-00	10	1.17	9.97	11.14	3.2	20.4	
Q1958-03	SS050P-SD30-050225-00	11	1.19	10.62	11.81	9.59	79.1	
Q1958-04	SS050P-SD31-050225-00	12	1.19	10.13	11.32	9.18	78.9	
Q1959-01	SOIL-PILE	13	1.18	10.18	11.36	10.11	87.7	
Q1960-01	344	14	1.00	1.00	2.00	2.00	100.0	stone sample,100% solids
Q1960-02	72-11933	15	1.15	10.38	11.53	10.24	87.6	
Q1961-01	EO-01-050225	16	1.18	10.08	11.26	10.53	92.8	
Q1961-02	EO-01-050225-E2	17	1.14	10.18	11.32	10.88	95.7	
Q1962-01	HD-01-050525	18	1.18	10.10	11.28	9.87	86.0	
Q1962-02	HD-01-050525-E2	19	1.19	10.43	11.62	9.91	83.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

VB 135668

WorkList Name : %1-050525-1 WorkList ID : 189298 Department : Wet-Chemistry Date : 05-05-2025 08:48:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1956-01	SB1-3-4	Solid	Percent Solids	Cool 4 deg C	CAMP02	L31	05/01/2025	Chemtech -SO
Q1956-02	SB2-4-5	Solid	Percent Solids	Cool 4 deg C	CAMP02	L31	05/02/2025	Chemtech -SO
Q1956-03	Q1956-02MS	Solid	Percent Solids	Cool 4 deg C	CAMP02	L31	05/02/2025	Chemtech -SO
Q1956-04	Q1956-02MSD	Solid	Percent Solids	Cool 4 deg C	CAMP02	L31	05/02/2025	Chemtech -SO
Q1956-06	SB91-3-4	Solid	Percent Solids	Cool 4 deg C	CAMP02	L31	05/01/2025	Chemtech -SO
Q1957-01	AT090P-SD05-050125-00	Solid	Percent Solids	Cool 4 deg C	WEST04	L51	05/01/2025	Chemtech -SO
Q1957-02	AT090P-SD03-050125-00	Solid	Percent Solids	Cool 4 deg C	WEST04	L51	05/01/2025	Chemtech -SO
Q1957-03	AT090P-SD04-050125-00	Solid	Percent Solids	Cool 4 deg C	WEST04	L51	05/01/2025	Chemtech -SO
Q1958-01	SS050P-SD28-043025-00	Solid	Percent Solids	Cool 4 deg C	WEST04	L51	04/30/2025	Chemtech -SO
Q1958-02	SS050P-SD29-043025-00	Solid	Percent Solids	Cool 4 deg C	WEST04	L51	04/30/2025	Chemtech -SO
Q1958-03	SS050P-SD30-050225-00	Solid	Percent Solids	Cool 4 deg C	WEST04	L51	05/02/2025	Chemtech -SO
Q1958-04	SS050P-SD31-050225-00	Solid	Percent Solids	Cool 4 deg C	WEST04	L51	05/02/2025	Chemtech -SO
Q1959-01	SOIL-PILE	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	05/05/2025	Chemtech -SO
Q1960-01	344	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	05/05/2025	Chemtech -SO
Q1960-02	72-11933	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	05/05/2025	Chemtech -SO
Q1961-01	EO-01-050225	Solid	Percent Solids	Cool 4 deg C	PSEG05	L51	05/05/2025	Chemtech -SO
Q1961-02	EO-01-050225-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	L51	05/05/2025	Chemtech -SO
Q1962-01	HD-01-050525	Solid	Percent Solids	Cool 4 deg C	PSEG05	L51	05/05/2025	Chemtech -SO
Q1962-02	HD-01-050525-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	L51	05/05/2025	Chemtech -SO

Date/Time 05/05/25 15:20

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 05/05/25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

SOP ID: M3541-ASE Extraction-14

Clean Up SOP #: Acid Cleanup

Matrix : Solid

Weigh By: EH

Balance check: RJ

Balance ID: EX-SC-2

pH Strip Lot#: N/A

Extraction By: RJ

Filter By: RJ

pH Meter ID: N/A

Hood ID: 3,7

Extraction Start Date : 05/06/2025

Extraction Start Time : 08:55

Extraction End Date : 05/06/2025

Extraction End Time : 11:55

Concentration By: EH

Supervisor By : RUPESH

Extraction Method: ☐ Separatory Funnel ☐ Continuous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☒ Soxhlet

Standardized Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24461
Surrogate	1.0ML	200 PPB	PP24460
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2601
Baked Na2SO4	N/A	EP2607
Hexane	N/A	E3933
H2SO4 1:1	N/A	EP2565
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40 ML Vial lot# 03-40 BTS723.

KD Bath ID: N/A

KD Bath Temperature: N/A

Envap ID: NEVAP-02


Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
5/6/25	RS (Ext-Lab)	Y.P. Peltier
12:00	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-14

Concentration Date: 05/06/2025

Sample ID	Client Sample ID	Test	g/mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB167877BL	ABLK877	PCB	30.02	N/A	ritesh	Evelyn	10			U7-1
PB167877BS	ALCS877	PCB	30.01	N/A	ritesh	Evelyn	10			2
Q1956-01	SB1-3-4	PCB	30.07	N/A	ritesh	Evelyn	10	M		3
Q1956-02	SB2-4-5	PCB	30.06	N/A	ritesh	Evelyn	10	M		4
Q1956-03	Q1956-02MS	PCB	30.02	N/A	ritesh	Evelyn	10	M		5
Q1956-04	Q1956-02MSD	PCB	30.03	N/A	ritesh	Evelyn	10	M		6
Q1956-06	SB91-3-4	PCB	30.08	N/A	ritesh	Evelyn	10	M		U6-1
Q1959-01	SOIL-PILE	PCB	30.06	N/A	ritesh	Evelyn	10	E		2
Q1960-01	344	PCB	30.02	N/A	ritesh	Evelyn	10		Stone	3
Q1960-02	72-11933	PCB	30.05	N/A	ritesh	Evelyn	10	E		4
Q1961-01	EO-01-050225	PCB	30.04	N/A	ritesh	Evelyn	10	E		5
Q1962-01	HD-01-050525	PCB	30.01	N/A	ritesh	Evelyn	10	E		6



RS
5/6

* Extracts relinquished on the same date as received.

167874
448491
8-53

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q1956 WorkList ID : 189338 Department : Extraction Date : 05-06-2025 08:47:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1956-01	SB1-3-4	Solid	PCB	Cool 4 deg C	CAMP02	L31	05/01/2025	8082A
Q1956-02	SB2-4-5	Solid	PCB	Cool 4 deg C	CAMP02	L31	05/02/2025	8082A
Q1956-03	Q1956-02MS	Solid	PCB	Cool 4 deg C	CAMP02	L31	05/02/2025	8082A
Q1956-04	Q1956-02MSD	Solid	PCB	Cool 4 deg C	CAMP02	L31	05/02/2025	8082A
Q1956-06	SB01-3-4	Solid	PCB	Cool 4 deg C	CAMP02	L31	05/01/2025	8082A
Q1959-01	SOIL-PILE	Solid	PCB	Cool 4 deg C	PSEG04	L41	05/05/2025	8082A
Q1960-01	344	Solid	PCB	Cool 4 deg C	PSEG03	L51	05/05/2025	8082A
Q1960-02	72-11933	Solid	PCB	Cool 4 deg C	PSEG03	L51	05/05/2025	8082A
Q1961-01	EO-01-050225	Solid	PCB	Cool 4 deg C	PSEG05	L51	05/05/2025	8082A
Q1962-01	HD-01-050525	Solid	PCB	Cool 4 deg C	PSEG05	L51	05/05/2025	8082A

Date/Time 05/06/25 8:50
Raw Sample Received by: RJ Geller-Vebj
Raw Sample Relinquished by: CP

Date/Time 05/06/25 9:15
Raw Sample Received by: CP
Raw Sample Relinquished by: RJ Geller-Vebj

SOP ID:	M3510C,3580A-Extraction PCB-14		
Clean Up SOP #:	Acid Cleanup	Extraction Start Date :	05/07/2025
Matrix :	Water	Extraction Start Time :	09:15
Welgh By:	N/A	Extraction By:	RS
Balance check:	N/A	Extraction End Date :	05/07/2025
Balance ID:	N/A	Filter By:	RS
pH Strip Lot#:	E3880	Extraction End Time :	14:00
		Concentration By:	EH
		Supervisor By :	RUPESH
Extraction Method:	☒ Separatory Funnel ☐ Continious Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☐ Soxhlet		

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24461
Surrogate	1.0ML	200 PPB	PP24460
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3930
Baked Na2SO4	N/A	EP2607
Hexane	N/A	E3933
H2SO4 1:1	N/A	EP2565
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40 ML Vial lot# 03-40BTS723, Q1931-01 Used Limited volume as sample is oily. Q1956-07 Limited volume received.


KD Bath ID:	WATER BATH-1,2	Envap ID:	NEVAP-02
KD Bath Temperature:	60 °C	Envap Temperature:	40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
5/7/25	RS (B4-Lab)	Y.P. Pest IP4B
14:05	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction PCB-14

Concentration Date: 05/07/2025

Sample ID	Client Sample ID	Test	g / <u>ml</u>	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB167890BL	ABLK890	PCB	1000	6	RUPESH	ritesh	10			SEP-1
PB167890BS	ALCS890	PCB	1000	6	RUPESH	ritesh	10			2
PB167890BS D	ALCSD890	PCB	1000	6	RUPESH	ritesh	10			3
Q1931-01	NWB-2119	PCB	100	6	RUPESH	ritesh	10	B		4
Q1956-07	FB-05022025	PCB	430	6	RUPESH	ritesh	5	F		5
Q1965-03	0502	PCB	1000	6	RUPESH	ritesh	10	B		6
Q1965-04	0505	PCB	990	6	RUPESH	ritesh	10	B		7



RS
517

167870
05/13/25

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q1931 WorkList ID : 199360 Department : Extraction Date : 05-07-2025 08:50:13

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1931-01	NWB-2119	Water	PCB	Cool 4 deg C	PSEG03	L51	05/01/2025	8082A
Q1956-07	FB-05022025	Water	PCB	Cool 4 deg C	CAMP02	L31	05/02/2025	8082A
Q1965-03	0502	Water	PCB	Cool 4 deg C	PSEG03	L41	05/06/2025	8082A
Q1965-04	0505	Water	PCB	Cool 4 deg C	PSEG03	L41	05/06/2025	8082A

Date/Time 5/7/25 9:10
Raw Sample Received by: RS (Ext-96b)
Raw Sample Relinquished by: CP SM

Date/Time 5/7/25 9:35
Raw Sample Received by: CP SM
Raw Sample Relinquished by: RS (Ext-66b)

Prep Standard - Chemical Standard Summary

Order ID : Q1956

Test : PCB

Prepbatch ID : PB167877,PB167890,

Sequence ID/Qc Batch ID: PP050625,pp050725,

Standard ID :

EP2565,EP2601,EP2607,PP24329,PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369,PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387,PP24460,PP24461,

Chemical ID :

E3551,E3804,E3877,E3916,E3917,E3930,E3933,M5173,P11522,P12699,P12702,P12931,P12936,P12949,P12955,P12957,P13355,P13356,P13373,P13381,P13589,P13591,P13697,P13702,P13830,P13878,P13883,W3112,W3177,

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
314	1.1 H2SO4 SOLN	EP2565	11/20/2024	05/20/2025	Rajesh Parikh	None	None	RUPESHKUMAR SHAH 11/20/2024

FROM 1000.00000ml of M5173 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
230	1:1ACETONE/HEXANE	EP2601	04/07/2025	10/03/2025	Rajesh Parikh	None	None	Riteshkumar Patel 04/07/2025

FROM 8000.00000ml of E3916 + 8000.00000ml of E3917 = Final Quantity: 8000.000 ml

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2607	04/25/2025	07/01/2025	RUPESHKUMAR SHAH	Extraction_SC ALE_2 (EX-SC-2)	None	Riteshkumar Patel 04/25/2025

FROM 4000.00000gram of E3551 = Final Quantity: 4000.000 gram

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24329	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P13356 + 9.00000ml of W3177 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
202	AR1660 1000/100 ppb working solution 1st source	PP24330	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13697 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
203	AR1660 750 PPB STD	PP24331	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24330 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
204	AR1660 500 PPB STD	PP24332	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24330 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
205	AR1660 250 PPB STD	PP24333	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24330 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
206	AR1660 50 PPB STD	PP24334	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24332 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
213	AR1221 1000 PPB WORKING SOLUTION	PP24335	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13702 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1079	AR1221 750 PPB STD	PP24336	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24335 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
222	AR1221 500 PPB STD	PP24337	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24335 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1080	AR1221 250 PPB STD	PP24338	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24335 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1081	AR1221 50 PPB STD	PP24339	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24337 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
214	AR1232 1000 PPB WORKING SOLUTION	PP24340	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13878 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1063	AR1232 750 PPB STD	PP24341	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24340 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
223	AR1232 500 PPB STD	PP24342	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24340 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1064	AR1232 250 PPB STD	PP24343	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24340 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1065	AR1232 50 PPB STD	PP24344	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24342 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
215	AR1242 1000 PPB WORKING STD	PP24345	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P12931 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1067	AR1242 750 PPB STD	PP24346	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.75000ml of PP24345 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
224	AR1242 500 PPB STD	PP24347	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24345 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1068	AR1242 250 PPB STD	PP24348	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24345 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1069	AR1242 50 PPB STD	PP24349	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24347 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
216	AR1248 1000 PPB WORKING STD	PP24350	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P12936 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1075	AR1248 750 PPB STD	PP24351	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24350 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
225	AR1248 500 PPB STD	PP24352	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24350 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1076	AR1248 250 PPB STD	PP24353	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24350 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1077	AR1248 50 PPB STD	PP24354	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24352 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
217	AR1254 1000 PPB WORKING STD	PP24355	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13830 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1071	AR1254 750 PPB STD	PP24356	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24355 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
226	AR1254 500 PPB STD	PP24357	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24355 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1072	AR1254 250 PPB STD	PP24358	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24355 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1073	AR1254 50 PPB STD	PP24359	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24357 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1529	AR1262 1000 PPB Working Solution	PP24360	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13883 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3753	AR1262 750 PPB STD	PP24361	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24360 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1530	AR1262 500 PPB STD	PP24362	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24360 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3754	AR1262 250 PPB STD	PP24363	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24360 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3755	AR1262 50 PPB STD	PP24364	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24362 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1532	AR1268 1000 PPB Working Solution	PP24365	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13381 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3820	AR1268 750 PPB STD	PP24366	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24365 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1533	AR1268 500 PPB STD	PP24367	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24365 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3821	AR1268 250 PPB STD	PP24368	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24365 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3822	AR1268 50 PPB STD	PP24369	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24367 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
404	AR1660 100 PPM Stock Solution 2nd Source	PP24370	03/18/2025	09/18/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12949 + 9.00000ml of E3804 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
405	AR1660 1000/100 PPB ICV STD	PP24371	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 98.50000ml of W3177 + 0.50000ml of PP24329 + 1.00000ml of PP24370 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
406	AR1660 500 PPB ICV	PP24372	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24371 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3789	AR1221 1000 PPB WORKING SOL.2ND SOURCE(AGILENT)	PP24373	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13373 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1886	AR1221 500 PPB ICV	PP24374	03/18/2025	08/12/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of E3877 + 0.50000ml of W3177 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1887	AR1232 1000 PPB Working Sol. 2nd Source	PP24375	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12699 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1888	AR1232 500 PPB ICV	PP24376	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24375 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1889	AR1242 1000 PPB Working Sol. 2nd Source	PP24377	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13589 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1891	AR1242 500 PPB ICV	PP24378	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24377 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1890	AR1248 1000 PPB Working Sol. 2nd Source	PP24379	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13591 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1892	AR1248 500 PPB ICV	PP24380	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24379 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1893	AR1254 1000 PPB Working Sol. 2nd Source	PP24381	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12957 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1894	AR1254 500 PPB ICV	PP24382	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24381 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3757	AR1262 1000 PPB Working Solution second source	PP24384	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12702 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3758	AR1262 500 PPB STD ICV	PP24385	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24384 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3817	AR1268 1000 ppb Working Soln. 2nd source	PP24386	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P11522 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3823	AR1268 500 PPB STD ICV	PP24387	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24386 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
465	200 PPB Pest/PCB Surrogate Spike	PP24460	04/11/2025	10/03/2025	Abdul Mirza	None	None	Yogesh Patel
								04/16/2025

FROM 1.00000ml of P13355 + 999.00000ml of E3917 = Final Quantity: 1000.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3857	5000 PPB PCB SPIKE SOLUTION 2ND SOURCE	PP24461	04/11/2025	10/03/2025	Abdul Mirza	None	None	Yogesh Patel 04/16/2025
<u>FROM</u> 0.50000ml of P12955 + 99.50000ml of E3917 = Final Quantity: 100.000 ml								

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	07/01/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	9005-05 / Acetone Ultra (cs/4x4L)	24E0761004	11/05/2025	10/01/2024 / Rajesh	09/25/2024 / Rajesh	E3804

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	08/12/2025	02/12/2025 / Rajesh	02/12/2025 / Rajesh	E3877

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	10/03/2025	04/03/2025 / Rajesh	03/31/2025 / Rajesh	E3916

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	10/03/2025	04/03/2025 / Rajesh	03/31/2025 / Rajesh	E3917

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25A0262002	02/20/2026	05/02/2025 / RUPESH	03/09/2025 / RUPESH	E3930

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	11/05/2025	05/05/2025 / RUPESH	04/23/2025 / RUPESH	E3933

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	0000281827	06/02/2025	06/01/2022 /	04/05/2022 / william	M5173

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-382-1 / Aroclor 1268	0006587800	09/18/2025	03/18/2025 / yogesh	02/21/2022 / Ankita	P11522

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc	91867 / Aroclor 1232 100 ug/mL	020823	09/18/2025	03/18/2025 / yogesh	08/07/2023 / Ankita	P12699

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc	x9166 / Aroclor 1262 100 ug/mL	060523	09/18/2025	03/18/2025 / yogesh	08/07/2023 / Ankita	P12702

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32009 / PCB Mix, Aroclor 1242, 1000ug/mL, Hexane, 1mL/ampul	a0203672	09/18/2025	03/18/2025 / yogesh	12/07/2023 / Ankita	P12931

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32010 / PCB Mix, Aroclor 1248, 1000ug/mL, Hexane, 1mL/ampul	a0202803	09/18/2025	03/18/2025 / yogesh	12/07/2023 / Ankita	P12936

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	09/18/2025	03/18/2025 / yogesh	12/20/2023 / Yogesh	P12949

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	10/11/2025	04/11/2025 / Abdul	12/20/2023 / Yogesh	P12955

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ Arochlor 1254	121823	04/03/2025	10/03/2024 / Ankita	12/20/2023 / Yogesh	P12957

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	10/11/2025	04/11/2025 / Abdul	04/22/2024 / Abdul	P13355

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	09/18/2025	03/18/2025 / yogesh	04/22/2024 / Abdul	P13356

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-292-1 / Aroclor 1221	0006783205	09/18/2025	03/18/2025 / yogesh	05/02/2024 / Ankita	P13373

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32410 / PCB Stock Solution, Aroclor 1268 Std, 1mL, Hexane	A0207475	09/18/2025	03/18/2025 / yogesh	05/03/2024 / Abdul	P13381

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-312-1 / Aroclor 1242	0006665550	09/18/2025	03/18/2025 / yogesh	10/14/2024 / Ankita	P13589

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-342-1 / Aroclor 1248	0006726317	09/18/2025	03/18/2025 / yogesh	10/14/2024 / Ankita	P13591

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32039 / PCB Mix, Aroclor 1016/1260, 1000ug/mL, hexane, 1mL/ampul	A0210629	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13697

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32007 / PCB Mix, Aroclor 1221, 1000ug/mL, Hexane, 1mL/ampul	A0215270	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13702

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32011 / PCB Mix, Aroclor 1254, 1000ug/mL, Hexane, 1mL/ampul	A0217391	09/18/2025	03/18/2025 / yogesh	12/09/2024 / Ankita	P13830

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32008 / PCB Mix, Aroclor 1232, 1000ug/mL, Hexane, 1mL/ampul	A0219655	09/18/2025	03/18/2025 / yogesh	01/23/2025 / Ankita	P13878

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32409 / PCB Stock Solution, Aroclor 1262 Std, 1mL, Hexane	A0220950	09/18/2025	03/18/2025 / yogesh	01/23/2025 / Ankita	P13883

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	08/22/2025	02/03/2025 / jignesh	01/31/2025 / jignesh	W3177



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/24/23 E 3551

RC-02-01, Ed. 3

Material No.: 9005-05
Batch No.: 24E0761004
Manufactured Date: 2024-05-02
Retest Date: 2029-05-01
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	≥ 99.5 %	99.8 %
Color (APHA)	≤ 10	< 5
Residue after Evaporation	≤ 5 ppm	< 1 ppm
Titration Acid (μeq/g)	≤ 0.3	0.1
Titration Base (μeq/g)	≤ 0.5	0.1
Water (H ₂ O)	≤ 0.5 %	0.1 %
Solubility in H ₂ O	Passes Test	Passes Test
Chloride (Cl)	≤ 0.2 ppm	< 0.2 ppm
Phosphate (PO ₄)	≤ 0.05 ppm	< 0.05 ppm
Trace Impurities – Aluminum (Al)	≤ 50.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 5.0 ppb
Trace Impurities – Barium (Ba)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Beryllium (Be)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Bismuth (Bi)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Boron (B)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Cadmium (Cd)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Calcium (Ca)	≤ 25.0 ppb	3.6 ppb
Trace Impurities – Chromium (Cr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Cobalt (Co)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Copper (Cu)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Gallium (Ga)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Germanium (Ge)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Gold (Au)	≤ 20 ppb	< 5 ppb
Trace Impurities – Iron (Fe)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Lead (Pb)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Lithium (Li)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Magnesium (Mg)	≤ 20 ppb	< 1 ppb
Trace Impurities – Manganese (Mn)	≤ 10.0 ppb	< 1.0 ppb

Recd by RP on 9/25/24

E 3804

>>> Continued on page 2 >>>

Acetone
CMOS

 **avantor™**



Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Nickel (Ni)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Niobium (Nb)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities – Potassium (K)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Silicon (Si)	≤ 50 ppb	< 10 ppb
Trace Impurities – Silver (Ag)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Sodium (Na)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Strontium (Sr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Tantalum (Ta)	≤ 50.0 ppb	< 5.0 ppb
Trace Impurities – Thallium (Tl)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Tin (Sn)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Titanium (Ti)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Vanadium (V)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Zinc (Zn)	≤ 20.0 ppb	7.9 ppb
Trace Impurities – Zirconium (Zr)	≤ 10.0 ppb	< 1.0 ppb
Particle Count – 0.5 µm and greater (Rion KS42AF)	≤ 100 par/ml	8 par/ml
Particle Count – 1.0 µm and greater (Rion KS42AF)	≤ 8 par/ml	2 par/ml

>>> Continued on page 3 >>>

Acetone
CMOS

 **avantor™**



Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result
------	---------------	--------

For Microelectronic Use

Country of Origin: USA
Packaging Site: Paris Mfg Ctr & DC



Michelle Bales
Sr. Manager, Quality Assurance

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Recd. by RP on 2/12/25

Harout Sahagian E3877

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.

If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT - 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A

Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Harout Sahagian

Harout Sahagian - Quality Control Manager - Fair Lawn

Recd by RP on 3/31/25

E 3946

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.
 If there are any questions with this certificate, please call at (800) 227-6701.
 *Based on suggested storage condition.

Acetone

BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H2762008

Manufactured Date: 2024-04-18

Expiration Date: 2027-04-18

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

Recd by RP on 03/31/25

E3917

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor



Material No.: 9266-A4
Batch No.: 25A0262002
Manufactured Date: 2024-11-21
Expiration Date: 2026-02-20
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	4
Assay (CH ₂ Cl ₂) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8 \%$	99.9 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.8 ppm
Titration Acid (μ eq/g)	≤ 0.3	<0.1
Chloride (Cl)	≤ 10 ppm	<5 ppm
Water (by KF, coulometric)	$\leq 0.02 \%$	<0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3930

J. Croak

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

 **avantorsTM**



Material No.: 9262-03

Batch No.: 25C0362005

Manufactured Date: 2025-01-29

Expiration Date: 2026-04-30

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

E3933



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Hydrochloric Acid, 36.5–38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis



Material No.: 9530-33
Batch No.: 0000281827
Manufactured Date: 2021/03/30
Retest Date: 2026/03/29
Revision No: 1

Certificate of Analysis

Test	Specification	Result
ACS – Assay (as HCl) (by acid–base titrn)	36.5 – 38.0 %	37.6
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	1
ACS – Specific Gravity at 60°/60°F	1.185 – 1.192	1.189
ACS – Bromide (Br)	<= 0.005 %	< 0.005
ACS – Extractable Organic Substances	<= 5 ppm	< 1
ACS – Free Chlorine (as Cl ₂)	<= 0.5 ppm	< 0.5
Phosphate (PO ₄)	<= 0.05 ppm	< 0.03
Sulfate (SO ₄)	<= 0.5 ppm	< 0.3
Sulfite (SO ₃)	<= 0.8 ppm	0.3
Ammonium (NH ₄)	<= 3 ppm	< 1
Trace Impurities – Arsenic (As)	<= 0.010 ppm	< 0.003
Trace Impurities – Aluminum (Al)	<= 10.0 ppb	0.5
Arsenic and Antimony (as As)	<= 5 ppb	< 3
Trace Impurities – Barium (Ba)	<= 1.0 ppb	< 0.2
Trace Impurities – Beryllium (Be)	<= 1.0 ppb	< 0.2
Trace Impurities – Bismuth (Bi)	<= 10.0 ppb	< 1.0
Trace Impurities – Boron (B)	<= 20.0 ppb	< 5.0
Trace Impurities – Cadmium (Cd)	<= 1.0 ppb	< 0.3
Trace Impurities – Calcium (Ca)	<= 50.0 ppb	15.0
Trace Impurities – Chromium (Cr)	<= 1.0 ppb	< 0.4
Trace Impurities – Cobalt (Co)	<= 1.0 ppb	< 0.3
Trace Impurities – Copper (Cu)	<= 1.0 ppb	< 0.1
Trace Impurities – Gallium (Ga)	<= 1.0 ppb	< 0.2

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Test	Specification	Result
Trace Impurities – Germanium (Ge)	<= 3.0 ppb	< 2.0
Trace Impurities – Gold (Au)	<= 4.0 ppb	3.0
Heavy Metals (as Pb)	<= 100 ppb	< 50
Trace Impurities – Iron (Fe)	<= 15.0 ppb	1.0
Trace Impurities – Lead (Pb)	<= 1.0 ppb	< 0.5
Trace Impurities – Lithium (Li)	<= 1.0 ppb	< 0.2
Trace Impurities – Magnesium (Mg)	<= 10.0 ppb	< 0.4
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	< 0.4
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	0.2
Trace Impurities – Molybdenum (Mo)	<= 10.0 ppb	< 5.0
Trace Impurities – Nickel (Ni)	<= 4.0 ppb	< 0.3
Trace Impurities – Niobium (Nb)	<= 1.0 ppb	< 0.2
Trace Impurities – Potassium (K)	<= 9.0 ppb	< 2.0
Trace Impurities – Selenium (Se), For Information Only	ppb	1.0
Trace Impurities – Silicon (Si)	<= 100.0 ppb	18.0
Trace Impurities – Silver (Ag)	<= 1.0 ppb	< 0.3
Trace Impurities – Sodium (Na)	<= 100.0 ppb	< 5.0
Trace Impurities – Strontium (Sr)	<= 1.0 ppb	< 0.2
Trace Impurities – Tantalum (Ta)	<= 1.0 ppb	< 0.9
Trace Impurities – Thallium (Tl)	<= 5.0 ppb	< 2.0
Trace Impurities – Tin (Sn)	<= 5.0 ppb	< 0.8
Trace Impurities – Titanium (Ti)	<= 1.0 ppb	< 0.2
Trace Impurities – Vanadium (V)	<= 1.0 ppb	< 0.2
Trace Impurities – Zinc (Zn)	<= 5.0 ppb	0.4
Trace Impurities – Zirconium (Zr)	<= 1.0 ppb	< 0.1

For Laboratory, Research or Manufacturing Use

Product Information (not specifications):

Appearance (clear, fuming liquid)

Meets ACS Specifications

Country of Origin: US

Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality



Certificate of Analysis

P11518
↓
P11522
AJ
02/21/22

Product Name: Aroclor 1268 Standard

Product Number: PP-382-1

Lot Issue Date: 09-Feb-2021

Lot Number: 0006587800

Expiration Date: 31-Mar-2029

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
Aroclor 1268	011100-14-4	RM00937	100.0 ± 0.5 µg/mL

Matrix: isooctane (2,2,4-trimethylpentane)

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.


Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026
Page: 1 of 1

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937



CERTIFIED WEIGHT REPORT

Part Number: **91867**
 Lot Number: **020823**
 Description: **WP 037 - Aroclor 1232**
PCB Technical Mixture
 Expiration Date: **020833**
 Recommended Storage: **Ambient (20 °C)**
 Nominal Concentration ($\mu\text{g/mL}$): **100**
 NIST Test ID#: **6UTB**

Solvent(
Aceton


Weight(s) shown below were combined and diluted to (mL): **100.0**
 5E-05 Balance Uncertainty
 0.057 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc ($\mu\text{g/mL}$)	Purity (%)	Uncertainty Purity	Target Weight (g)
1. Aroclor 1232	17	45-6A	100	100	0.5	0.01000

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measure Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments

GC3-M1 Analysis by Melissa Stonier
 Column ID SPB-608 30 meter X 0.53mm X 5 μm film thickness
 Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
 Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
 Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
 Rate = 8°C/min, Total run time = 35 min
 Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
 Standard injection = 1.5 μL , Range=3






110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32009 **Lot No.:** A0203672

Description: Aroclor® 1242 Standard

Aroclor® 1242 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: January 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

p12928

→

p12932

AJ

12/27/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Aroclor 1242	53469-21-9	01141	—%	1,004.7 µg/mL	+/- 55.7515

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2ul



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Russ Boothamer

Russ Boothamer - Operations Technician I

Date Mixed: 26-Oct-2023

Balance Serial #

B442140311

Jennifer Polino

Jennifer Polino - Operations Tech III - ARM QC

Date Passed: 06-Nov-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FW 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32010 **Lot No.:** A0202803

Description: Aroclor® 1248 Standard

Aroclor® 1248 Standard 1,000µg/mL, Hexane, 1mL/ampul

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: January 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P1210967
P1210977
P1210973
P1210973

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1248	12672-29-6	13897600	---%	1,001.7 µg/mL	+/- 55.5850

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2ul



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Laith Clemente
Laith Clemente - Operations Technician I

Date Mixed: 03-Oct-2023

Balance Serial #

1128360905

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 09-Oct-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FW 80397



Certified Reference Material CRM

CERTIFIED WEIGHT REPORT

Part Number:
20064

Solvent(s):
Hexane

Lot#
273615

Lot Number:
022023

Formulated By:
Benson Chan

DATE
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Reviewed By:
Pedro L. Rentas

DATE
022023

Expiration Date:
022033

Expanded
Uncertainty
(+/-) (µg/mL)

4.0

Recommended Storage:
Ambient (20 °C)

Actual
Conc (µg/mL)

1002.8

Nominal Concentration (µg/mL):
1000

Actual
Weight(g)

0.20060

NIST Test ID#:
6UTB

Target
Weight(g)

0.20004

Weight(s) shown below were combined and diluted to (mL):
200.0

Purity
(%)

100

5E-05 Balance Uncertainty
0.010 Flask Uncertainty

Conc (µg/mL)

1000

SDS Information
(Solvent Safety Info. On Attached pg.)

OSHA PEL (TWA)


LD50

PR2946 718
↓
12/20/23
P1295

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11086-82-5	0.5mg/m3	orl-rat 1315mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stortier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 280°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C, FID Signal = Etdaq Channel 1
Standard Injection = 1.5µL, Range=3





Certified Reference Material CRM

CERTIFIED WEIGHT REPORT

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

5E-05 Balance Uncertainty
0.010 Flask Uncertainty

Nominal Conc (µg/mL):
200.0

Weight(s) shown below were combined and diluted to (mL):

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

5E-05 Balance Uncertainty
0.010 Flask Uncertainty


Nominal Conc (µg/mL):
200.0

Weight(s) shown below were combined and diluted to (mL):

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11086-82-5	0.5mg/m3	orl-rat 1315mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stortier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 280°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C, FID Signal = Etdaq Channel 1
Standard Injection = 1.5µL, Range=3





CERTIFIED WEIGHT REPORT

Part Number: 99139
Lot Number: 121823
Description: Aroclor 1254

Expiration Date: 121833
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 100
NIST Test ID#: 6UTB

Volume(s) shown below were combined and diluted to (mL): 20.0
Note: Aroclor 1254 is a mix of isomers.

Solvent(s): Iso-octane
Lot#: 82227

Formulated By: Anthony Mahoney	121823	DATE
Reviewed By: Pedro L. Rentas	121823	DATE

P12956 / Y.P.
L
P12957 / 12/18/23


Note: Aroclor 1254 is a mix of isomers.

Compound		Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Initial Uncertainty	Pipette (mL)	Conc.(µg/mL)	Initial Conc.(µg/mL)	Final Conc.(µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1254		79100	121823	0.10	2.00	0.017	0.017	1003.3	100.1	11097-69-1	1.8	0.5mg/m3 (skin)	or-rat 1295mg/kg	

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X 5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150 °C (Time 1 = 4 min), Temp 2 = 280 °C (Time 2 = 13.5 min)
Rate = 8 °C/min, Total run time = 35 min
Injector temp. = 200 °C, FID Temp. = 300 °C, FID Signal = Etdaq Channel 1
Standard injection = 1.5µL, Range=3

Peak No.	Name	FID RT (min.)
1	Total Aroclor 1254	18.12





110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
BAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024

Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
SAUF
04/25/2025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024


Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
10


SAUF
04/25/2025



ISO 17034

Reference Material Certificate

Product Information Sheet

Product Name: Aroclor 1221 Standard
Product Number: PP-292-1
Storage Conditions: Store at Room Temperature (15° to 30°C).

Lot Number: 0006783205
Lot Issue Date: 20-Feb-2024
Expiration Date: 31-Mar-2032

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
Aroclor 1221	100.3 ±	0.5 µg/mL	011104-28-2	NT01017

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13342
AJ
05106124
P13343

Page: 1 of 2

CSD-QA-015.2

ISO 17025

Cert No. AT-1937


250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034
Cert No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO
9001:2015 Quality Management System. Cert# 951215321

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-015.2

ISO 17025
Cert No. AT-1937

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32410 **Lot No.:** A0207475

Description : Aroclor® 1268 Standard

Aroclor® 1268 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1268	11100-14-4	10947000	---%	1,000.0 µg/mL	+/- 55.4925

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P 13380
↓
P 13381
↓
SAZU
05/6/2024

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

Split ratio 500:1

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 06-Feb-2024

Balance Serial # B442140311

Dillon Murphy
Dillon Murphy - Operations Technician I


Date Passed: 09-Feb-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13380 } ②
↓
P13381
1

[Signature]
05/6/2024

ISO 17034



Agilent

Trusted Answers

Reference Material Certificate

Product Name: Aroclor 1242 Standard

Lot Number: 0006665550

Product Number: PP-312-1

Lot Issue Date: 08-Feb-2022

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 31-Jan-2027

Component Name	CERTIFIED VALUES		CAS#	Analyte Lot
	Concentration	Expanded Uncertainty		
Aroclor 1242	100.4	± 0.5 µg/mL	053469-21-9	NT01020

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13589
↓
P13590AJ
10/14/24

ISO 17034



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO 9001:2015
Quality Management System. Cert# 951215321

Page: 2 of 2

www.agilent.com/quality/

CSD-QA-015.1



ISO 17025
Cert No. AT-

Reference Material Certificate
Product Information Sheet

Product Name: Aroclor 1248 Standard

Lot Number: 0006726317

Product Number: PP-342-1

Lot Issue Date: 27-Jan-2023

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 28-Feb-2031

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
Aroclor 1248	100.3 ±	0.5 µg/mL	012672-29-6	NT01582

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material (RM) standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above. Purity values are taken from approved vendor raw material certificates.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference (RM) standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference (RM) standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard (RM) is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13591
↓
P13592

AS
10/14/2024

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:



Monica Bourgeois
QMS Representative



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32039 **Lot No.:** A0210629

Description : Aroclor® 1016/1260 Mix
Aroclor® 1016/1260 Mix 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13697
↓
P13701 } Y.P.
10/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1016	12674-11-2	07	----%	1,005.3 µg/mL	+/- 55.7809
2	Aroclor 1260	11096-82-5	1320657	----%	1,000.0 µg/mL	+/- 55.4850

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

0.2µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B442140311


Dillan Murphy - Operations Technician I

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32007 **Lot No.:** A0215270

Description : Aroclor® 1221 Standard

Aroclor® 1221 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13902
P13703 } Y.P.
10/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1221	11104-28-2	14969200	----%	1,005.0 µg/mL	+/- 55.7700

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 16-Aug-2024

Balance Serial # 1128360905

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 20-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32011 **Lot No.:** A0217391

Description : Aroclor® 1254 Standard

Aroclor® 1254 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1254	11097-69-1	124-191-B	----%	1,004.7 µg/mL	+/- 55.7515

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13830
↓
P13832
AJ
12/09/24

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 02-Oct-2024

Balance Serial # C322230531

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 07-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32008

Lot No.: A0219655

Description : Aroclor® 1232 Standard

Aroclor® 1232 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2031

Storage: 25°C nominal

Handling:

This product contains PCBs.

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1232	11141-16-5	15665-01	---%	1,007.0 µg/mL	+/- 55.8810

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13878



AJ
01/28/25

P13880

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 02-Dec-2024

Balance Serial #

C322230531

Brittany Federhik

Brittany Federhiko - Operations Tech I

Date Passed: 05-Dec-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32409 **Lot No.:** A0220950

Description: Aroclor® 1262 Standard

Aroclor® 1262 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: April 30, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1262	37324-23-5	10849100	----%	1,002.0 µg/mL	+/- 55.6035

Solvent: Hexane

CAS # 110-54-3
Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13882



AJ
01/28/25

P13883

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckal - Mix Technician

Date Mixed: 09-Jan-2025 Balance Serial # C322230531



Britiany Federinko - Operations Tech I

Date Passed: 14-Jan-2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: CDM Smith
ADDRESS: 110 Fieldcrest Avenue #8 6th Floor
CITY Edison STATE: NJ ZIP: 08837
ATTENTION: Marcie Encinas
PHONE: 732-590-4679 FAX: 732-225-7851

CLIENT PROJECT INFORMATION

PROJECT NAME: Bergen Point Project
PROJECT NO.: 99939 LOCATION: W. Babylon, NY
PROJECT MANAGER: Marcie Encinas
e-mail: ENcinasMA@CDMSmith.com
PHONE: 732-590-4679 FAX: 732-225-7851

CLIENT BILLING INFORMATION

BILL TO: CDM Smith PO#:
ADDRESS: 110 Fieldcrest Avenue, #8 Floor 6th
CITY Edison STATE: NJ ZIP: 08837
ATTENTION: Marcie Encinas PHONE: 732-590-4690

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: _____ DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☒ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☒ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other _____
☐ EDD FORMAT _____

1. TCL VOC
2. TCL SVOC
3. TAL METALS
4. PCBs
5. PESTICIDES
6. HERBICIDES
7. PRO/GRO
8. FULL TCLP
9. RCRA CHARAC

PRESERVATIVES

COMMENTS

← Specify Preservatives
A-HCl D-NaOH
B-HNO3 E-ICE
C-H2SO4 F-OTHER

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES										
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	SB1-3-4'	Soil		X	5/1/25	9:00AM	13	X	X	X	X	X	X	X			E
2.	SB2-4-5'	Soil		X	5/2/25	12:00PM	13	X	X	X	X	X	X	X			E
3.	COMP1	Soil	X		5/2/25	1:30PM	3								X	X	E
4.	SB91-3-4'	Soil		X	5/1/25	9:30AM	13	X	X	X	X	X	X	X	X		E
5.	SB2-MS	Soil		X	5/2/25	12:30PM	13	X	X	X	X	X	X	X			E
6.	SB2-MSD	Soil		X	5/2/25	12:40PM	13	X	X	X	X	X	X	X			E
7.	FB-05022025	Aqueous		X	5/2/25	11:00AM	9	X	X	X	X	X	X	X			E, A, B
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <i>[Signature]</i>	DATE/TIME: 1404 5.2.2025	RECEIVED BY: <i>[Signature]</i>	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☒ COOLER TEMP 3 °C
RELINQUISHED BY SAMPLER: 2. <i>[Signature]</i>	DATE/TIME:	RECEIVED BY:	Comments: IR Can #1 (Adjusted Factor + 1)
RELINQUISHED BY SAMPLER: 3. <i>[Signature]</i>	DATE/TIME: 1820 5.2.2025	RECEIVED BY:	

Page ____ of

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete
☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1956	CAMP02	Order Date : 5/2/2025 3:14:35 PM	Project Mgr :
Client Name : CDM Smith		Project Name : Bergen Point Fueling System	Report Type : NYS ASP A
Client Contact : Marcie Ann Encinas		Receive DateTime : 5/2/2025 12:00:00 AM	EDD Type : EQUIS
Invoice Name : CDM Smith		Purchase Order : 18:26	Hard Copy Date :
Invoice Contact : Marcie Ann Encinas			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1956-01	SB1-3-4	Solid	05/01/2025	09:00					
					VOC-TCLVOA-10		8260D	10 Bus. Days	
Q1956-02	SB2-4-5	Solid	05/02/2025	12:00					
					VOC-TCLVOA-10		8260D	10 Bus. Days	
Q1956-03	Q1956-02MS	Solid	05/02/2025	12:30					
					VOC-TCLVOA-10		8260D	10 Bus. Days	
Q1956-04	Q1956-02MSD	Solid	05/02/2025	12:40					
					VOC-TCLVOA-10		8260D	10 Bus. Days	
Q1956-06	SB19-3-4 SB91	Solid	05/01/2025	09:30					
					VOC-TCLVOA-10		8260D	10 Bus. Days	
Q1956-07	FB-05022025	Water	05/02/2025	11:00					
					VOC-TCLVOA-10		8260-Low	10 Bus. Days	

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1956	CAMP02	Order Date : 5/2/2025 3:14:35 PM	Project Mgr :
Client Name : CDM Smith		Project Name : Bergen Point Fueling System	Report Type : NYS ASP A
Client Contact : Marcie Ann Encinas		Receive DateTime : 5/2/2025 12:00:00 AM	EDD Type : EQUIS
Invoice Name : CDM Smith		Purchase Order : 18-26	Hard Copy Date :
Invoice Contact : Marcie Ann Encinas			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1956-09 TB		water	5/2/25	12:00	VOC-TCLVOA-10		8260-LOW		10 Bus Days.

Relinquished By : 

Date / Time : 5/5/25 0835

Received By : 

Date / Time : 5/5/25 08:35

Storage Area : VOA Refridgerator Room

Rep # 6
FZ-2