

Surrogate Summary

SDG No.: Q1956

Client: CDM Smith

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP071388.D	PIBLK-PP071388.D	Tetrachloro-m-xylene	1	20	17.1	85		60	140
		Decachlorobiphenyl	1	20	17.6	88		60	140
		Tetrachloro-m-xylene	2	20	17.2	86		60	140
		Decachlorobiphenyl	2	20	17.6	88		60	140
I.BLK-PP071804.D	PIBLK-PP071804.D	Tetrachloro-m-xylene	1	20	17.1	86		60	140
		Decachlorobiphenyl	1	20	19.0	95		60	140
		Tetrachloro-m-xylene	2	20	19.0	95		60	140
		Decachlorobiphenyl	2	20	20.2	101		60	140
PB167877BL	PB167877BL	Tetrachloro-m-xylene	1	20	20.6	103		32	144
		Decachlorobiphenyl	1	20	21.9	109		32	175
		Tetrachloro-m-xylene	2	20	22.9	115		32	144
		Decachlorobiphenyl	2	20	24.6	123		32	175
PB167877BS	PB167877BS	Tetrachloro-m-xylene	1	20	20.7	104		32	144
		Decachlorobiphenyl	1	20	22.2	111		32	175
		Tetrachloro-m-xylene	2	20	22.4	112		32	144
		Decachlorobiphenyl	2	20	24.1	121		32	175
Q1956-01	SB1-3-4	Tetrachloro-m-xylene	1	20	20.9	104		32	144
		Decachlorobiphenyl	1	20	16.1	81		32	175
		Tetrachloro-m-xylene	2	20	23.5	117		32	144
		Decachlorobiphenyl	2	20	18.3	91		32	175
Q1956-02	SB2-4-5	Tetrachloro-m-xylene	1	20	20.7	104		32	144
		Decachlorobiphenyl	1	20	17.2	86		32	175
		Tetrachloro-m-xylene	2	20	22.6	113		32	144
		Decachlorobiphenyl	2	20	18.0	90		32	175
Q1956-03MS	SB2-4-5MS	Tetrachloro-m-xylene	1	20	22.6	113		32	144
		Decachlorobiphenyl	1	20	21.1	105		32	175
		Tetrachloro-m-xylene	2	20	23.7	119		32	144
		Decachlorobiphenyl	2	20	23.3	117		32	175
Q1956-04MSD	SB2-4-5MSD	Tetrachloro-m-xylene	1	20	21.8	109		32	144
		Decachlorobiphenyl	1	20	21.0	105		32	175
		Tetrachloro-m-xylene	2	20	21.2	106		32	144
		Decachlorobiphenyl	2	20	22.7	114		32	175
Q1956-06	SB91-3-4	Tetrachloro-m-xylene	1	20	19.7	99		32	144
		Decachlorobiphenyl	1	20	14.9	74		32	175
		Tetrachloro-m-xylene	2	20	22.8	114		32	144
		Decachlorobiphenyl	2	20	16.2	81		32	175
I.BLK-PP071819.D	PIBLK-PP071819.D	Tetrachloro-m-xylene	1	20	17.0	85		60	140
		Decachlorobiphenyl	1	20	18.7	93		60	140
		Tetrachloro-m-xylene	2	20	19.6	98		60	140
		Decachlorobiphenyl	2	20	21.9	109		60	140
I.BLK-PP071839.D	PIBLK-PP071839.D	Tetrachloro-m-xylene	1	20	16.3	81		60	140

Surrogate Summary

SDG No.: Q1956

Client: CDM Smith

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP071839.D	PIBLK-PP071839.D	Decachlorobiphenyl	1	20	16.5	83		60	140
		Tetrachloro-m-xylene	2	20	18.1	91		60	140
PB167890BL	PB167890BL	Decachlorobiphenyl	2	20	19.2	96		60	140
		Tetrachloro-m-xylene	1	20	20.3	102		16	158
		Decachlorobiphenyl	1	20	20.3	101		10	173
		Tetrachloro-m-xylene	2	20	21.9	110		16	158
		Decachlorobiphenyl	2	20	23.4	117		10	173
		Tetrachloro-m-xylene	1	20	20.4	102		16	158
PB167890BS	PB167890BS	Decachlorobiphenyl	1	20	20.6	103		10	173
		Tetrachloro-m-xylene	2	20	20.5	102		16	158
		Decachlorobiphenyl	2	20	22.1	110		10	173
		Tetrachloro-m-xylene	1	20	19.7	99		16	158
PB167890BSD	PB167890BSD	Decachlorobiphenyl	1	20	19.8	99		10	173
		Tetrachloro-m-xylene	2	20	20.7	104		16	158
		Decachlorobiphenyl	2	20	22.1	111		10	173
		Tetrachloro-m-xylene	1	20	21.4	107		16	158
		Decachlorobiphenyl	1	20	12.3	62		10	173
		Tetrachloro-m-xylene	2	20	23.5	117		16	158
Q1956-07	FB-05022025	Decachlorobiphenyl	2	20	13.8	69		10	173
		Tetrachloro-m-xylene	1	20	16.2	81		60	140
		Decachlorobiphenyl	1	20	16.6	83		60	140
		Tetrachloro-m-xylene	2	20	17.9	90		60	140
I.BLK-PP071851.D	PIBLK-PP071851.D	Decachlorobiphenyl	2	20	19.1	96		60	140