

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL01 Date Analyzed: 05/08/2025

Lab Sample No.: PSTDCCC050 Data File : PD088434.D Time Analyzed: 17:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.706	6.606	6.806	56.580	50.000	13.2
4,4'-DDE	6.197	6.097	6.297	55.070	50.000	10.1
4,4'-DDT	7.022	6.922	7.122	55.680	50.000	11.4
Aldrin	5.272	5.173	5.373	55.840	50.000	11.7
alpha-BHC	4.000	3.901	4.101	56.080	50.000	12.2
alpha-Chlordane	6.028	5.929	6.129	55.480	50.000	11.0
beta-BHC	4.516	4.416	4.616	54.220	50.000	8.4
Decachlorobiphenyl	9.075	8.975	9.175	53.340	50.000	6.7
delta-BHC	4.764	4.665	4.865	55.650	50.000	11.3
Dieldrin	6.348	6.249	6.449	55.760	50.000	11.5
Endosulfan I	6.075	5.976	6.176	55.130	50.000	10.3
Endosulfan II	6.787	6.687	6.887	54.420	50.000	8.8
Endosulfan sulfate	7.150	7.050	7.250	55.170	50.000	10.3
Endrin	6.576	6.476	6.676	55.920	50.000	11.8
Endrin aldehyde	6.916	6.816	7.016	55.830	50.000	11.7
Endrin ketone	7.631	7.531	7.731	55.550	50.000	11.1
gamma-BHC (Lindane)	4.331	4.232	4.432	54.990	50.000	10.0
gamma-Chlordane	5.947	5.847	6.047	55.410	50.000	10.8
Heptachlor	4.930	4.831	5.031	55.690	50.000	11.4
Heptachlor epoxide	5.691	5.592	5.792	55.150	50.000	10.3
Methoxychlor	7.494	7.395	7.595	54.990	50.000	10.0
Tetrachloro-m-xylene	3.551	3.452	3.652	53.540	50.000	7.1

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL01 Date Analyzed: 05/08/2025

Lab Sample No.: PSTDCCC050 Data File : PD088434.D Time Analyzed: 17:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.932	5.834	6.034	52.600	50.000	5.2
4,4'-DDE	5.378	5.280	5.480	53.540	50.000	7.1
4,4'-DDT	6.186	6.088	6.288	53.240	50.000	6.5
Aldrin	4.371	4.273	4.473	52.680	50.000	5.4
alpha-BHC	3.395	3.296	3.496	51.780	50.000	3.6
alpha-Chlordane	5.193	5.094	5.294	51.660	50.000	3.3
beta-BHC	4.027	3.928	4.128	51.200	50.000	2.4
Decachlorobiphenyl	8.076	7.977	8.177	51.180	50.000	2.4
delta-BHC	4.264	4.165	4.365	52.070	50.000	4.1
Dieldrin	5.515	5.417	5.617	52.210	50.000	4.4
Endosulfan I	5.249	5.151	5.351	49.810	50.000	-0.4
Endosulfan II	6.082	5.984	6.184	51.940	50.000	3.9
Endosulfan sulfate	6.485	6.386	6.586	52.030	50.000	4.1
Endrin	5.791	5.693	5.893	52.470	50.000	4.9
Endrin aldehyde	6.261	6.163	6.363	52.440	50.000	4.9
Endrin ketone	6.994	6.895	7.095	52.050	50.000	4.1
gamma-BHC (Lindane)	3.731	3.633	3.833	51.100	50.000	2.2
gamma-Chlordane	5.128	5.030	5.230	51.820	50.000	3.6
Heptachlor	4.085	3.986	4.186	51.480	50.000	3.0
Heptachlor epoxide	4.875	4.777	4.977	51.660	50.000	3.3
Methoxychlor	6.757	6.659	6.859	52.930	50.000	5.9
Tetrachloro-m-xylene	2.882	2.783	2.983	51.060	50.000	2.1