

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL02 Date Analyzed: 05/08/2025

Lab Sample No.: PSTDCCC050 Data File : PD088440.D Time Analyzed: 18:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.705	6.606	6.806	55.120	50.000	10.2
4,4'-DDE	6.196	6.097	6.297	53.580	50.000	7.2
4,4'-DDT	7.021	6.922	7.122	54.480	50.000	9.0
Aldrin	5.271	5.173	5.373	54.580	50.000	9.2
alpha-BHC	4.000	3.901	4.101	54.730	50.000	9.5
alpha-Chlordane	6.027	5.929	6.129	54.110	50.000	8.2
beta-BHC	4.515	4.416	4.616	52.990	50.000	6.0
Decachlorobiphenyl	9.074	8.975	9.175	52.380	50.000	4.8
delta-BHC	4.763	4.665	4.865	54.300	50.000	8.6
Dieldrin	6.347	6.249	6.449	54.450	50.000	8.9
Endosulfan I	6.074	5.976	6.176	53.850	50.000	7.7
Endosulfan II	6.786	6.687	6.887	53.120	50.000	6.2
Endosulfan sulfate	7.150	7.050	7.250	53.970	50.000	7.9
Endrin	6.574	6.476	6.676	54.410	50.000	8.8
Endrin aldehyde	6.915	6.816	7.016	54.450	50.000	8.9
Endrin ketone	7.630	7.531	7.731	54.500	50.000	9.0
gamma-BHC (Lindane)	4.330	4.232	4.432	53.830	50.000	7.7
gamma-Chlordane	5.946	5.847	6.047	53.960	50.000	7.9
Heptachlor	4.930	4.831	5.031	54.600	50.000	9.2
Heptachlor epoxide	5.691	5.592	5.792	53.570	50.000	7.1
Methoxychlor	7.493	7.395	7.595	53.870	50.000	7.7
Tetrachloro-m-xylene	3.550	3.452	3.652	52.440	50.000	4.9

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL02 Date Analyzed: 05/08/2025

Lab Sample No.: PSTDCCC050 Data File : PD088440.D Time Analyzed: 18:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.932	5.834	6.034	51.020	50.000	2.0
4,4'-DDE	5.376	5.280	5.480	50.630	50.000	1.3
4,4'-DDT	6.186	6.088	6.288	51.610	50.000	3.2
Aldrin	4.371	4.273	4.473	51.050	50.000	2.1
alpha-BHC	3.395	3.296	3.496	50.150	50.000	0.3
alpha-Chlordane	5.192	5.094	5.294	50.030	50.000	0.1
beta-BHC	4.027	3.928	4.128	49.560	50.000	-0.9
Decachlorobiphenyl	8.075	7.977	8.177	49.950	50.000	-0.1
delta-BHC	4.264	4.165	4.365	50.480	50.000	1.0
Dieldrin	5.515	5.417	5.617	50.490	50.000	1.0
Endosulfan I	5.249	5.151	5.351	48.260	50.000	-3.5
Endosulfan II	6.082	5.984	6.184	50.330	50.000	0.7
Endosulfan sulfate	6.484	6.386	6.586	50.330	50.000	0.7
Endrin	5.791	5.693	5.893	50.750	50.000	1.5
Endrin aldehyde	6.260	6.163	6.363	50.650	50.000	1.3
Endrin ketone	6.994	6.895	7.095	50.560	50.000	1.1
gamma-BHC (Lindane)	3.731	3.633	3.833	49.490	50.000	-1.0
gamma-Chlordane	5.128	5.030	5.230	50.160	50.000	0.3
Heptachlor	4.085	3.986	4.186	49.730	50.000	-0.5
Heptachlor epoxide	4.875	4.777	4.977	50.030	50.000	0.1
Methoxychlor	6.757	6.659	6.859	51.320	50.000	2.6
Tetrachloro-m-xylene	2.882	2.783	2.983	49.450	50.000	-1.1