

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL03 Date Analyzed: 05/06/2025

Lab Sample No.: PSTDCCC050 Data File : PL095562.D Time Analyzed: 12:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.750	6.653	6.853	47.930	50.000	-4.1
4,4'-DDE	6.241	6.143	6.343	44.580	50.000	-10.8
4,4'-DDT	7.065	6.969	7.169	46.230	50.000	-7.5
Aldrin	5.317	5.218	5.418	45.200	50.000	-9.6
alpha-BHC	4.046	3.946	4.146	42.420	50.000	-15.2
alpha-Chlordane	6.072	5.975	6.175	46.510	50.000	-7.0
beta-BHC	4.563	4.464	4.664	42.780	50.000	-14.4
Decachlorobiphenyl	9.119	9.022	9.222	43.410	50.000	-13.2
delta-BHC	4.810	4.712	4.912	43.730	50.000	-12.5
Dieldrin	6.394	6.296	6.496	45.950	50.000	-8.1
Endosulfan I	6.120	6.023	6.223	45.890	50.000	-8.2
Endosulfan II	6.832	6.735	6.935	49.680	50.000	-0.6
Endosulfan sulfate	7.195	7.098	7.298	47.500	50.000	-5.0
Endrin	6.620	6.523	6.723	46.610	50.000	-6.8
Endrin aldehyde	6.961	6.864	7.064	47.100	50.000	-5.8
Endrin ketone	7.677	7.579	7.779	46.190	50.000	-7.6
gamma-BHC (Lindane)	4.376	4.277	4.477	42.940	50.000	-14.1
gamma-Chlordane	5.991	5.894	6.094	47.060	50.000	-5.9
Heptachlor	4.975	4.876	5.076	43.410	50.000	-13.2
Heptachlor epoxide	5.736	5.639	5.839	47.050	50.000	-5.9
Methoxychlor	7.537	7.440	7.640	45.900	50.000	-8.2
Tetrachloro-m-xylene	3.595	3.495	3.695	42.360	50.000	-15.3

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL03 Date Analyzed: 05/06/2025

Lab Sample No.: PSTDCCC050 Data File : PL095562.D Time Analyzed: 12:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.952	5.849	6.049	54.080	50.000	8.2
4,4'-DDE	5.399	5.297	5.497	51.030	50.000	2.1
4,4'-DDT	6.206	6.104	6.304	53.070	50.000	6.1
Aldrin	4.393	4.290	4.490	53.940	50.000	7.9
alpha-BHC	3.422	3.316	3.516	55.360	50.000	10.7
alpha-Chlordane	5.214	5.111	5.311	52.930	50.000	5.9
beta-BHC	4.049	3.944	4.144	52.020	50.000	4.0
Decachlorobiphenyl	8.090	7.989	8.189	45.640	50.000	-8.7
delta-BHC	4.285	4.180	4.380	54.260	50.000	8.5
Dieldrin	5.534	5.432	5.632	54.340	50.000	8.7
Endosulfan I	5.270	5.167	5.367	47.690	50.000	-4.6
Endosulfan II	6.100	5.997	6.197	51.990	50.000	4.0
Endosulfan sulfate	6.502	6.398	6.598	51.270	50.000	2.5
Endrin	5.809	5.707	5.907	55.690	50.000	11.4
Endrin aldehyde	6.278	6.175	6.375	52.000	50.000	4.0
Endrin ketone	7.007	6.904	7.104	58.750	50.000	17.5
gamma-BHC (Lindane)	3.755	3.650	3.850	53.480	50.000	7.0
gamma-Chlordane	5.149	5.047	5.247	52.660	50.000	5.3
Heptachlor	4.109	4.004	4.204	51.890	50.000	3.8
Heptachlor epoxide	4.897	4.793	4.993	52.620	50.000	5.2
Methoxychlor	6.775	6.675	6.875	51.610	50.000	3.2
Tetrachloro-m-xylene	2.912	2.806	3.006	50.590	50.000	1.2