

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050725\
 Data File : PL095572.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2025 17:50
 Operator : AR\AJ
 Sample : Q1956-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 SB2-4-5

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 07 01:43:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042325.M
 Quant Title : GC Extractables
 QLast Update : Wed Apr 23 16:23:18 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.593	2.912	65106991	81116331	19.943	25.510 #
28)	SA Decachlor...	9.118	8.090	42691472	49042970	15.408	16.573
Target Compounds							
6)	B beta-BHC	0.000	4.020f	0	5513970	N.D.	2.656 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050725\
Data File : PL095572.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2025 17:50
Operator : AR\AJ
Sample : Q1956-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
SB2-4-5

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 07 01:43:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042325.M
Quant Title : GC Extractables
QLast Update : Wed Apr 23 16:23:18 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.593 min
Delta R.T.: -0.002 min
Response: 65106991
Conc: 19.94 ng/ml

Instrument :
ECD_L
ClientSampleId :
SB2-4-5




#1 Tetrachloro-m-xylene

R.T.: 2.912 min
Delta R.T.: 0.007 min
Response: 81116331
Conc: 25.51 ng/ml




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 4.564 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 4.020 min
Delta R.T.: -0.024 min
Response: 5513970
Conc: 2.66 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.118 min
Delta R.T.: -0.005 min
Response: 42691472
Conc: 15.41 ng/ml

Instrument :
ECD_L
ClientSampleId :
SB2-4-5



#28 Decachlorobiphenyl

R.T.: 8.090 min
Delta R.T.: 0.000 min
Response: 49042970
Conc: 16.57 ng/ml