

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL01 Date Analyzed: 05/12/2025

Lab Sample No.: HSTDCCC750 Data File : PS030138.D Time Analyzed: 18:58

| COMPOUND         | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|-------|-----------|-------|--------------------|-------------------|------|
|                  |       | FROM      | TO    |                    |                   |      |
| 2,4,5-TP(Silvex) | 8.851 | 8.750     | 8.950 | 717.430            | 712.500           | 0.7  |
| 2,4-D            | 8.009 | 7.910     | 8.110 | 700.620            | 705.000           | -0.6 |
| 2,4-DCAA         | 6.934 | 6.834     | 7.034 | 744.030            | 750.000           | -0.8 |

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL01 Date Analyzed: 05/12/2025

Lab Sample No.: HSTDCCC750 Data File : PS030138.D Time Analyzed: 18:58

| COMPOUND         | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D  |
|------------------|-------|----------------------|-------|--------------------|-------------------|-----|
| 2,4,5-TP(Silvex) | 9.523 | 9.423                | 9.623 | 737.880            | 712.500           | 3.6 |
| 2,4-D            | 8.645 | 8.544                | 8.744 | 715.360            | 705.000           | 1.5 |
| 2,4-DCAA         | 7.454 | 7.354                | 7.554 | 767.100            | 750.000           | 2.3 |