

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL02 Date Analyzed: 05/12/2025

Lab Sample No.: HSTDCCC750 Data File : PS030146.D Time Analyzed: 22:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	8.851	8.750	8.950	714.210	712.500	0.2
2,4-D	8.009	7.910	8.110	694.380	705.000	-1.5
2,4-DCAA	6.933	6.834	7.034	738.380	750.000	-1.5

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL02 Date Analyzed: 05/12/2025

Lab Sample No.: HSTDCCC750 Data File : PS030146.D Time Analyzed: 22:35

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.523	9.423	9.623	726.310	712.500	1.9
2,4-D	8.645	8.544	8.744	712.660	705.000	1.1
2,4-DCAA	7.454	7.354	7.554	755.150	750.000	0.7