

## **Report of Analysis**

Client:						Date Collected:			
Project:	Project: Bergen Point Fueling System					Date Received:			
Client Sample ID:	PB167968	BL				SDG No.:	Q1956		
Lab Sample ID:	PB167968	BL				Matrix:	TCLP		
Analytical Method:	8151A					% Solid:	0	Decanted:	
Sample Wt/Vol:	1000	Units:	mL			Final Vol:	10000	uL	
Soil Aliquot Vol:			uL			Test:	TCLP Herb	icide	
Extraction Type:						Injection Volume :			
GPC Factor :	1.0	]	PH :						
Prep Method :	SW3510C								
File ID/Qc Batch:	Dilution:		Prep Date			Date Analyzed	Prep Batch ID		
PS030139.D	1	1		12/25 12:00		05/12/25 19:46	PB167968		
CAS Number	Parameter		Conc.	Qualifier	MDL		LOQ/O	CRQL	Units
TARGETS									
94-75-7	2,4-D		2.00	U	0.92			2.00	ug/L
93-72-1	2,4,5-TP (Silvex)		2.00	U	0.78			2.00	ug/L
SURROGATES 19719-28-9	2,4-DCAA		540		39 - 175			108%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates > 25% difference for detected

concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

- J = Estimated Value
- B = Analyte Found in Associated Method Blank
- N = Presumptive Evidence of a Compound
- \* = Values outside of QC limits
- D = Dilution

 $\mathbf{S}=\mathbf{Indicates}$  estimated value where valid five-point calibration

was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit