

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL01 Date Analyzed: 05/13/2025

Lab Sample No.: PSTDCCC050 Data File : PD088499.D Time Analyzed: 09:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.074	8.975	9.175	48.950	50.000	-2.1
Endrin	6.576	6.476	6.676	53.150	50.000	6.3
gamma-BHC (Lindane)	4.332	4.232	4.432	53.830	50.000	7.7
Heptachlor	4.931	4.831	5.031	55.670	50.000	11.3
Heptachlor epoxide	5.692	5.592	5.792	54.860	50.000	9.7
Methoxychlor	7.494	7.395	7.595	51.790	50.000	3.6
Tetrachloro-m-xylene	3.552	3.452	3.652	51.960	50.000	3.9

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL01 Date Analyzed: 05/13/2025

Lab Sample No.: PSTDCCC050 Data File : PD088499.D Time Analyzed: 09:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.075	7.977	8.177	42.160	50.000	-15.7
Endrin	5.791	5.693	5.893	46.530	50.000	-6.9
gamma-BHC (Lindane)	3.731	3.633	3.833	47.220	50.000	-5.6
Heptachlor	4.085	3.986	4.186	47.280	50.000	-5.4
Heptachlor epoxide	4.875	4.777	4.977	47.370	50.000	-5.3
Methoxychlor	6.757	6.659	6.859	45.070	50.000	-9.9
Tetrachloro-m-xylene	2.882	2.783	2.983	47.090	50.000	-5.8