

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL02 Date Analyzed: 05/13/2025

Lab Sample No.: PSTDCCC050 Data File : PD088515.D Time Analyzed: 14:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.080	8.975	9.175	48.470	50.000	-3.1
Endrin	6.582	6.476	6.676	53.290	50.000	6.6
gamma-BHC (Lindane)	4.338	4.232	4.432	55.030	50.000	10.1
Heptachlor	4.937	4.831	5.031	55.970	50.000	11.9
Heptachlor epoxide	5.698	5.592	5.792	54.730	50.000	9.5
Methoxychlor	7.500	7.395	7.595	51.030	50.000	2.1
Tetrachloro-m-xylene	3.558	3.452	3.652	52.540	50.000	5.1

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL02 Date Analyzed: 05/13/2025

Lab Sample No.: PSTDCCC050 Data File : PD088515.D Time Analyzed: 14:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.076	7.977	8.177	41.330	50.000	-17.3
Endrin	5.792	5.693	5.893	47.220	50.000	-5.6
gamma-BHC (Lindane)	3.732	3.633	3.833	47.870	50.000	-4.3
Heptachlor	4.086	3.986	4.186	47.490	50.000	-5.0
Heptachlor epoxide	4.876	4.777	4.977	47.790	50.000	-4.4
Methoxychlor	6.758	6.659	6.859	45.030	50.000	-9.9
Tetrachloro-m-xylene	2.882	2.783	2.983	47.460	50.000	-5.1