

Cover Page

Order ID : Q1956

Project ID : Bergen Point Fueling System

Client : CDM Smith

Lab Sample Number

Q1956-01
Q1956-02
Q1956-03
Q1956-04
Q1956-05
Q1956-06
Q1956-07
Q1956-09

Client Sample Number

SB1-3-4
SB2-4-5
Q1956-02MS
Q1956-02MSD
COMP1
SB91-3-4
FB-05022025
TB

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 5/15/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following "Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1956

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: MAHESH PATEL

Date: 05/15/2025

LAB CHRONICLE

OrderID:	Q1956	OrderDate:	5/2/2025 3:14:35 PM
Client:	CDM Smith	Project:	Bergen Point Fueling System
Contact:	Marcie Ann Encinas	Location:	L31,VOA Ref. #2 Soil,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1956-01	SB1-3-4	SOIL			05/01/25			05/02/25
			Diesel Range Organics	8015D		05/13/25	05/13/25	
			Gasoline Range Organics	8015D			05/06/25	
			Herbicide	8151A		05/07/25	05/08/25	
			PCB	8082A		05/06/25	05/06/25	
			Pesticide-TCL	8081B		05/06/25	05/06/25	
Q1956-02	SB2-4-5	SOIL			05/02/25			05/02/25
			Diesel Range Organics	8015D		05/13/25	05/13/25	
			Gasoline Range Organics	8015D			05/06/25	
			Herbicide	8151A		05/07/25	05/08/25	
			PCB	8082A		05/06/25	05/06/25	
			Pesticide-TCL	8081B		05/06/25	05/06/25	
Q1956-05	COMP1	TCLP			05/02/25			05/02/25
			TCLP Herbicide	8151A		05/12/25	05/12/25	
			TCLP Pesticide	8081B		05/12/25	05/13/25	
Q1956-06	SB91-3-4	SOIL			05/01/25			05/02/25
			Diesel Range Organics	8015D		05/13/25	05/13/25	
			Gasoline Range Organics	8015D			05/06/25	
			Herbicide	8151A		05/07/25	05/08/25	
			PCB	8082A		05/06/25	05/06/25	
			Pesticide-TCL	8081B		05/06/25	05/06/25	
Q1956-07	FB-05022025	Water			05/02/25			05/02/25
			Diesel Range Organics	8015D		05/08/25	05/08/25	
			Gasoline Range Organics	8015D			05/06/25	
			Herbicide	8151A		05/08/25	05/08/25	
			PCB	8082A		05/07/25	05/07/25	
			Pesticide-TCL	8081B		05/08/25	05/08/25	

Hit Summary Sheet
SW-846

SDG No.: Q1956

Order ID: Q1956

Client: CDM Smith

Project ID: Bergen Point Fueling System

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000



QC SUMMARY

Surrogate Summary

SDG No.: Q1956

Client: CDM Smith

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088142.D	PIBLK-PD088142.D	Decachlorobiphenyl	1	20	23.2	116		43	140
		Tetrachloro-m-xylene	1	20	20.8	104		77	126
		Decachlorobiphenyl	2	20	23.3	116		43	140
		Tetrachloro-m-xylene	2	20	21.4	107		77	126
I.BLK-PD088497.D	PIBLK-PD088497.D	Decachlorobiphenyl	1	20	19.7	99		43	140
		Tetrachloro-m-xylene	1	20	18.5	92		77	126
		Decachlorobiphenyl	2	20	17.4	87		43	140
		Tetrachloro-m-xylene	2	20	17.6	88		77	126
PB167969BL	PB167969BL	Decachlorobiphenyl	1	20	21.2	106		43	140
		Tetrachloro-m-xylene	1	20	21.0	105		77	126
		Decachlorobiphenyl	2	20	16.6	83		43	140
		Tetrachloro-m-xylene	2	20	20.7	104		77	126
PB167969BS	PB167969BS	Decachlorobiphenyl	1	20	21.3	107		43	140
		Tetrachloro-m-xylene	1	20	21.2	106		77	126
		Decachlorobiphenyl	2	20	17.7	88		43	140
		Tetrachloro-m-xylene	2	20	21.1	105		77	126
PB167851TB	PB167851TB	Decachlorobiphenyl	1	20	21.7	109		43	140
		Tetrachloro-m-xylene	1	20	21.6	108		77	126
		Decachlorobiphenyl	2	20	17.7	88		43	140
		Tetrachloro-m-xylene	2	20	21.2	106		77	126
Q1956-05	COMP1	Decachlorobiphenyl	1	20	21.9	110		43	140
		Tetrachloro-m-xylene	1	20	21.5	107		77	126
		Decachlorobiphenyl	2	20	17.6	88		43	140
		Tetrachloro-m-xylene	2	20	19.1	96		77	126
Q1956-05MS	COMP1MS	Decachlorobiphenyl	1	20	21.6	108		43	140
		Tetrachloro-m-xylene	1	20	21.8	109		77	126
		Decachlorobiphenyl	2	20	17.9	89		43	140
		Tetrachloro-m-xylene	2	20	19.2	96		77	126
Q1956-05MSD	COMP1MSD	Decachlorobiphenyl	1	20	21.6	108		43	140
		Tetrachloro-m-xylene	1	20	22.3	111		77	126
		Decachlorobiphenyl	2	20	17.9	90		43	140
		Tetrachloro-m-xylene	2	20	18.9	95		77	126
I.BLK-PD088514.D	PIBLK-PD088514.D	Decachlorobiphenyl	1	20	18.7	94		43	140
		Tetrachloro-m-xylene	1	20	18.1	90		77	126
		Decachlorobiphenyl	2	20	15.7	79		43	140
		Tetrachloro-m-xylene	2	20	16.9	85		77	126

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1956

Client: CDM Smith

Analytical Method: 8081B

DataFile : PD088512.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	COMP1MS											
Q1956-05MS	gamma-BHC (Lindane)	5	0	6.10	ug/L	122				60	152	
	Heptachlor	5	0	6.00	ug/L	120				56	147	
	Heptachlor epoxide	5	0	6.10	ug/L	122				77	143	
	Endrin	5	0	6.00	ug/L	120				76	144	
	Methoxychlor	5	0	5.60	ug/L	112				70	142	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1956

Client: CDM Smith

Analytical Method: 8081B

DataFile : PD088513.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec Qual	RPD Qual	Low	Limits High	RPD
Client Sample ID:	COMP1MSD									
Q1956-05MSD	gamma-BHC (Lindane)	5	0	6.00	ug/L	120	2	60	152	20
	Heptachlor	5	0	6.00	ug/L	120	0	56	147	20
	Heptachlor epoxide	5	0	6.10	ug/L	122	0	77	143	20
	Endrin	5	0	6.00	ug/L	120	0	76	144	20
	Methoxychlor	5	0	5.60	ug/L	112	0	70	142	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1956

Client: CDM Smith

Analytical Method: **8081B** **Datafile :** PD088509.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167969BS	gamma-BHC (Lindane)	0.5	0.51	ug/L	103				82	129	
	Heptachlor	0.5	0.53	ug/L	106				79	127	
	Heptachlor epoxide	0.5	0.52	ug/L	104				81	124	
	Endrin	0.5	0.48	ug/L	97				81	128	
	Methoxychlor	0.5	0.47	ug/L	94				78	108	

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167969BL

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956

SAS No.: Q1956 SDG NO.: Q1956

Lab Sample ID: PB167969BL

Lab File ID: PD088508.D

Matrix: (soil/water) water

Extraction: (Type) SEPF

Sulfur Cleanup: (Y/N) N

Date Extracted: 05/12/2025

Date Analyzed (1): 05/13/2025

Date Analyzed (2): 05/13/2025

Time Analyzed (1): 11:57

Time Analyzed (2): 11:57

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB167969BS	PB167969BS	PD088509.D	05/13/2025	05/13/2025
PB167851TB	PB167851TB	PD088510.D	05/13/2025	05/13/2025
COMP1	Q1956-05	PD088511.D	05/13/2025	05/13/2025
COMP1MS	Q1956-05MS	PD088512.D	05/13/2025	05/13/2025
COMP1MSD	Q1956-05MSD	PD088513.D	05/13/2025	05/13/2025

COMMENTS: _____



SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Bergen Point Fueling System	Date Received:	05/12/25
Client Sample ID:	PB167851TB	SDG No.:	Q1956
Lab Sample ID:	PB167851TB	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	100 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088510.D	1	05/12/25 12:25	05/13/25 12:24	PB167969

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.50	U	0.037	0.50	ug/L
76-44-8	Heptachlor	0.50	U	0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	0.50	U	0.096	0.50	ug/L
72-20-8	Endrin	0.50	U	0.032	0.50	ug/L
72-43-5	Methoxychlor	0.50	U	0.11	0.50	ug/L
8001-35-2	Toxaphene	10.0	U	1.70	10.0	ug/L
57-74-9	Chlordane	5.00	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	21.7		43 - 140	109%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.6		77 - 126	108%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
 Data File : PD088510.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2025 12:24
 Operator : AR\AJ
 Sample : PB167851TB
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB167851TB

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 22:40:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.552	2.883	43130125	309.6E6	21.593	21.174
28) SA Decachlor...	9.074	8.074	71932047	326.4E6	21.744	17.661

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
Data File : PD088510.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2025 12:24
Operator : AR\AJ
Sample : PB167851TB
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB167851TB

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 22:40:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.552 min
Delta R.T.: 0.000 min
Response: 43130125
Conc: 21.59 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB167851TB




#1 Tetrachloro-m-xylene

R.T.: 2.883 min
Delta R.T.: 0.000 min
Response: 309604326
Conc: 21.17 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.074 min
Delta R.T.: 0.000 min
Response: 71932047
Conc: 21.74 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.074 min
Delta R.T.: -0.002 min
Response: 326369834
Conc: 17.66 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/02/25
Project:	Bergen Point Fueling System	Date Received:	05/02/25
Client Sample ID:	COMP1	SDG No.:	Q1956
Lab Sample ID:	Q1956-05	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	TCLP Pesticide
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088511.D	1	05/12/25 12:25	05/13/25 12:37	PB167969

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.50	U	0.037	0.50	ug/L
76-44-8	Heptachlor	0.50	U	0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	0.50	U	0.096	0.50	ug/L
72-20-8	Endrin	0.50	U	0.032	0.50	ug/L
72-43-5	Methoxychlor	0.50	U	0.11	0.50	ug/L
8001-35-2	Toxaphene	10.0	U	1.70	10.0	ug/L
57-74-9	Chlordane	5.00	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	21.9		43 - 140	110%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.5		77 - 126	107%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
 Data File : PD088511.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2025 12:37
 Operator : AR\AJ
 Sample : Q1956-05
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 COMP1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 22:40:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.552	2.883	42877793	279.6E6	21.467	19.119
28) SA Decachlor...	9.075	8.075	72553667	325.8E6	21.932	17.630

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
Data File : PD088511.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2025 12:37
Operator : AR\AJ
Sample : Q1956-05
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
COMP1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 22:40:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.552 min
Delta R.T.: 0.000 min
Response: 42877793
Conc: 21.47 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP1




#1 Tetrachloro-m-xylene

R.T.: 2.883 min
Delta R.T.: 0.000 min
Response: 279554151
Conc: 19.12 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.075 min
Delta R.T.: 0.000 min
Response: 72553667
Conc: 21.93 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.075 min
Delta R.T.: -0.002 min
Response: 325803897
Conc: 17.63 ng/ml



CALIBRATION SUMMARY

RETENTION TIMES OF INITIAL CALIBRATION

Contract:	<u>CAMP02</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1956</u>	SAS No.:	<u>Q1956</u>	SDG NO.:	<u>Q1956</u>
Instrument ID:	<u>ECD_D</u>	Calibration Date(s):		<u>04/18/2025</u>	<u>04/18/2025</u>		
		Calibration Times:		<u>13:56</u>	<u>14:51</u>		

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 100 =	<u>PD088124.D</u>	RT 075 =	<u>PD088125.D</u>
	RT 050 =	PD088126.D	RT 025 =	PD088127.D
			RT 005 =	PD088128.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Contract: CAMP02
Lab Code: CHEM **Case No.:** Q1956 **SAS No.:** Q1956 **SDG NO.:** Q1956
Instrument ID: ECD_D **Calibration Date(s):** 04/18/2025 04/18/2025
Calibration Times: 13:56 14:51

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 100 = <u>PD088124.D</u>	RT 075 = <u>PD088125.D</u>
RT 050 = <u>PD088126.D</u>	RT 025 = <u>PD088127.D</u>	RT 005 = <u>PD088128.D</u>

COMPOUND	RT 100	RT 075	RT 050	RT 025	RT 005	MEAN RT	RT WINDOW	
							FROM	TO
Decachlorobiphenyl	8.09	8.08	8.08	8.08	8.08	8.08	7.98	8.18
Endrin	5.81	5.79	5.79	5.79	5.79	5.80	5.70	5.90
gamma-BHC (Lindane)	3.75	3.73	3.73	3.73	3.73	3.74	3.64	3.84
Heptachlor	4.10	4.09	4.09	4.09	4.09	4.09	3.99	4.19
Heptachlor epoxide	4.89	4.88	4.88	4.88	4.88	4.88	4.78	4.98
Methoxychlor	6.77	6.76	6.76	6.76	6.76	6.76	6.66	6.86
Tetrachloro-m-xylene	2.90	2.88	2.88	2.88	2.88	2.89	2.79	2.99



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Instrument ID: ECD_D

Calibration Date(s): 04/18/2025 04/18/2025

Calibration Times: 13:56 14:51

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD088124.D</u>	CF 075 =	<u>PD088125.D</u>		
CF 050 =		<u>PD088126.D</u>	CF 025 =	<u>PD088127.D</u>	CF 005 =	<u>PD088128.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
Decachlorobiphenyl	3080820000	3141130000	3178140000	3290360000	3850090000	3308110000	9
Endrin	3128180000	3077300000	2935450000	2806460000	2965440000	2982570000	4
gamma-BHC (Lindane)	4507210000	4365130000	4155030000	3887470000	4022700000	4187510000	6
Heptachlor	4264490000	4159110000	3978190000	3781390000	4011590000	4038950000	5
Heptachlor epoxide	3670910000	3598060000	3486330000	3389810000	3724620000	3573950000	4
Methoxychlor	1461300000	1466360000	1467850000	1483290000	1618930000	1499540000	4
Tetrachloro-m-xylene	1982340000	2006790000	1938680000	1923660000	2135510000	1997400000	4



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Instrument ID: ECD_D

Calibration Date(s): 04/18/2025 04/18/2025

Calibration Times: 13:56 14:51

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD088124.D</u>	CF 075 =	<u>PD088125.D</u>		
CF 050 =		<u>PD088126.D</u>	CF 025 =	<u>PD088127.D</u>	CF 005 =	<u>PD088128.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
Decachlorobiphenyl	16767000000	17098200000	17470300000	18387600000	22674800000	18479600000	13
Endrin	16756200000	17001500000	17464500000	18215500000	21574100000	18202300000	11
gamma-BHC (Lindane)	20140800000	20226900000	20558100000	21137200000	25630400000	21538700000	11
Heptachlor	19826900000	20021500000	20522700000	21298800000	25320700000	21398100000	11
Heptachlor epoxide	17359900000	17576100000	18079700000	18879200000	22576700000	18894300000	11
Methoxychlor	8272330000	8467510000	8839240000	9290490000	10735700000	9121060000	11
Tetrachloro-m-xylene	13615300000	13685800000	14010800000	14551200000	17245000000	14621600000	10



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Instrument ID: ECD_D Date(s) Analyzed: 04/18/2025 04/18/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	4.72	4.62	4.82	161995000
		2	5.24	5.14	5.34	160780000
		3	5.95	5.85	6.05	666185000
		4	6.03	5.93	6.13	797101000
		5	6.87	6.77	6.97	133051000
Toxaphene	500	1	6.24	6.14	6.34	23597900
		2	6.44	6.34	6.54	34064600
		3	7.15	7.05	7.25	65326300
		4	7.57	7.47	7.67	82045400
		5	7.93	7.83	8.03	47715200



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Instrument ID: ECD_D Date(s) Analyzed: 04/18/2025 04/18/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	3.91	3.81	4.01	767527000
		2	4.49	4.39	4.59	783082000
		3	5.13	5.03	5.23	2396830000
		4	5.19	5.09	5.29	2031740000
		5	6.09	5.99	6.19	931551000
Toxaphene	500	1	5.48	5.38	5.58	130682000
		2	5.65	5.55	5.75	89513500
		3	6.76	6.66	6.86	417352000
		4	7.20	7.10	7.30	302911000
		5	7.33	7.23	7.43	207761000

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
 Data File : PD088124.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2025 13:56
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDICC100

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 04/19/2025

Supervised By :mohammad ahmed 04/22/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 04:25:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 04:25:16 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.551	2.900	198.2E6	1361.5E6	101.114	98.568
28) SA Decachlor...	9.075	8.093	308.1E6	1676.7E6	98.445	97.946
Target Compounds						
2) A alpha-BHC	4.001	3.413	478.3E6	2191.3E6	104.846	99.598
3) MA gamma-BHC...	4.332	3.749	450.7E6	2014.1E6	104.066	98.975
4) MA Heptachlor	4.931	4.103	426.4E6	1982.7E6	103.473	98.276
5) MB Aldrin	5.273	4.389	419.1E6	1944.0E6	103.451	98.579
6) B beta-BHC	4.516	4.045	160.4E6	847.4E6	100.801	97.843
7) B delta-BHC	4.765	4.282	444.5E6	2017.9E6	104.335	99.275
8) B Heptachlo...	5.693	4.892	367.1E6	1736.0E6	102.579	98.354m
9) A Endosulfan I	6.076	5.268	345.0E6	1644.8E6	102.166	97.631
10) B gamma-Chl...	5.947	5.146	375.8E6	1896.0E6	103.034	98.757
11) B alpha-Chl...	6.029	5.211	371.0E6	1819.3E6	102.498	98.477
12) B 4,4'-DDE	6.197	5.396	346.7E6	1834.5E6	103.381	98.584
13) MA Dieldrin	6.349	5.533	375.0E6	1853.7E6	103.019	98.382
14) MA Endrin	6.576	5.810	312.8E6	1675.6E6	103.179	97.930
15) B Endosulfa...	6.787	6.101	310.4E6	1601.0E6	101.581	97.777
16) A 4,4'-DDD	6.706	5.950	265.8E6	1515.5E6	103.156	97.940
17) MA 4,4'-DDT	7.023	6.204	292.3E6	1643.2E6	102.967	99.053
18) B Endrin al...	6.917	6.279	229.5E6	1206.4E6	101.060	97.405
19) B Endosulfa...	7.151	6.502	289.7E6	1562.9E6	101.567	97.534
20) A Methoxychlor	7.494	6.774	146.1E6	827.2E6	99.776	96.687
21) B Endrin ke...	7.631	7.011	312.9E6	1699.7E6	101.477	97.396
22) Mirex	8.115	7.205	219.9E6	1333.1E6	98.857	97.475

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
Data File : PD088124.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2025 13:56
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDICC100

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 04/19/2025

Supervised By :mohammad ahmed 04/22/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 04:25:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 04:25:16 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
 Data File : PD088125.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2025 14:10
 Operator : AR\AJ
 Sample : PSTDICC075
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 04:38:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 04:38:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.552	2.883	150.5E6	1026.4E6	76.770	74.309
28)	SA Decachlor...	9.074	8.076	235.6E6	1282.4E6	75.279	74.910
Target Compounds							
2)	A alpha-BHC	4.001	3.396	344.5E6	1637.3E6	75.524	74.417
3)	MA gamma-BHC...	4.332	3.732	327.4E6	1517.0E6	75.589	74.548
4)	MA Heptachlor	4.931	4.086	311.9E6	1501.6E6	75.687	74.430
5)	MB Aldrin	5.273	4.372	305.2E6	1470.4E6	75.338	74.562
6)	B beta-BHC	4.516	4.028	119.2E6	643.1E6	74.880	74.260
7)	B delta-BHC	4.765	4.265	321.2E6	1514.3E6	75.391	74.502
8)	B Heptachlo...	5.693	4.876	269.9E6	1318.2E6	75.407	74.392
9)	A Endosulfan I	6.076	5.251	254.8E6	1255.1E6	75.455	74.495
10)	B gamma-Chl...	5.948	5.130	277.4E6	1429.7E6	76.052	74.470
11)	B alpha-Chl...	6.029	5.195	271.9E6	1376.2E6	75.106	74.493
12)	B 4,4'-DDE	6.197	5.379	264.5E6	1393.5E6	78.883	74.883
13)	MA Dieldrin	6.349	5.517	274.2E6	1403.5E6	75.327	74.488
14)	MA Endrin	6.576	5.793	230.8E6	1275.1E6	76.125	74.523
15)	B Endosulfa...	6.788	6.084	229.6E6	1217.8E6	75.122	74.374
16)	A 4,4'-DDD	6.706	5.934	194.0E6	1155.3E6	75.311	74.663
17)	MA 4,4'-DDT	7.022	6.188	215.1E6	1237.2E6	75.763	74.584
18)	B Endrin al...	6.916	6.263	170.1E6	920.9E6	74.896	74.348
19)	B Endosulfa...	7.151	6.486	213.5E6	1192.6E6	74.839	74.424
20)	A Methoxychlor	7.494	6.759	110.0E6	635.1E6	75.092	74.226
21)	B Endrin ke...	7.631	6.995	230.5E6	1294.9E6	74.768	74.201
22)	Mirex	8.115	7.189	164.0E6	1014.9E6	73.732	74.208



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
Data File : PD088125.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2025 14:10
Operator : AR\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 04:38:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 04:38:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
 Data File : PD088126.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2025 14:24
 Operator : AR\AJ
 Sample : PSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 03:52:56 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 03:51:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.552	2.883	96933779	700.5E6	50.000	50.000
28) SA Decachlor...	9.075	8.077	158.9E6	873.5E6	50.000	50.000
Target Compounds						
2) A alpha-BHC	4.001	3.396	217.0E6	1104.5E6	50.000	50.000
3) MA gamma-BHC...	4.332	3.733	207.8E6	1027.9E6	50.000	50.000
4) MA Heptachlor	4.931	4.086	198.9E6	1026.1E6	50.000	50.000
5) MB Aldrin	5.273	4.373	195.6E6	1000.0E6	50.000	50.000
6) B beta-BHC	4.516	4.028	78949582	442.4E6	50.000	50.000
7) B delta-BHC	4.765	4.265	203.8E6	1023.7E6	50.000	50.000
8) B Heptachlo...	5.692	4.877	174.3E6	904.0E6	50.000	50.000
9) A Endosulfan I	6.076	5.251	165.2E6	862.3E6	50.000	50.000
10) B gamma-Chl...	5.947	5.130	176.8E6	971.9E6	50.000	50.000
11) B alpha-Chl...	6.029	5.194	176.5E6	937.8E6	50.000	50.000
12) B 4,4'-DDE	6.197	5.380	162.0E6	943.6E6	50.000	50.000
13) MA Dieldrin	6.349	5.517	176.5E6	957.3E6	50.000	50.000
14) MA Endrin	6.576	5.793	146.8E6	873.2E6	50.000	50.000
15) B Endosulfa...	6.787	6.084	150.4E6	836.9E6	50.000	50.000
16) A 4,4'-DDD	6.706	5.934	124.8E6	789.6E6	50.000	50.000
17) MA 4,4'-DDT	7.022	6.188	137.8E6	837.3E6	50.000	50.000
18) B Endrin al...	6.916	6.263	112.3E6	635.4E6	50.000	50.000
19) B Endosulfa...	7.150	6.486	140.4E6	821.0E6	50.000	50.000
20) A Methoxychlor	7.495	6.759	73392410	442.0E6	50.000	50.000
21) B Endrin ke...	7.631	6.995	151.9E6	895.3E6	50.000	50.000
22) Mirex	8.115	7.190	112.5E6	701.1E6	50.000	50.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
Data File : PD088126.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2025 14:24
Operator : AR\AJ
Sample : PSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 03:52:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 03:51:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
 Data File : PD088127.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2025 14:37
 Operator : AR\AJ
 Sample : PSTDICC025
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 03:53:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 03:51:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.551	2.883	48091567	363.8E6	24.806	25.964
28)	SA Decachlor...	9.075	8.076	82258935	459.7E6	25.883	26.313
Target Compounds							
2)	A alpha-BHC	4.001	3.396	99341615	563.0E6	22.886	25.485
3)	MA gamma-BHC...	4.332	3.733	97186739	528.4E6	23.390	25.704
4)	MA Heptachlor	4.932	4.087	94534629	532.5E6	23.763	25.945
5)	MB Aldrin	5.273	4.373	92978704	517.9E6	23.769	25.894
6)	B beta-BHC	4.516	4.028	39603540	232.3E6	25.082	26.257
7)	B delta-BHC	4.765	4.265	96272367	525.8E6	23.621	25.682
8)	B Heptachlo...	5.693	4.877	84745348	472.0E6	24.308	26.106
9)	A Endosulfan I	6.076	5.251	80453509	451.6E6	24.350	26.184
10)	B gamma-Chl...	5.948	5.130	85075416	502.2E6	24.058	25.835
11)	B alpha-Chl...	6.029	5.194	85660342	486.4E6	24.270	25.934
12)	B 4,4'-DDE	6.198	5.380	76777577	489.5E6	23.696	25.940
13)	MA Dieldrin	6.349	5.517	84286018	497.2E6	23.874	25.966
14)	MA Endrin	6.576	5.793	70161483	455.4E6	23.901	26.075
15)	B Endosulfa...	6.788	6.085	74367338	438.3E6	24.727	26.186
16)	A 4,4'-DDD	6.707	5.934	59400090	409.0E6	23.808	25.900
17)	MA 4,4'-DDT	7.023	6.188	65746503	426.6E6	23.864	25.475
18)	B Endrin al...	6.917	6.263	55861887	333.7E6	24.862	26.261
19)	B Endosulfa...	7.151	6.487	69350501	430.0E6	24.701	26.189
20)	A Methoxychlor	7.495	6.759	37082247	232.3E6	25.263	26.276
21)	B Endrin ke...	7.632	6.995	74459471	469.5E6	24.513	26.222
22)	Mirex	8.116	7.190	58667661	372.6E6	26.079	26.570



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
Data File : PD088127.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2025 14:37
Operator : AR\AJ
Sample : PSTDICC025
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 03:53:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 03:51:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
 Data File : PD088128.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2025 14:51
 Operator : AR\AJ
 Sample : PSTDICC005
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDICC005

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 04/19/2025

Supervised By :mohammad ahmed 04/22/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 03:53:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 03:51:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.552	2.883	10677571	86225193	5.508	6.154
28) SA Decachlor...	9.075	8.076	19250443	113.4E6	6.057	6.490
Target Compounds						
2) A alpha-BHC	4.001	3.396	19344032	129.8E6	4.456	5.874 #
3) MA gamma-BHC...	4.332	3.733	20113508	128.2E6	4.841	6.234 #
4) MA Heptachlor	4.931	4.086	20057945	126.6E6	5.042	6.169
5) MB Aldrin	5.273	4.372	19278191	121.3E6	4.928	6.063
6) B beta-BHC	4.516	4.028	8800396	55407278	5.573	6.262
7) B delta-BHC	4.764	4.265	19851779	122.2E6	4.871m	5.967
8) B Heptachlo...	5.693	4.876	18623118	112.9E6	5.342	6.244
9) A Endosulfan I	6.076	5.251	17577772	107.8E6	5.320	6.253
10) B gamma-Chl...	5.948	5.130	18576714	119.7E6	5.253	6.157
11) B alpha-Chl...	6.029	5.194	18782686	116.3E6	5.322	6.201
12) B 4,4'-DDE	6.197	5.379	15929607	115.5E6	4.916	6.117
13) MA Dieldrin	6.349	5.517	17671896	116.9E6	5.006	6.106
14) MA Endrin	6.576	5.793	14827210	107.9E6	5.051	6.177
15) B Endosulfa...	6.787	6.084	17419714	105.5E6	5.792	6.301
16) A 4,4'-DDD	6.706	5.933	12295875	96096151	4.928	6.085
17) MA 4,4'-DDT	7.022	6.188	13556042	91724476	4.921	5.478
18) B Endrin al...	6.916	6.262	12478868	80534454	5.554	6.338
19) B Endosulfa...	7.151	6.486	15272374	102.6E6	5.440	6.248
20) A Methoxychlor	7.495	6.758	8094629	53678604	5.515	6.073
21) B Endrin ke...	7.631	6.995	16074498	112.9E6	5.292	6.304
22) Mirex	8.116	7.189	13799458	91589011	6.134	6.532

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
Data File : PD088128.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2025 14:51
Operator : AR\AJ
Sample : PSTDICC005
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDICC005

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 04/19/2025

Supervised By :mohammad ahmed 04/22/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 03:53:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 03:51:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
 Data File : PD088131.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2025 15:32
 Operator : AR\AJ
 Sample : PCHLORICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PCHLORICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 05:48:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 05:47:09 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.552	2.883	96235160	865.8E6	50.000	50.000
28)	SA Decachlor...	9.074	8.077	154.6E6	878.3E6	50.000	50.000
Target Compounds							
23)	Chlordane-1	4.717	3.909	80997430	383.8E6	500.000	500.000
24)	Chlordane-2	5.243	4.491	80390039	391.5E6	500.000	500.000
25)	Chlordane-3	5.948	5.130	333.1E6	1198.4E6	500.000	500.000
26)	Chlordane-4	6.034	5.194	398.6E6	1015.9E6	500.000	500.000
27)	Chlordane-5	6.873	6.094	66525644	465.8E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
Data File : PD088131.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2025 15:32
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 05:48:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 05:47:09 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
 Data File : PD088136.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2025 16:40
 Operator : AR\AJ
 Sample : PTOXICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PTOXICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 06:17:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:16:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds						
1) SA Tetrachlo...	3.552	2.883	101.0E6	721.8E6	50.000	50.000
7) SA Decachlor...	9.074	8.077	164.9E6	890.5E6	50.000	50.000
Target Compounds						
2) Toxaphene-1	6.243	5.479	11798963	65341019	500.000	500.000
3) Toxaphene-2	6.443	5.650	17032300	44756730	500.000	500.000
4) Toxaphene-3	7.149	6.760	32663151	208.7E6	500.000	500.000
5) Toxaphene-4	7.565	7.202	41022693	151.5E6	500.000	500.000
6) Toxaphene-5	7.931	7.333	23857581	103.9E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
Data File : PD088136.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2025 16:40
Operator : AR\AJ
Sample : PTOXICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 06:17:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:16:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
 Data File : PD088139.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2025 17:21
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD041825

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 03:53:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 03:51:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.552	2.883	98011498	705.8E6	50.556	50.375
28) SA Decachlor...	9.074	8.076	156.5E6	873.1E6	49.237	49.975
Target Compounds						
2) A alpha-BHC	4.001	3.396	221.4E6	1115.2E6	51.003	50.483
3) MA gamma-BHC...	4.332	3.732	211.2E6	1037.4E6	50.823	50.460
4) MA Heptachlor	4.931	4.086	202.0E6	1034.4E6	50.784	50.400
5) MB Aldrin	5.273	4.372	199.5E6	1010.7E6	50.990	50.536
6) B beta-BHC	4.516	4.028	80388581	446.0E6	50.911	50.415
7) B delta-BHC	4.765	4.265	207.4E6	1030.8E6	50.892	50.347
8) B Heptachlo...	5.693	4.876	177.5E6	911.0E6	50.911	50.388
9) A Endosulfan I	6.076	5.250	167.8E6	870.5E6	50.788	50.475
10) B gamma-Chl...	5.947	5.129	179.7E6	979.0E6	50.819	50.365
11) B alpha-Chl...	6.028	5.194	179.5E6	945.1E6	50.851	50.389
12) B 4,4'-DDE	6.197	5.379	162.4E6	948.3E6	50.132	50.246
13) MA Dieldrin	6.349	5.516	179.4E6	966.1E6	50.821	50.456
14) MA Endrin	6.576	5.792	148.1E6	879.3E6	50.462	50.348
15) B Endosulfa...	6.787	6.084	151.1E6	841.2E6	50.228	50.256
16) A 4,4'-DDD	6.706	5.933	124.5E6	794.4E6	49.895	50.303
17) MA 4,4'-DDT	7.022	6.187	137.1E6	844.1E6	49.781	50.406
18) B Endrin al...	6.916	6.262	112.7E6	640.6E6	50.151	50.415
19) B Endosulfa...	7.150	6.486	140.8E6	824.1E6	50.153	50.192
20) A Methoxychlor	7.494	6.758	73171542	444.8E6	49.850	50.322
21) B Endrin ke...	7.631	6.994	151.9E6	901.7E6	50.003	50.358
22) Mirex	8.115	7.189	112.8E6	707.4E6	50.149	50.451



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
Data File : PD088139.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2025 17:21
Operator : AR\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
ICVPD041825

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 03:53:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 03:51:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
 Data File : PD088140.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2025 17:35
 Operator : AR\AJ
 Sample : PCHLORICV500
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD041825CHLOR

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 05:48:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 05:47:09 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.552	2.883	98326390	875.4E6	51.087	50.557
28) SA Decachlor...	9.073	8.076	155.6E6	883.8E6	50.337	50.312
Target Compounds						
23) Chlordane-1	4.716	3.908	83912348	391.8E6	517.994	510.427
24) Chlordane-2	5.243	4.490	83436128	397.6E6	518.946	507.710
25) Chlordane-3	5.948	5.129	344.9E6	1215.6E6	517.682	507.183
26) Chlordane-4	6.034	5.193	413.5E6	1029.1E6	518.799	506.534
27) Chlordane-5	6.872	6.093	67686335	471.4E6	508.724	506.043



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
Data File : PD088140.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2025 17:35
Operator : AR\AJ
Sample : PCHLORICV500
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
ICVPD041825CHLOR

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 05:48:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 05:47:09 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
 Data File : PD088141.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2025 17:49
 Operator : AR\AJ
 Sample : PTOXICV500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD041825TOX

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 06:18:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:16:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds						
1) SA Tetrachlo...	3.551	2.883	102.5E6	730.7E6	50.719	50.617
7) SA Decachlor...	9.075	8.076	162.9E6	900.9E6	49.398	50.583
Target Compounds						
2) Toxaphene-1	6.242	5.479	11935495	66985324	505.786	512.582
3) Toxaphene-2	6.442	5.650	17155801	45180209	503.625	504.731
4) Toxaphene-3	7.149	6.760	32633182	210.1E6	499.541	503.358
5) Toxaphene-4	7.564	7.201	41575512	152.0E6	506.738	501.932
6) Toxaphene-5	7.930	7.332	23954033	105.1E6	502.021	505.684

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
Data File : PD088141.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2025 17:49
Operator : AR\AJ
Sample : PTOXICV500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 06:18:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:16:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Continuing Calib Date: 05/13/2025 Initial Calibration Date(s): 04/18/2025 04/18/2025

Continuing Calib Time: 09:55 Initial Calibration Time(s): 13:56 14:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.07	9.08	8.98	9.18	0.01
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.93	4.93	4.83	5.03	0.00
Heptachlor epoxide	5.69	5.69	5.59	5.79	0.00
Endrin	6.58	6.58	6.48	6.68	0.00
Methoxychlor	7.49	7.50	7.40	7.60	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Continuing Calib Date: 05/13/2025 Initial Calibration Date(s): 04/18/2025 04/18/2025

Continuing Calib Time: 09:55 Initial Calibration Time(s): 13:56 14:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.08	8.08	7.98	8.18	0.01
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
gamma-BHC (Lindane)	3.73	3.73	3.63	3.83	0.00
Heptachlor	4.09	4.09	3.99	4.19	0.00
Heptachlor epoxide	4.88	4.88	4.78	4.98	0.00
Endrin	5.79	5.79	5.69	5.89	0.00
Methoxychlor	6.76	6.76	6.66	6.86	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL01 Date Analyzed: 05/13/2025

Lab Sample No.: PSTDCCC050 Data File : PD088499.D Time Analyzed: 09:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.074	8.975	9.175	48.950	50.000	-2.1
Endrin	6.576	6.476	6.676	53.150	50.000	6.3
gamma-BHC (Lindane)	4.332	4.232	4.432	53.830	50.000	7.7
Heptachlor	4.931	4.831	5.031	55.670	50.000	11.3
Heptachlor epoxide	5.692	5.592	5.792	54.860	50.000	9.7
Methoxychlor	7.494	7.395	7.595	51.790	50.000	3.6
Tetrachloro-m-xylene	3.552	3.452	3.652	51.960	50.000	3.9



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

 Contract: CAMP02

 Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

 GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

 Client Sample No.: CCAL01 Date Analyzed: 05/13/2025

 Lab Sample No.: PSTDCCC050 Data File : PD088499.D Time Analyzed: 09:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.075	7.977	8.177	42.160	50.000	-15.7
Endrin	5.791	5.693	5.893	46.530	50.000	-6.9
gamma-BHC (Lindane)	3.731	3.633	3.833	47.220	50.000	-5.6
Heptachlor	4.085	3.986	4.186	47.280	50.000	-5.4
Heptachlor epoxide	4.875	4.777	4.977	47.370	50.000	-5.3
Methoxychlor	6.757	6.659	6.859	45.070	50.000	-9.9
Tetrachloro-m-xylene	2.882	2.783	2.983	47.090	50.000	-5.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
 Data File : PD088499.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2025 09:55
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 05/14/2025

Supervised By :mohammad ahmed 05/15/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 12:07:14 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.552	2.882	103.8E6	688.6E6	51.956	47.092
28) SA Decachlor...	9.074	8.075	161.9E6	779.1E6	48.950	42.157
Target Compounds						
2) A alpha-BHC	4.001	3.395	235.7E6	1094.6E6	54.670	47.878
3) MA gamma-BHC...	4.332	3.731	225.4E6	1017.0E6	53.833	47.217
4) MA Heptachlor	4.931	4.085	224.8E6	1011.7E6	55.666	47.278
5) MB Aldrin	5.273	4.369	216.3E6	990.6E6	54.767	47.619m
6) B beta-BHC	4.516	4.027	86646772	436.7E6	53.374	47.195
7) B delta-BHC	4.765	4.263	222.9E6	1017.8E6	54.032	47.868
8) B Heptachlo...	5.692	4.875	196.1E6	895.0E6	54.857	47.367
9) A Endosulfan I	6.075	5.248	184.3E6	751.4E6	54.565	41.715
10) B gamma-Chl...	5.948	5.128	200.7E6	958.7E6	55.406	47.238
11) B alpha-Chl...	6.028	5.192	198.4E6	921.6E6	54.970	47.013
12) B 4,4'-DDE	6.197	5.376	172.6E6	929.5E6	52.343	47.198m
13) MA Dieldrin	6.349	5.515	195.1E6	940.0E6	54.683	47.158
14) MA Endrin	6.576	5.791	158.5E6	847.0E6	53.147	46.533
15) B Endosulfa...	6.787	6.083	164.9E6	819.2E6	52.760	46.752
16) A 4,4'-DDD	6.706	5.932	138.9E6	771.8E6	55.229	47.099
17) MA 4,4'-DDT	7.023	6.186	147.3E6	798.3E6	53.044	46.910
18) B Endrin al...	6.917	6.261	117.5E6	593.2E6	50.893	44.601
19) B Endosulfa...	7.151	6.485	151.8E6	782.6E6	52.791	45.679
20) A Methoxychlor	7.494	6.757	77657457	411.1E6	51.787	45.073
21) B Endrin ke...	7.632	6.994	162.1E6	842.5E6	52.519	45.044
22) Mirex	8.115	7.188	120.6E6	657.5E6	51.345	44.364

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
Data File : PD088499.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2025 09:55
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 05/14/2025

Supervised By :mohammad ahmed 05/15/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 12:07:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Continuing Calib Date: 05/13/2025 Initial Calibration Date(s): 04/18/2025 04/18/2025

Continuing Calib Time: 14:22 Initial Calibration Time(s): 13:56 14:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.08	9.08	8.98	9.18	0.00
Tetrachloro-m-xylene	3.56	3.55	3.45	3.65	-0.01
gamma-BHC (Lindane)	4.34	4.33	4.23	4.43	-0.01
Heptachlor	4.94	4.93	4.83	5.03	-0.01
Heptachlor epoxide	5.70	5.69	5.59	5.79	-0.01
Endrin	6.58	6.58	6.48	6.68	0.00
Methoxychlor	7.50	7.50	7.40	7.60	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

Continuing Calib Date: 05/13/2025 Initial Calibration Date(s): 04/18/2025 04/18/2025

Continuing Calib Time: 14:22 Initial Calibration Time(s): 13:56 14:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.08	8.08	7.98	8.18	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
gamma-BHC (Lindane)	3.73	3.73	3.63	3.83	0.00
Heptachlor	4.09	4.09	3.99	4.19	0.00
Heptachlor epoxide	4.88	4.88	4.78	4.98	0.00
Endrin	5.79	5.79	5.69	5.89	0.00
Methoxychlor	6.76	6.76	6.66	6.86	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL02 Date Analyzed: 05/13/2025

Lab Sample No.: PSTDCCC050 Data File : PD088515.D Time Analyzed: 14:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.080	8.975	9.175	48.470	50.000	-3.1
Endrin	6.582	6.476	6.676	53.290	50.000	6.6
gamma-BHC (Lindane)	4.338	4.232	4.432	55.030	50.000	10.1
Heptachlor	4.937	4.831	5.031	55.970	50.000	11.9
Heptachlor epoxide	5.698	5.592	5.792	54.730	50.000	9.5
Methoxychlor	7.500	7.395	7.595	51.030	50.000	2.1
Tetrachloro-m-xylene	3.558	3.452	3.652	52.540	50.000	5.1



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL02 Date Analyzed: 05/13/2025

Lab Sample No.: PSTDCCC050 Data File : PD088515.D Time Analyzed: 14:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.076	7.977	8.177	41.330	50.000	-17.3
Endrin	5.792	5.693	5.893	47.220	50.000	-5.6
gamma-BHC (Lindane)	3.732	3.633	3.833	47.870	50.000	-4.3
Heptachlor	4.086	3.986	4.186	47.490	50.000	-5.0
Heptachlor epoxide	4.876	4.777	4.977	47.790	50.000	-4.4
Methoxychlor	6.758	6.659	6.859	45.030	50.000	-9.9
Tetrachloro-m-xylene	2.882	2.783	2.983	47.460	50.000	-5.1

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
 Data File : PD088515.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2025 14:22
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDCCC050

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 05/14/2025

Supervised By :mohammad ahmed 05/15/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 22:41:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.558	2.882	104.9E6	693.9E6	52.538	47.459
28)	SA Decachlor...	9.080	8.076	160.3E6	763.7E6	48.470	41.326
Target Compounds							
2)	A alpha-BHC	4.007	3.395	240.0E6	1107.8E6	55.655	48.459
3)	MA gamma-BHC...	4.338	3.732	230.5E6	1031.1E6	55.035	47.870
4)	MA Heptachlor	4.937	4.086	226.1E6	1016.2E6	55.968	47.489
5)	MB Aldrin	5.279	4.370	219.6E6	1003.7E6	55.594	48.248m
6)	B beta-BHC	4.523	4.028	88197919	446.0E6	54.330	48.192
7)	B delta-BHC	4.771	4.265	233.6E6	1034.6E6	56.622	48.661
8)	B Heptachlo...	5.698	4.876	195.6E6	903.0E6	54.730	47.793
9)	A Endosulfan I	6.082	5.249	187.2E6	750.2E6	55.444	41.646
10)	B gamma-Chl...	5.952	5.129	196.6E6	967.4E6	54.275m	47.663
11)	B alpha-Chl...	6.035	5.194	203.7E6	929.1E6	56.447	47.399
12)	B 4,4'-DDE	6.203	5.377	172.9E6	935.0E6	52.408	47.477m
13)	MA Dieldrin	6.354	5.516	195.7E6	948.4E6	54.837	47.580
14)	MA Endrin	6.582	5.792	158.9E6	859.6E6	53.292	47.223
15)	B Endosulfa...	6.793	6.084	165.8E6	823.2E6	53.036	46.981
16)	A 4,4'-DDD	6.712	5.933	143.9E6	776.3E6	57.225	47.376
17)	MA 4,4'-DDT	7.028	6.188	146.0E6	793.8E6	52.550	46.649
18)	B Endrin al...	6.922	6.262	118.6E6	597.6E6	51.376	44.930
19)	B Endosulfa...	7.156	6.486	152.4E6	783.1E6	53.003	45.708
20)	A Methoxychlor	7.500	6.758	76520112	410.7E6	51.029	45.032
21)	B Endrin ke...	7.637	6.995	161.9E6	841.9E6	52.446	45.009
22)	Mirex	8.120	7.189	119.6E6	647.8E6	50.918	43.707

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
Data File : PD088515.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2025 14:22
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 05/14/2025

Supervised By :mohammad ahmed 05/15/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 22:41:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No. (PEM): PEM - PD088122.D Date Analyzed: 04/18/2025

Lab Sample No.(PEM): PEM Time Analyzed: 13:29

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	9.075	8.970	9.180	23.220	20.000	16.1
Tetrachloro-m-xylene	3.551	3.500	3.600	21.610	20.000	8.1
alpha-BHC	4.000	3.950	4.050	9.950	10.000	-0.5
beta-BHC	4.516	4.470	4.570	11.360	10.000	13.6
gamma-BHC (Lindane)	4.331	4.280	4.380	10.480	10.000	4.8
Endrin	6.576	6.510	6.650	51.530	50.000	3.1
4,4'-DDT	7.023	6.950	7.090	110.510	100.000	10.5
Methoxychlor	7.494	7.420	7.560	265.100	250.000	6.0

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No. (PEM): PEM - PD088122.D Date Analyzed: 04/18/2025

Lab Sample No.(PEM): PEM Time Analyzed: 13:29

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	8.076	7.980	8.180	22.950	20.000	14.8
Tetrachloro-m-xylene	2.883	2.830	2.930	22.200	20.000	11.0
alpha-BHC	3.396	3.350	3.450	11.720	10.000	17.2
beta-BHC	4.028	3.980	4.080	12.460	10.000	24.6
gamma-BHC (Lindane)	3.732	3.680	3.780	11.580	10.000	15.8
Endrin	5.793	5.720	5.860	49.780	50.000	-0.4
4,4'-DDT	6.187	6.120	6.260	102.610	100.000	2.6
Methoxychlor	6.758	6.690	6.830	215.580	250.000	-13.8

PEM
 Data File: PD088122.D Date Acquired 4/18/2025 13:29
 Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin	6.58	153680370.7	162192230.2	8511859.47	Down 5.25
Endrin aldehyde	6.92	3420195.012			
Endrin ketone	7.63	5091664.461			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.79	906106216.2	976124672.3	70018456.1	7.17
Endrin aldehyde #2	6.26	27341767.91			
Endrin ketone #2	6.99	42676688.15			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	7.02	306944307.8	307584368.9	640061.113	0.21
4,4'-DDE	0.00	0			
4,4'-DDD	6.71	640061.113			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.19	1746026847	1755776541	9749694.14	0.56
4,4'-DDE #2	0.00	0			
4,4'-DDD #2	5.94	9749694.138			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
 Data File : PD088122.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2025 13:29
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 04/19/2025
 Supervised By :mohammad ahmed 04/22/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 06:41:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.551	2.883	43154195	324.6E6	21.605	22.198
2)	SA Decachlor...	9.075	8.076	76807188	424.1E6	23.218	22.951
Target Compounds							
2)	A alpha-BHC	4.000	3.396	42909357	267.8E6	9.951	11.716
3)	MA gamma-BHC...	4.331	3.732	43874391	249.5E6	10.477	11.583
6)	B beta-BHC	4.516	4.028	18444986	115.3E6	11.362	12.462
14)	MA Endrin	6.576	5.793	153.7E6	906.1E6	51.526	49.780
16)	A 4,4'-DDD	6.708	5.936	640061	9749694	0.255	0.595m#
17)	MA 4,4'-DDT	7.023	6.187	306.9E6	1746.0E6	110.509	102.608
18)	B Endrin al...	6.915	6.262	3420195	27341768	1.482	2.056 #
20)	A Methoxychlor	7.494	6.758	397.5E6	1966.3E6	265.100	215.579
21)	B Endrin ke...	7.630	6.994	5091664	42676688	1.650	2.282 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
Data File : PD088122.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2025 13:29
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 04/19/2025
Supervised By : mohammad ahmed 04/22/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 06:41:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No. (PEM): PEM - PD088498.D Date Analyzed: 05/13/2025

Lab Sample No.(PEM): PEM Time Analyzed: 09:41

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	9.074	8.970	9.170	23.730	20.000	18.7
Tetrachloro-m-xylene	3.551	3.500	3.600	23.660	20.000	18.3
alpha-BHC	4.000	3.950	4.050	10.800	10.000	8.0
beta-BHC	4.516	4.470	4.570	12.260	10.000	22.6
gamma-BHC (Lindane)	4.331	4.280	4.380	11.190	10.000	11.9
Endrin	6.575	6.500	6.650	59.430	50.000	18.9
4,4'-DDT	7.021	6.950	7.090	119.580	100.000	19.6
Methoxychlor	7.494	7.420	7.560	270.970	250.000	8.4

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No. (PEM): PEM - PD088498.D Date Analyzed: 05/13/2025

Lab Sample No.(PEM): PEM Time Analyzed: 09:41

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	8.073	7.970	8.170	20.850	20.000	4.3
Tetrachloro-m-xylene	2.881	2.830	2.930	22.400	20.000	12.0
alpha-BHC	3.394	3.340	3.440	11.630	10.000	16.3
beta-BHC	4.026	3.980	4.080	12.030	10.000	20.3
gamma-BHC (Lindane)	3.730	3.680	3.780	11.540	10.000	15.4
Endrin	5.790	5.720	5.860	51.700	50.000	3.4
4,4'-DDT	6.185	6.110	6.260	99.870	100.000	-0.1
Methoxychlor	6.755	6.680	6.830	200.220	250.000	-19.9

PEM
 Data File: PD088498.D Date Acquired 5/13/2025 9:41
 Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin	6.57	177244119.1	181454592.7	4210473.63	2.32
Endrin aldehyde	6.92	1009509.166			
Endrin ketone	7.63	3200964.467			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.79	941050922.5	977842641.7	36791719.2	3.76
Endrin aldehyde #2	6.26	12256261.66			
Endrin ketone #2	6.99	24535457.53			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	7.02	332149556.5	336484886.9	4335330.38	1.29
4,4'-DDE	0.00	0			
4,4'-DDD	6.70	4335330.375			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.18	1699406352	1725577036	26170683.3	1.52
4,4'-DDE #2	0.00	0			
4,4'-DDD #2	5.93	26170683.25			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
 Data File : PD088498.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2025 09:41
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 05/14/2025
 Supervised By :mohammad ahmed 05/15/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 12:06:49 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.551	2.881	47263112	327.5E6	23.662	22.398
28)	SA Decachlor...	9.074	8.073	78502399	385.3E6	23.730	20.847
Target Compounds							
2)	A alpha-BHC	4.000	3.394	46551953	265.9E6	10.796	11.632
3)	MA gamma-BHC...	4.331	3.730	46855378	248.5E6	11.189	11.539
6)	B beta-BHC	4.516	4.026	19897245	111.3E6	12.257	12.027
14)	MA Endrin	6.575	5.790	177.2E6	941.1E6	59.427	51.699
16)	A 4,4'-DDD	6.705	5.931	4335330	26170683	1.724	1.597
17)	MA 4,4'-DDT	7.021	6.185	332.1E6	1699.4E6	119.584	99.868
18)	B Endrin al...	6.917	6.259	1009509	12256262	0.437	0.921 #
20)	A Methoxychlor	7.494	6.755	406.3E6	1826.2E6	270.965	200.217 #
21)	B Endrin ke...	7.629	6.992	3200964	24535458	1.037m	1.312 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
Data File : PD088498.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2025 09:41
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 05/14/2025
Supervised By : mohammad ahmed 05/15/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 12:06:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
Data File : PD088123.D
Acq On : 18 Apr 2025 13:43
Operator : AR\AJ
Sample : RESCHK
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e

Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Title : GC Extractables
Last Update : Sat Apr 19 06:32:27 2025
Integrator: ChemStation

RT#1	RT#2	Resolution
3.552	5.948	100.00%
5.948	6.077	100.00%
6.077	6.198	100.00%
6.198	6.349	100.00%
6.349	7.151	100.00%
7.151	7.495	100.00%
7.495	7.631	100.00%
7.631	9.075	100.00%

Signal #2

2.883	5.130	100.00%
5.130	5.251	100.00%
5.251	5.379	100.00%
5.379	5.517	100.00%
5.517	6.486	100.00%
6.486	6.759	100.00%
6.759	6.995	100.00%
6.995	8.076	100.00%

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
 Data File : PD088123.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2025 13:43
 Operator : AR\AJ
 Sample : RESCHK
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 RESCHK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 06:03:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.552	2.883	37870717	290.8E6	18.960	19.885
28)	SA Decachlor...	9.075	8.076	66771462	372.0E6	20.184	20.130
Target Compounds							
9)	A Endosulfan I	6.077	5.251	29410187	177.2E6	8.709	9.836
10)	B gamma-Chl...	5.948	5.130	32853106	209.1E6	9.070	10.304
12)	B 4,4'-DDE	6.198	5.379	60393989	395.9E6	18.311	20.100
13)	MA Dieldrin	6.349	5.517	64800077	392.3E6	18.159	19.682
19)	B Endosulfa...	7.151	6.486	54061534	339.7E6	18.799	19.830
20)	A Methoxychlor	7.495	6.758	137.4E6	780.0E6	91.608	85.519
21)	B Endrin ke...	7.631	6.995	57978754	371.5E6	18.784	19.864



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
Data File : PD088123.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2025 13:43
Operator : AR\AJ
Sample : RESCHK
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
RESCHK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 06:03:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Analytical Sequence

Client: CDM Smith	SDG No.: Q1956
Project: Bergen Point Fueling System	Instrument ID: ECD_D
GC Column: ZB-MR1	ID: 0.32 (mm) Inst. Calib. Date(s): 04/18/2025 04/18/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
PEM	PEM	04/18/2025	13:29	PD088122.D	9.08	3.55
RESCHK	RESCHK	04/18/2025	13:43	PD088123.D	9.08	3.55
PSTDICC100	PSTDICC100	04/18/2025	13:56	PD088124.D	9.08	3.55
PSTDICC075	PSTDICC075	04/18/2025	14:10	PD088125.D	9.07	3.55
PSTDICC050	PSTDICC050	04/18/2025	14:24	PD088126.D	9.08	3.55
PSTDICC025	PSTDICC025	04/18/2025	14:37	PD088127.D	9.08	3.55
PSTDICC005	PSTDICC005	04/18/2025	14:51	PD088128.D	9.08	3.55
PCHLORICC500	PCHLORICC500	04/18/2025	15:32	PD088131.D	9.07	3.55
PTOXICC500	PTOXICC500	04/18/2025	16:40	PD088136.D	9.07	3.55
LBLK	LBLK	04/18/2025	18:02	PD088142.D	9.08	3.55
LBLK	LBLK	05/13/2025	09:28	PD088497.D	9.07	3.55
PEM	PEM	05/13/2025	09:41	PD088498.D	9.07	3.55
PSTDCCC050	PSTDCCC050	05/13/2025	09:55	PD088499.D	9.07	3.55
PB167969BL	PB167969BL	05/13/2025	11:57	PD088508.D	9.07	3.55
PB167969BS	PB167969BS	05/13/2025	12:10	PD088509.D	9.07	3.55
PB167851TB	PB167851TB	05/13/2025	12:24	PD088510.D	9.07	3.55
COMPI	Q1956-05	05/13/2025	12:37	PD088511.D	9.08	3.55
COMP1MS	Q1956-05MS	05/13/2025	12:51	PD088512.D	9.07	3.55
COMP1MSD	Q1956-05MSD	05/13/2025	13:04	PD088513.D	9.07	3.55
LBLK	LBLK	05/13/2025	13:18	PD088514.D	9.07	3.55
PSTDCCC050	PSTDCCC050	05/13/2025	14:22	PD088515.D	9.08	3.56

Analytical Sequence

Client: CDM Smith

SDG No.: Q1956

Project: Bergen Point Fueling System

Instrument ID: ECD_D

GC Column: ZB-MR2

ID: 0.32 (mm)

Inst. Calib. Date(s): 04/18/2025 04/18/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
PEM	PEM	04/18/2025	13:29	PD088122.D	8.08	2.88
RESCHK	RESCHK	04/18/2025	13:43	PD088123.D	8.08	2.88
PSTDICC100	PSTDICC100	04/18/2025	13:56	PD088124.D	8.09	2.90
PSTDICC075	PSTDICC075	04/18/2025	14:10	PD088125.D	8.08	2.88
PSTDICC050	PSTDICC050	04/18/2025	14:24	PD088126.D	8.08	2.88
PSTDICC025	PSTDICC025	04/18/2025	14:37	PD088127.D	8.08	2.88
PSTDICC005	PSTDICC005	04/18/2025	14:51	PD088128.D	8.08	2.88
PCHLORICC500	PCHLORICC500	04/18/2025	15:32	PD088131.D	8.08	2.88
PTOXICC500	PTOXICC500	04/18/2025	16:40	PD088136.D	8.08	2.88
LBLK	LBLK	04/18/2025	18:02	PD088142.D	8.08	2.88
LBLK	LBLK	05/13/2025	09:28	PD088497.D	8.07	2.88
PEM	PEM	05/13/2025	09:41	PD088498.D	8.07	2.88
PSTDCCC050	PSTDCCC050	05/13/2025	09:55	PD088499.D	8.08	2.88
PB167969BL	PB167969BL	05/13/2025	11:57	PD088508.D	8.07	2.88
PB167969BS	PB167969BS	05/13/2025	12:10	PD088509.D	8.08	2.88
PB167851TB	PB167851TB	05/13/2025	12:24	PD088510.D	8.07	2.88
COMPI	Q1956-05	05/13/2025	12:37	PD088511.D	8.08	2.88
COMP1MS	Q1956-05MS	05/13/2025	12:51	PD088512.D	8.08	2.88
COMP1MSD	Q1956-05MSD	05/13/2025	13:04	PD088513.D	8.08	2.88
LBLK	LBLK	05/13/2025	13:18	PD088514.D	8.08	2.88
PSTDCCC050	PSTDCCC050	05/13/2025	14:22	PD088515.D	8.08	2.88

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

COMP1MS

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956

SAS No.: Q1956 SDG NO.: Q1956

Lab Sample ID: Q1956-05MS

Date(s) Analyzed: 05/13/2025 05/13/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Methoxychlor	1	7.49	7.44	7.54	5.60	15.4
	2	6.76	6.71	6.81	4.80	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	6.10	15.9
	2	3.73	3.68	3.78	5.20	
Heptachlor	1	4.93	4.88	4.98	6.00	18.2
	2	4.09	4.04	4.14	5.00	
Heptachlor epoxide	1	5.69	5.64	5.74	6.10	15.9
	2	4.88	4.83	4.93	5.20	
Endrin	1	6.58	6.53	6.63	6.00	16.2
	2	5.79	5.74	5.84	5.10	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

COMP1MSD

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956

SAS No.: Q1956 SDG NO.: Q1956

Lab Sample ID: Q1956-05MSD

Date(s) Analyzed: 05/13/2025 05/13/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Methoxychlor	1	7.49	7.44	7.54	5.60	15.4
	2	6.76	6.71	6.81	4.80	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	6.00	16.2
	2	3.73	3.68	3.78	5.10	
Heptachlor	1	4.93	4.88	4.98	6.00	18.2
	2	4.09	4.04	4.14	5.00	
Heptachlor epoxide	1	5.69	5.64	5.74	6.10	15.9
	2	4.88	4.83	4.93	5.20	
Endrin	1	6.58	6.53	6.63	6.00	16.2
	2	5.79	5.74	5.84	5.10	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB167969BS

Contract: CAMP02

Lab Code: CHEM Case No.: Q1956

SAS No.: Q1956 SDG NO.: Q1956

Lab Sample ID: PB167969BS

Date(s) Analyzed: 05/13/2025 05/13/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endrin	1	6.58	6.53	6.63	0.48	11.1
	2	5.79	5.74	5.84	0.43	
Methoxychlor	1	7.49	7.44	7.54	0.47	13.5
	2	6.76	6.71	6.81	0.41	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	0.51	8.9
	2	3.73	3.68	3.78	0.47	
Heptachlor	1	4.93	4.88	4.98	0.53	12.2
	2	4.09	4.04	4.14	0.47	
Heptachlor epoxide	1	5.69	5.64	5.74	0.52	10.5
	2	4.87	4.82	4.92	0.47	



QC SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Bergen Point Fueling System	Date Received:	
Client Sample ID:	PB167969BL	SDG No.:	Q1956
Lab Sample ID:	PB167969BL	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088508.D	1	05/12/25 12:25	05/13/25 11:57	PB167969

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.050	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.050	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.050	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.050	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.050	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	1.00	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.50	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	21.2		43 - 140	106%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.0		77 - 126	105%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
 Data File : PD088508.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2025 11:57
 Operator : AR\AJ
 Sample : PB167969BL
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB167969BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 22:39:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.551	2.883	41919598	302.8E6	20.987	20.707
28) SA Decachlor...	9.074	8.074	70135821	305.8E6	21.201	16.548

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
Data File : PD088508.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2025 11:57
Operator : AR\AJ
Sample : PB167969BL
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB167969BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 22:39:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.551 min
Delta R.T.: 0.000 min
Response: 41919598
Conc: 20.99 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB167969BL




#1 Tetrachloro-m-xylene

R.T.: 2.883 min
Delta R.T.: 0.000 min
Response: 302773914
Conc: 20.71 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.074 min
Delta R.T.: -0.001 min
Response: 70135821
Conc: 21.20 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.074 min
Delta R.T.: -0.003 min
Response: 305800377
Conc: 16.55 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	04/18/25
Project:	Bergen Point Fueling System	Date Received:	04/18/25
Client Sample ID:	PIBLK-PD088142.D	SDG No.:	Q1956
Lab Sample ID:	I.BLK-PD088142.D	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088142.D	1		04/18/25	PD041825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.050	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.050	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.050	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.050	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.050	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	1.00	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.50	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	23.3		43 - 140	116%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.4		77 - 126	107%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
 Data File : PD088142.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2025 18:02
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 06:42:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.552	2.883	41440080	312.2E6	20.747	21.349
28) SA Decachlor...	9.075	8.077	76796349	430.2E6	23.215	23.281

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD041825\
Data File : PD088142.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2025 18:02
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 06:42:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.552 min
Delta R.T.: 0.000 min
Response: 41440080
Conc: 20.75 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.883 min
Delta R.T.: 0.000 min
Response: 312150762
Conc: 21.35 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.075 min
Delta R.T.: 0.000 min
Response: 76796349
Conc: 23.21 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.077 min
Delta R.T.: 0.000 min
Response: 430226938
Conc: 23.28 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/13/25
Project:	Bergen Point Fueling System	Date Received:	05/13/25
Client Sample ID:	PIBLK-PD088497.D	SDG No.:	Q1956
Lab Sample ID:	I.BLK-PD088497.D	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088497.D	1		05/13/25	PD051325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.050	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.050	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.050	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.050	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.050	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	1.00	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.50	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	19.7		43 - 140	99%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.5		77 - 126	92%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
 Data File : PD088497.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2025 09:28
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 12:06:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.551	2.881	36901453	257.0E6	18.475	17.575
28) SA Decachlor...	9.074	8.074	65198097	321.4E6	19.709	17.391

Target Compounds


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
Data File : PD088497.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2025 09:28
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 12:06:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.551 min
Delta R.T.: 0.000 min
Response: 36901453
Conc: 18.47 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.881 min
Delta R.T.: -0.002 min
Response: 256973831
Conc: 17.57 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.074 min
Delta R.T.: 0.000 min
Response: 65198097
Conc: 19.71 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.074 min
Delta R.T.: -0.003 min
Response: 321387887
Conc: 17.39 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/13/25
Project:	Bergen Point Fueling System	Date Received:	05/13/25
Client Sample ID:	PIBLK-PD088514.D	SDG No.:	Q1956
Lab Sample ID:	I.BLK-PD088514.D	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088514.D	1		05/13/25	pd051325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.050	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.050	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.050	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.050	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.050	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	1.00	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.50	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	18.7		43 - 140	94%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.1		77 - 126	90%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
Data File : PD088514.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2025 13:18
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 22:41:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.552	2.882	36117016	247.6E6	18.082	16.935
28) SA Decachlor...	9.074	8.075	61900960	290.6E6	18.712	15.727

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
Data File : PD088514.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2025 13:18
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 22:41:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.552 min
Delta R.T.: 0.000 min
Response: 36117016
Conc: 18.08 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.882 min
Delta R.T.: 0.000 min
Response: 247611102
Conc: 16.93 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.074 min
Delta R.T.: 0.000 min
Response: 61900960
Conc: 18.71 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.075 min
Delta R.T.: -0.002 min
Response: 290627633
Conc: 15.73 ng/ml

Report of Analysis

Client:	CDM Smith		Date Collected:		
Project:	Bergen Point Fueling System		Date Received:		
Client Sample ID:	PB167969BS		SDG No.:	Q1956	
Lab Sample ID:	PB167969BS		Matrix:	TCLP	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088509.D	1	05/12/25 12:25	05/13/25 12:10	PB167969

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.51		0.0037	0.050	ug/L
76-44-8	Heptachlor	0.53		0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.52		0.0096	0.050	ug/L
72-20-8	Endrin	0.48		0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.47		0.011	0.050	ug/L
8001-35-2	Toxaphene	1.00	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.50	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	21.3		43 - 140	107%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.2		77 - 126	106%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
 Data File : PD088509.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2025 12:10
 Operator : AR\AJ
 Sample : PB167969BS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB167969BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 22:39:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.552	2.883	42287586	307.8E6	21.171	21.051
28)	SA Decachlor...	9.074	8.075	70584158	326.5E6	21.337	17.669
Target Compounds							
2)	A alpha-BHC	4.001	3.395	224.6E6	1088.7E6	52.094	47.623
3)	MA gamma-BHC...	4.332	3.731	214.8E6	1010.8E6	51.303	46.927
4)	MA Heptachlor	4.931	4.085	213.2E6	1000.0E6	52.780	46.732
5)	MB Aldrin	5.273	4.370	207.5E6	987.2E6	52.532	47.454
6)	B beta-BHC	4.516	4.026	82305767	440.2E6	50.700	47.569
7)	B delta-BHC	4.765	4.263	215.8E6	1012.7E6	52.323	47.631
8)	B Heptachlo...	5.692	4.874	185.1E6	880.5E6	51.785	46.600
9)	A Endosulfan I	6.076	5.249	174.7E6	835.2E6	51.725	46.365
10)	B gamma-Chl...	5.947	5.127	185.6E6	938.5E6	51.237	46.240
11)	B alpha-Chl...	6.028	5.192	185.7E6	903.2E6	51.458	46.077
12)	B 4,4'-DDE	6.197	5.377	166.3E6	907.9E6	50.429	46.103
13)	MA Dieldrin	6.348	5.514	185.3E6	916.0E6	51.914	45.954
14)	MA Endrin	6.575	5.790	144.2E6	787.5E6	48.352	43.265
15)	B Endosulfa...	6.786	6.082	155.1E6	779.6E6	49.602	44.495
16)	A 4,4'-DDD	6.705	5.932	130.2E6	756.7E6	51.787	46.177
17)	MA 4,4'-DDT	7.022	6.186	137.1E6	731.5E6	49.366	42.990
18)	B Endrin al...	6.915	6.261	116.0E6	585.2E6	50.280	43.996
19)	B Endosulfa...	7.149	6.484	145.0E6	747.9E6	50.422	43.654
20)	A Methoxychlor	7.493	6.756	70243663	373.3E6	46.843	40.929
21)	B Endrin ke...	7.631	6.993	156.8E6	807.0E6	50.798	43.145
22)	Mirex	8.115	7.188	111.3E6	588.5E6	47.375	39.705



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
Data File : PD088509.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2025 12:10
Operator : AR\AJ
Sample : PB167969BS
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB167969BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 22:39:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	CDM Smith	Date Collected:	05/02/25
Project:	Bergen Point Fueling System	Date Received:	05/02/25
Client Sample ID:	COMP1MS	SDG No.:	Q1956
Lab Sample ID:	Q1956-05MS	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	100 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088512.D	1	05/12/25 12:25	05/13/25 12:51	PB167969

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	6.10		0.037	0.50	ug/L
76-44-8	Heptachlor	6.00		0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	6.10		0.096	0.50	ug/L
72-20-8	Endrin	6.00		0.032	0.50	ug/L
72-43-5	Methoxychlor	5.60		0.11	0.50	ug/L
8001-35-2	Toxaphene	10.0	U	1.70	10.0	ug/L
57-74-9	Chlordane	5.00	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	21.6		43 - 140	108%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.8		77 - 126	109%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
 Data File : PD088512.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2025 12:51
 Operator : AR\AJ
 Sample : Q1956-05MS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

COMP1MS

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 05/14/2025

Supervised By :mohammad ahmed 05/15/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 22:40:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.552	2.883	43512328	281.2E6	21.785	19.234
28) SA Decachlor...	9.073	8.075	71443577	330.1E6	21.597	17.864
Target Compounds						
2) A alpha-BHC	4.001	3.395	264.3E6	1198.5E6	61.298	52.423
3) MA gamma-BHC...	4.332	3.732	254.6E6	1115.6E6	60.794	51.797
4) MA Heptachlor	4.931	4.085	242.9E6	1075.9E6	60.150	50.282
5) MB Aldrin	5.273	4.371	234.9E6	1062.4E6	59.474	51.068
6) B beta-BHC	4.517	4.027	96618112	489.9E6	59.517	52.935
7) B delta-BHC	4.765	4.264	270.3E6	1132.9E6	65.533	53.282
8) B Heptachlo...	5.692	4.875	217.5E6	987.3E6	60.845	52.254
9) A Endosulfan I	6.076	5.249	204.9E6	933.4E6	60.668	51.821
10) B gamma-Chl...	5.947	5.128	219.2E6	1050.4E6	60.514	51.755
11) B alpha-Chl...	6.028	5.193	217.9E6	1016.8E6	60.363	51.872
12) B 4,4'-DDE	6.197	5.378	196.2E6	1013.9E6	59.478	51.484
13) MA Dieldrin	6.348	5.515	218.9E6	1028.0E6	61.355	51.571
14) MA Endrin	6.575	5.791	179.4E6	931.6E6	60.137	51.183
15) B Endosulfa...	6.786	6.083	182.3E6	906.9E6	58.320	51.757
16) A 4,4'-DDD	6.706	5.932	151.9E6	850.3E6	60.392	51.893
17) MA 4,4'-DDT	7.021	6.186	159.7E6	852.6E6	57.498m	50.102
18) B Endrin al...	6.916	6.260	138.6E6	684.4E6	60.033	51.458
19) B Endosulfa...	7.150	6.484	171.4E6	856.9E6	59.614	50.016
20) A Methoxychlor	7.493	6.756	84337634	435.9E6	56.242	47.795
21) B Endrin ke...	7.631	6.993	182.6E6	915.8E6	59.163	48.962
22) Mirex	8.115	7.188	129.8E6	678.2E6	55.266	45.762

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
Data File : PD088512.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2025 12:51
Operator : AR\AJ
Sample : Q1956-05MS
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

COMP1MS

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 05/14/2025

Supervised By :mohammad ahmed 05/15/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 22:40:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	CDM Smith	Date Collected:	05/02/25
Project:	Bergen Point Fueling System	Date Received:	05/02/25
Client Sample ID:	COMP1MSD	SDG No.:	Q1956
Lab Sample ID:	Q1956-05MSD	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	100 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088513.D	1	05/12/25 12:25	05/13/25 13:04	PB167969

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	6.00		0.037	0.50	ug/L
76-44-8	Heptachlor	6.00		0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	6.10		0.096	0.50	ug/L
72-20-8	Endrin	6.00		0.032	0.50	ug/L
72-43-5	Methoxychlor	5.60		0.11	0.50	ug/L
8001-35-2	Toxaphene	10.0	U	1.70	10.0	ug/L
57-74-9	Chlordane	5.00	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	21.6		43 - 140	108%	SPK: 20
877-09-8	Tetrachloro-m-xylene	22.3		77 - 126	111%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
 Data File : PD088513.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2025 13:04
 Operator : AR\AJ
 Sample : Q1956-05MSD
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 COMP1MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 22:40:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.552	2.883	44518692	277.1E6	22.288	18.952
28)	SA Decachlor...	9.074	8.075	71327239	330.8E6	21.561	17.901
Target Compounds							
2)	A alpha-BHC	4.001	3.395	262.2E6	1190.2E6	60.797	52.062
3)	MA gamma-BHC...	4.332	3.731	252.4E6	1109.0E6	60.270	51.491
4)	MA Heptachlor	4.931	4.085	242.6E6	1067.9E6	60.057	49.906
5)	MB Aldrin	5.273	4.371	234.0E6	1058.8E6	59.237	50.899
6)	B beta-BHC	4.516	4.027	96206207	486.0E6	59.263	52.519
7)	B delta-BHC	4.765	4.264	271.7E6	1127.6E6	65.870	53.035
8)	B Heptachlo...	5.692	4.875	216.6E6	984.3E6	60.601	52.093
9)	A Endosulfan I	6.076	5.249	203.9E6	931.7E6	60.387	51.728
10)	B gamma-Chl...	5.947	5.128	217.9E6	1050.1E6	60.146	51.738
11)	B alpha-Chl...	6.028	5.192	216.5E6	1015.7E6	59.978	51.813
12)	B 4,4'-DDE	6.197	5.377	195.3E6	1012.6E6	59.205	51.415
13)	MA Dieldrin	6.348	5.515	219.1E6	1026.8E6	61.411	51.511
14)	MA Endrin	6.575	5.791	179.5E6	932.2E6	60.170	51.214
15)	B Endosulfa...	6.787	6.083	182.9E6	910.4E6	58.517	51.956
16)	A 4,4'-DDD	6.706	5.932	152.0E6	850.9E6	60.447	51.929
17)	MA 4,4'-DDT	7.022	6.186	158.9E6	854.3E6	57.199	50.206
18)	B Endrin al...	6.916	6.261	138.5E6	683.3E6	60.019	51.375
19)	B Endosulfa...	7.150	6.484	171.5E6	861.1E6	59.638	50.261
20)	A Methoxychlor	7.494	6.757	84387655	436.5E6	56.276	47.860
21)	B Endrin ke...	7.631	6.993	182.5E6	916.3E6	59.126	48.988
22)	Mirex	8.115	7.188	130.0E6	678.1E6	55.350	45.750



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051325\
Data File : PD088513.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2025 13:04
Operator : AR\AJ
Sample : Q1956-05MSD
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
COMP1MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 22:40:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integration Report

Sequence:

PD041825

Instrument

ECD_d

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PD088122.D	4,4"-DDD #2	Abdul	4/19/2025 2:06:43 PM	mohammad	4/22/2025 2:20:46	Peak Integrated by Software
PSTDICC100	PD088124.D	Heptachlor epoxide #2	Abdul	4/19/2025 2:06:47 PM	mohammad	4/22/2025 2:20:46	Peak Integrated by Software
PSTDICC005	PD088128.D	delta-BHC	Abdul	4/19/2025 2:20:23 PM	mohammad	4/22/2025 2:20:46	Peak Integrated by Software
PEM	PD088143.D	4,4"-DDD	Abdul	4/19/2025 2:07:12 PM	mohammad	4/22/2025 2:20:46	Peak Integrated by Software
PEM	PD088143.D	Endrin aldehyde	Abdul	4/19/2025 2:07:12 PM	mohammad	4/22/2025 2:20:46	Peak Integrated by Software

Manual Integration Report

Sequence:	PD051325	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PD088498.D	Endrin ketone	Abdul	5/14/2025 8:32:32 AM	 	 	Peak Integrated by Software
PSTDCCC050	PD088499.D	4,4"-DDE #2	Abdul	5/14/2025 8:32:36 AM	 	 	Peak Integrated by Software
PSTDCCC050	PD088499.D	Aldrin #2	Abdul	5/14/2025 8:32:36 AM	 	 	Peak Integrated by Software
Q1956-05MS	PD088512.D	4,4"-DDT	Abdul	5/14/2025 8:32:51 AM	 	 	Peak Integrated by Software
PSTDCCC050	PD088515.D	4,4"-DDE #2	Abdul	5/14/2025 8:32:54 AM	 	 	Peak Integrated by Software
PSTDCCC050	PD088515.D	Aldrin #2	Abdul	5/14/2025 8:32:54 AM	 	 	Peak Integrated by Software
PSTDCCC050	PD088515.D	gamma-Chlordane	Abdul	5/14/2025 8:32:54 AM	 	 	Peak Integrated by Software
PSTDCCC050	PD088524.D	4,4"-DDE #2	Abdul	5/14/2025 8:33:05 AM	 	 	Peak Integrated by Software
PSTDCCC050	PD088524.D	Aldrin #2	Abdul	5/14/2025 8:33:05 AM	 	 	Peak Integrated by Software
PSTDCCC050	PD088524.D	gamma-Chlordane	Abdul	5/14/2025 8:33:05 AM	 	 	Peak Integrated by Software
PEM	PD088536.D	Endrin aldehyde	Abdul	5/14/2025 8:33:27 AM	 	 	Peak Integrated by Software
PSTDCCC050	PD088537.D	4,4"-DDE #2	Abdul	5/14/2025 8:33:30 AM	 	 	Peak Integrated by Software
PSTDCCC050	PD088537.D	Aldrin #2	Abdul	5/14/2025 8:33:30 AM	 	 	Peak Integrated by Software

Manual Integration Report

Sequence:

PD051325

Instrument

ECD_d

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PSTDCCC050	PD088537.D	gamma-Chlordane	Abdul	5/14/2025 8:33:30 AM	 	 	Peak Integrated by Software
PSTDCCC050	PD088541.D	4,4"-DDE #2	Abdul	5/14/2025 8:33:35 AM	 	 	Peak Integrated by Software
PSTDCCC050	PD088541.D	Aldrin #2	Abdul	5/14/2025 8:33:35 AM	 	 	Peak Integrated by Software
PSTDCCC050	PD088541.D	gamma-Chlordane	Abdul	5/14/2025 8:33:35 AM	 	 	Peak Integrated by Software

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD041825

Review By	Abdul	Review On	4/19/2025 2:07:56 PM
Supervise By	mohammad	Supervise On	4/22/2025 2:20:46 AM
SubDirectory	PD041825	HP Acquire Method	HP Processing Method pd041825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD088120.D	18 Apr 2025 13:02	AR\AJ	Ok
2	I.BLK	PD088121.D	18 Apr 2025 13:15	AR\AJ	Ok
3	PEM	PD088122.D	18 Apr 2025 13:29	AR\AJ	Ok,M
4	RESCHK	PD088123.D	18 Apr 2025 13:43	AR\AJ	Ok
5	PSTDICC100	PD088124.D	18 Apr 2025 13:56	AR\AJ	Ok,M
6	PSTDICC075	PD088125.D	18 Apr 2025 14:10	AR\AJ	Ok
7	PSTDICC050	PD088126.D	18 Apr 2025 14:24	AR\AJ	Ok
8	PSTDICC025	PD088127.D	18 Apr 2025 14:37	AR\AJ	Ok
9	PSTDICC005	PD088128.D	18 Apr 2025 14:51	AR\AJ	Ok,M
10	PCHLORICC1000	PD088129.D	18 Apr 2025 15:05	AR\AJ	Ok
11	PCHLORICC750	PD088130.D	18 Apr 2025 15:18	AR\AJ	Ok
12	PCHLORICC500	PD088131.D	18 Apr 2025 15:32	AR\AJ	Ok
13	PCHLORICC250	PD088132.D	18 Apr 2025 15:46	AR\AJ	Ok
14	PCHLORICC050	PD088133.D	18 Apr 2025 15:59	AR\AJ	Ok,M
15	PTOXICC1000	PD088134.D	18 Apr 2025 16:13	AR\AJ	Ok,M
16	PTOXICC750	PD088135.D	18 Apr 2025 16:27	AR\AJ	Ok
17	PTOXICC500	PD088136.D	18 Apr 2025 16:40	AR\AJ	Ok
18	PTOXICC250	PD088137.D	18 Apr 2025 16:54	AR\AJ	Ok,M
19	PTOXICC100	PD088138.D	18 Apr 2025 17:08	AR\AJ	Ok,M
20	PSTDICV050	PD088139.D	18 Apr 2025 17:21	AR\AJ	Ok
21	PCHLORICV500	PD088140.D	18 Apr 2025 17:35	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD041825

Review By	Abdul	Review On	4/19/2025 2:07:56 PM
Supervise By	mohammad	Supervise On	4/22/2025 2:20:46 AM
SubDirectory	PD041825	HP Acquire Method	HP Processing Method pd041825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PD088141.D	18 Apr 2025 17:49	AR\AJ	Ok
23	I.BLK	PD088142.D	18 Apr 2025 18:02	AR\AJ	Ok
24	PEM	PD088143.D	18 Apr 2025 18:16	AR\AJ	Ok,M
25	PSTDCCC050	PD088144.D	18 Apr 2025 18:30	AR\AJ	Not Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD051325

Review By	Abdul	Review On	5/14/2025 8:34:03 AM
Supervise By		Supervise On	
SubDirectory	PD051325	HP Acquire Method	HP Processing Method pd041825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD088496.D	13 May 2025 09:14	AR\AJ	Ok
2	I.BLK	PD088497.D	13 May 2025 09:28	AR\AJ	Ok
3	PEM	PD088498.D	13 May 2025 09:41	AR\AJ	Ok,NS
4	PSTDCCC050	PD088499.D	13 May 2025 09:55	AR\AJ	Ok,NS
5	PB167950BL	PD088500.D	13 May 2025 10:08	AR\AJ	Ok
6	Q1982-01	PD088501.D	13 May 2025 10:22	AR\AJ	Ok
7	Q1982-02	PD088502.D	13 May 2025 10:35	AR\AJ	Ok,NS
8	Q1982-04	PD088503.D	13 May 2025 10:49	AR\AJ	Ok
9	Q1982-05	PD088504.D	13 May 2025 11:02	AR\AJ	Ok
10	Q1982-07	PD088505.D	13 May 2025 11:16	AR\AJ	Ok,NS
11	PB167925BL	PD088506.D	13 May 2025 11:29	AR\AJ	Ok
12	Q1983-01DL	PD088507.D	13 May 2025 11:43	AR\AJ	Ok,NS
13	PB167969BL	PD088508.D	13 May 2025 11:57	AR\AJ	Ok
14	PB167969BS	PD088509.D	13 May 2025 12:10	AR\AJ	Ok
15	PB167851TB	PD088510.D	13 May 2025 12:24	AR\AJ	Ok
16	Q1956-05	PD088511.D	13 May 2025 12:37	AR\AJ	Ok
17	Q1956-05MS	PD088512.D	13 May 2025 12:51	AR\AJ	Ok,NS
18	Q1956-05MSD	PD088513.D	13 May 2025 13:04	AR\AJ	Ok
19	I.BLK	PD088514.D	13 May 2025 13:18	AR\AJ	Ok
20	PSTDCCC050	PD088515.D	13 May 2025 14:22	AR\AJ	Ok,NS
21	Q2007-06	PD088516.D	13 May 2025 14:35	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD051325

Review By	Abdul	Review On	5/14/2025 8:34:03 AM
Supervise By		Supervise On	
SubDirectory	PD051325	HP Acquire Method	HP Processing Method pd041825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q2007-12	PD088517.D	13 May 2025 14:49	AR\AJ	Ok,NS
23	Q2007-18	PD088518.D	13 May 2025 15:03	AR\AJ	Ok
24	Q2007-24	PD088519.D	13 May 2025 15:16	AR\AJ	Ok
25	Q2007-30	PD088520.D	13 May 2025 15:30	AR\AJ	Ok
26	Q2007-36	PD088521.D	13 May 2025 15:44	AR\AJ	Ok,NS
27	PB167925BS	PD088522.D	13 May 2025 16:16	AR\AJ	Ok
28	I.BLK	PD088523.D	13 May 2025 16:35	AR\AJ	Ok
29	PSTDCCC050	PD088524.D	13 May 2025 16:48	AR\AJ	Ok,NS
30	PB167959BL	PD088525.D	13 May 2025 17:02	AR\AJ	Ok
31	PB167959BS	PD088526.D	13 May 2025 17:15	AR\AJ	Ok
32	Q1982-08	PD088527.D	13 May 2025 17:29	AR\AJ	Ok,NS
33	Q2010-01	PD088528.D	13 May 2025 17:43	AR\AJ	Ok,NS
34	Q2010-02	PD088529.D	13 May 2025 17:56	AR\AJ	Ok
35	Q2010-03	PD088530.D	13 May 2025 18:10	AR\AJ	Ok
36	Q2010-04	PD088531.D	13 May 2025 18:24	AR\AJ	Ok
37	Q2014-01	PD088532.D	13 May 2025 18:38	AR\AJ	Ok,NS
38	Q2014-03	PD088533.D	13 May 2025 18:51	AR\AJ	Ok,NS
39	Q2014-05	PD088534.D	13 May 2025 19:05	AR\AJ	Not Ok
40	I.BLK	PD088535.D	13 May 2025 19:18	AR\AJ	Ok
41	PEM	PD088536.D	13 May 2025 19:32	AR\AJ	Ok,NS
42	PSTDCCC050	PD088537.D	13 May 2025 19:46	AR\AJ	Ok,NS
43	Q2017-01	PD088538.D	13 May 2025 19:59	AR\AJ	Ok
44	Q2017-05	PD088539.D	13 May 2025 20:13	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD051325

Review By	Abdul	Review On	5/14/2025 8:34:03 AM
Supervise By		Supervise On	
SubDirectory	PD051325	HP Acquire Method	HP Processing Method pd041825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	I.BLK	PD088540.D	13 May 2025 20:27	AR\AJ	Ok
46	PSTDCCC050	PD088541.D	13 May 2025 20:40	AR\AJ	Ok,NS

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD041825

Review By	Abdul	Review On	4/19/2025 2:07:56 PM
Supervise By	mohammad	Supervise On	4/22/2025 2:20:46 AM
SubDirectory	PD041825	HP Acquire Method	HP Processing Method pd041825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24095
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD088120.D	18 Apr 2025 13:02		AR\AJ	Ok
2	I.BLK	I.BLK	PD088121.D	18 Apr 2025 13:15		AR\AJ	Ok
3	PEM	PEM	PD088122.D	18 Apr 2025 13:29		AR\AJ	Ok,M
4	RESCHK	RESCHK	PD088123.D	18 Apr 2025 13:43		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PD088124.D	18 Apr 2025 13:56		AR\AJ	Ok,M
6	PSTDICC075	PSTDICC075	PD088125.D	18 Apr 2025 14:10		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PD088126.D	18 Apr 2025 14:24		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PD088127.D	18 Apr 2025 14:37		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PD088128.D	18 Apr 2025 14:51		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PD088129.D	18 Apr 2025 15:05		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PD088130.D	18 Apr 2025 15:18		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PD088131.D	18 Apr 2025 15:32		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PD088132.D	18 Apr 2025 15:46		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PD088133.D	18 Apr 2025 15:59		AR\AJ	Ok,M
15	PTOXICC1000	PTOXICC1000	PD088134.D	18 Apr 2025 16:13		AR\AJ	Ok,M
16	PTOXICC750	PTOXICC750	PD088135.D	18 Apr 2025 16:27		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PD088136.D	18 Apr 2025 16:40		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PD088137.D	18 Apr 2025 16:54		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD041825

Review By	Abdul	Review On	4/19/2025 2:07:56 PM
Supervise By	mohammad	Supervise On	4/22/2025 2:20:46 AM
SubDirectory	PD041825	HP Acquire Method	HP Processing Method pd041825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PD088138.D	18 Apr 2025 17:08		AR\AJ	Ok,M
20	PSTDICV050	ICVPD041825	PD088139.D	18 Apr 2025 17:21		AR\AJ	Ok
21	PCHLORICV500	ICVPD041825CHLOR	PD088140.D	18 Apr 2025 17:35		AR\AJ	Ok
22	PTOXICV500	ICVPD041825TOX	PD088141.D	18 Apr 2025 17:49		AR\AJ	Ok
23	I.BLK	I.BLK	PD088142.D	18 Apr 2025 18:02		AR\AJ	Ok
24	PEM	PEM	PD088143.D	18 Apr 2025 18:16		AR\AJ	Ok,M
25	PSTDCCC050	PSTDCCC050	PD088144.D	18 Apr 2025 18:30	ccc high	AR\AJ	Not Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD051325

Review By	Abdul	Review On	5/14/2025 8:34:03 AM
Supervise By		Supervise On	
SubDirectory	PD051325	HP Acquire Method	HP Processing Method pd041825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD088496.D	13 May 2025 09:14		AR\AJ	Ok
2	I.BLK	I.BLK	PD088497.D	13 May 2025 09:28		AR\AJ	Ok
3	PEM	PEM	PD088498.D	13 May 2025 09:41		AR\AJ	Ok,NS
4	PSTDCCC050	PSTDCCC050	PD088499.D	13 May 2025 09:55		AR\AJ	Ok,NS
5	PB167950BL	PB167950BL	PD088500.D	13 May 2025 10:08		AR\AJ	Ok
6	Q1982-01	TP-1	PD088501.D	13 May 2025 10:22		AR\AJ	Ok
7	Q1982-02	TP-2B	PD088502.D	13 May 2025 10:35		AR\AJ	Ok,NS
8	Q1982-04	TP-4	PD088503.D	13 May 2025 10:49		AR\AJ	Ok
9	Q1982-05	TP-5	PD088504.D	13 May 2025 11:02		AR\AJ	Ok
10	Q1982-07	TP-8	PD088505.D	13 May 2025 11:16		AR\AJ	Ok,NS
11	PB167925BL	PB167925BL	PD088506.D	13 May 2025 11:29		AR\AJ	Ok
12	Q1983-01DL	OR-636-COMP-01DL	PD088507.D	13 May 2025 11:43		AR\AJ	Ok,NS
13	PB167969BL	PB167969BL	PD088508.D	13 May 2025 11:57		AR\AJ	Ok
14	PB167969BS	PB167969BS	PD088509.D	13 May 2025 12:10		AR\AJ	Ok
15	PB167851TB	PB167851TB	PD088510.D	13 May 2025 12:24		AR\AJ	Ok
16	Q1956-05	COMP1	PD088511.D	13 May 2025 12:37		AR\AJ	Ok
17	Q1956-05MS	COMP1MS	PD088512.D	13 May 2025 12:51		AR\AJ	Ok,NS
18	Q1956-05MSD	COMP1MSD	PD088513.D	13 May 2025 13:04		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD051325

Review By	Abdul	Review On	5/14/2025 8:34:03 AM
Supervise By		Supervise On	
SubDirectory	PD051325	HP Acquire Method	HP Processing Method pd041825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	I.BLK	I.BLK	PD088514.D	13 May 2025 13:18		AR\AJ	Ok
20	PSTDCCC050	PSTDCCC050	PD088515.D	13 May 2025 14:22		AR\AJ	Ok,NS
21	Q2007-06	OR-636-COMP-10	PD088516.D	13 May 2025 14:35		AR\AJ	Ok
22	Q2007-12	OR-636-COMP-11	PD088517.D	13 May 2025 14:49		AR\AJ	Ok,NS
23	Q2007-18	OR-636-COMP-12	PD088518.D	13 May 2025 15:03		AR\AJ	Ok
24	Q2007-24	OR-636-COMP-13	PD088519.D	13 May 2025 15:16		AR\AJ	Ok
25	Q2007-30	OR-636-COMP-14	PD088520.D	13 May 2025 15:30		AR\AJ	Ok
26	Q2007-36	OR-636-COMP-15	PD088521.D	13 May 2025 15:44		AR\AJ	Ok,NS
27	PB167925BS	PB167925BS	PD088522.D	13 May 2025 16:16		AR\AJ	Ok
28	I.BLK	I.BLK	PD088523.D	13 May 2025 16:35		AR\AJ	Ok
29	PSTDCCC050	PSTDCCC050	PD088524.D	13 May 2025 16:48		AR\AJ	Ok,NS
30	PB167959BL	PB167959BL	PD088525.D	13 May 2025 17:02		AR\AJ	Ok
31	PB167959BS	PB167959BS	PD088526.D	13 May 2025 17:15		AR\AJ	Ok
32	Q1982-08	TP-9	PD088527.D	13 May 2025 17:29		AR\AJ	Ok,NS
33	Q2010-01	TP-10	PD088528.D	13 May 2025 17:43		AR\AJ	Ok,NS
34	Q2010-02	TP-13	PD088529.D	13 May 2025 17:56		AR\AJ	Ok
35	Q2010-03	TP-14	PD088530.D	13 May 2025 18:10		AR\AJ	Ok
36	Q2010-04	TP-17	PD088531.D	13 May 2025 18:24		AR\AJ	Ok
37	Q2014-01	MOO-25-0134	PD088532.D	13 May 2025 18:38		AR\AJ	Ok,NS

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD051325

Review By	Abdul	Review On	5/14/2025 8:34:03 AM
Supervise By		Supervise On	
SubDirectory	PD051325	HP Acquire Method	HP Processing Method pd041825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	Q2014-03	MOO-25-0145	PD088533.D	13 May 2025 18:51		AR\AJ	Ok,NS
39	Q2014-05	MOO-25-0148	PD088534.D	13 May 2025 19:05	F flag in TCX	AR\AJ	Not Ok
40	I.BLK	I.BLK	PD088535.D	13 May 2025 19:18		AR\AJ	Ok
41	PEM	PEM	PD088536.D	13 May 2025 19:32		AR\AJ	Ok,NS
42	PSTDCCC050	PSTDCCC050	PD088537.D	13 May 2025 19:46		AR\AJ	Ok,NS
43	Q2017-01	MH-I	PD088538.D	13 May 2025 19:59		AR\AJ	Ok
44	Q2017-05	MH-J	PD088539.D	13 May 2025 20:13		AR\AJ	Ok
45	I.BLK	I.BLK	PD088540.D	13 May 2025 20:27		AR\AJ	Ok
46	PSTDCCC050	PSTDCCC050	PD088541.D	13 May 2025 20:40		AR\AJ	Ok,NS

M : Manual Integration


SOP ID :	M1311-TCLP-15	
SDG No :	N/A	Start Prep Date : 05/05/2025 Time : 15:00
Welgh By :	JP	End Prep Date : 05/06/2025 Time : 08:35
Balance ID :	WC SC-7	Combination Ratio : 20
pH Meter ID :	WC PH METER-1	ZHE Cleaning Batch : ⁷⁰ N/A
Extraction By :	JP	Initial Room Temperature: 23 °C
Filter By :	JP	Final Room Temperature: 22 °C
Pippete ID :	WC	TCLP Technician Signature : 
Tumbler ID :	T-1	Supervisor By : 
TCLP Filter ID :	115525	

Standard Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP112795
HCL-TCLP,1N	N/A	WP112797
HNO3-TCLP,1N	N/A	WP112799
pH Strips	N/A	W1931,W1934,W3171,W3172
pH Strips	W1940,W1941,W1942	W3166,W1938,W1939,
1 Liter Amber	N/A	90924-08
120ml Plastic bottle	N/A	2738
1:1 HNO3	N/A	MP84041

Extraction Conformance/Non-Conformance Comments:

Matrix sikes are added after fiftation and tumbling. TUMBLER T-1 checked,30 rpm. Particle size reduction is not required. Q1960-03 IS USED FOR MS-MSD.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
05/06/25 11:00	 /ECP Room	S/S.  EXT
	Preparation Group	Analysis Group 

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB167851TB	LEB851	07	N/A	2000	N/A	N/A	N/A	4.93	1.0	T-1
Q1947-04	MH-P	01	100.02	2000	N/A	N/A	N/A	3.0	1.5	T-1
Q1947-08	MH-O	02	100.03	2000	N/A	N/A	N/A	3.5	1.0	T-1
Q1951-04	MH-KK	03	100.04	2000	N/A	N/A	N/A	3.5	1.5	T-1
Q1956-05	COMP1	04	100.02	2000	N/A	N/A	N/A	5.6	1.5	T-1
Q1959-01	SOIL-PILE	05	100.03	2000	N/A	N/A	N/A	5.5	1.0	T-1
Q1960-03	72-11933	06	100.02	2000	N/A	N/A	N/A	5.6	1.5	T-1

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
PB167851TB	LEB851	N/A	N/A	N/A	N/A	N/A	N/A
Q1947-04	MH-P	N/A	N/A	N/A	N/A	100	N/A
Q1947-08	MH-O	N/A	N/A	N/A	N/A	100	N/A
Q1951-04	MH-KK	N/A	N/A	N/A	N/A	100	N/A
Q1956-05	COMP1	N/A	N/A	N/A	N/A	100	N/A
Q1959-01	SOIL-PILE	N/A	N/A	N/A	N/A	100	N/A
Q1960-03	72-11933	N/A	N/A	N/A	N/A	100	N/A

Hot Block ID : WC S-1 / WC S-2

Thermometer ID : N/A

SampleID	ClientID	Sample Weight (g)	Volume DI Water (mL)	PH after 5 min stir	PH after 10 min stir	Extraction Fluid 1 or 2	pH Extraction Fluid
PB167851TB	LEB851	N/A	N/A	N/A	N/A	#1	4.93
Q1947-04	MH-P	5.02	96.5	5.6	1.5	#1	4.93
Q1947-08	MH-O	5.01	96.5	6.2	2.0	#1	4.93
Q1951-04	MH-KK	5.02	96.5	6.0	2.0	#1	4.93
Q1956-05	COMP1	5.02	96.5	7.6	3.0	#1	4.93
Q1959-01	SOIL-PILE	5.01	96.5	7.2	3.0	#1	4.93
Q1960-03	72-11933	5.02	96.5	7.0	3.0	#1	4.93

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tcip q1956

WorkList ID : 189319

Department : TCLP Extraction

Date : 05-05-2025 11:14:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1947-04	MH-P	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L41	05/02/2025	1311
Q1947-08	MH-O	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L41	05/02/2025	1311
Q1951-04	MH-KK	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L31	05/02/2025	1311
Q1956-05	COMP1	Solid	TCLP Extraction	Cool 4 deg C	CAMP02	L31	05/02/2025	1311
Q1959-01	SOIL-PILE	Solid	TCLP Extraction	Cool 4 deg C	PSEG04	L41	05/05/2025	1311
Q1960-03	72-11933	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L51	05/05/2025	1311

Date/Time 05/05/25 14:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 05/05/25 17:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

SOP ID: M3510C,3580A-Extraction Pesticide-16
Clean Up SOP #: N/A
Matrix : Water
Welgh By: N/A **Extraction By:** RS
Balance check: N/A **Filter By:** RS
Balance ID: N/A **pH Meter ID:** N/A
pH Strip Lot#: E3880 **Hood ID:** 4,5,6,7
Extraction Method: ☒ Separatory Funnel ☐ Continious Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☐ Soxhlet

Extraction Start Date : 05/12/2025
Extraction Start Time : 12:25
Extraction End Date : 05/12/2025
Extraction End Time : 17:10
Concentration By: EH
Supervisor By : RUPESH

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	500 PPB	PP24285
Surrogate	1.0ML	200 PPB	PP24460
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3930
Baked Na2SO4	N/A	EP2611
Hexane	N/A	E3933
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS723.

KD Bath ID: WATER BATH-1,2 **Envap ID:** NEVAP-02
KD Bath Temperature: 60 °C **Envap Temperature:** 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
5/12/25	RS (Ext-Lab)	SR Pest/PCB Lab
17:15	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction Pesticide-16

Concentration Date: 05/12/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB167851TB	PB167851TB	TCLP Pesticide	100	6	RUPESH	ritesh	10			SEP-1
PB167943TB	PB167943TB	TCLP Pesticide	100	6	RUPESH	ritesh	10			2
PB167969BL	PBLK969	TCLP Pesticide	1000	6	RUPESH	ritesh	10			3
PB167969BS	PLCS969	TCLP Pesticide	1000	6	RUPESH	ritesh	10			4
Q1956-05	COMP1	TCLP Pesticide	100	6	RUPESH	ritesh	10	A		5
Q1956-05MS	COMP1MS	TCLP Pesticide	100	6	RUPESH	ritesh	10	A		6
Q1956-05MS D	COMP1MSD	TCLP Pesticide	100	6	RUPESH	ritesh	10	A		7
Q2007-06	OR-636-COMP-10	TCLP Pesticide	100	6	RUPESH	ritesh	10	A		8
Q2007-12	OR-636-COMP-11	TCLP Pesticide	100	6	RUPESH	ritesh	10	A		9
Q2007-18	OR-636-COMP-12	TCLP Pesticide	100	6	RUPESH	ritesh	10	A		10
Q2007-24	OR-636-COMP-13	TCLP Pesticide	100	6	RUPESH	ritesh	10	A		11
Q2007-30	OR-636-COMP-14	TCLP Pesticide	100	6	RUPESH	ritesh	10	A		12
Q2007-36	OR-636-COMP-15	TCLP Pesticide	100	6	RUPESH	ritesh	10	A		13



Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB167851TB	LEB851	07	N/A	2000	N/A	N/A	N/A	4.93	1.0	T-1
Q1947-04	MH-P	01	100.02	2000	N/A	N/A	N/A	3.0	1.5	T-1
Q1947-08	MH-O	02	100.03	2000	N/A	N/A	N/A	3.5	1.0	T-1
Q1951-04	MH-KK	03	100.04	2000	N/A	N/A	N/A	3.5	1.5	T-1
Q1956-05	COMP1	04	100.02	2000	N/A	N/A	N/A	5.6	1.5	T-1
Q1959-01	SOIL-PILE	05	100.03	2000	N/A	N/A	N/A	5.5	1.0	T-1
Q1960-03	72-11933	06	100.02	2000	N/A	N/A	N/A	5.6	1.5	T-1

05/06/25
112.00

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB167943TB	LEB943	13	N/A	2000	N/A	N/A	N/A	4.94	1.0	T-2
Q2004-02	VNJ-220	01	100.02	2000	N/A	N/A	N/A	3.5	1.5	T-1
Q2004-04	2811	02	100.03	2000	N/A	N/A	N/A	3.0	1.0	T-1
Q2004-06	R0202	03	100.04	2000	N/A	N/A	N/A	5.8	1.5	T-1
Q2004-08	205510	04	100.02	2000	N/A	N/A	N/A	5.0	1.0	T-1
Q2004-10	SOIL-STOCK	05	100.03	2000	N/A	N/A	N/A	5.0	1.5	T-1
Q2004-12	STONE-STOCK	06	100.02	2000	N/A	N/A	N/A	6.2	1.5	T-1
Q2007-06	OR-636-COMP-10	07	100.01	2000	N/A	N/A	N/A	3.0	1.0	T-1
Q2007-12	OR-636-COMP-11	08	100.02	2000	N/A	N/A	N/A	5.6	1.0	T-1
Q2007-18	OR-636-COMP-12	09	100.03	2000	N/A	N/A	N/A	3.5	1.5	T-1
Q2007-24	OR-636-COMP-13	10	100.04	2000	N/A	N/A	N/A	5.6	1.5	T-1
Q2007-30	OR-636-COMP-14	11	100.03	2000	N/A	N/A	N/A	3.5	1.0	T-2
Q2007-36	OR-636-COMP-15	12	100.02	2000	N/A	N/A	N/A	5.6	1.5	T-2

05/12/15

11:30



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: CDM Smith
ADDRESS: 110 Fieldcrest Avenue #8 6th Floor
CITY Edison STATE: NJ ZIP: 08837
ATTENTION: Marcie Encinas
PHONE: 732-590-4679 FAX: 732-225-7851

CLIENT PROJECT INFORMATION

PROJECT NAME: Bergen Point Project
PROJECT NO.: 99939 LOCATION: W. Babylon, NY
PROJECT MANAGER: Marcie Encinas
e-mail: ENcinasMA@CDMSmith.com
PHONE: 732-590-4679 FAX: 732-225-7851

CLIENT BILLING INFORMATION

BILL TO: CDM Smith PO#:
ADDRESS: 110 Fieldcrest Avenue, #8 Floor 6th
CITY Edison STATE: NJ ZIP: 08837
ATTENTION: Marcie Encinas PHONE: 732-590-4690

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: _____ DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☒ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☒ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other _____
☐ EDD FORMAT _____

1. TCL VOC
2. TCL SVOC
3. TAL METALS
4. PCBs
5. PESTICIDES
6. HERBICIDES
7. PRO/GRO
8. FULL TCLP
9. RCRA CHARAC

PRESERVATIVES

COMMENTS

← Specify Preservatives
A-HCl D-NaOH
B-HNO3 E-ICE
C-H2SO4 F-OTHER

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES										
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	SB1-3-4'	Soil		X	5/1/25	9:00AM	13	X	X	X	X	X	X	X			E
2.	SB2-4-5'	Soil		X	5/2/25	12:00PM	13	X	X	X	X	X	X	X			E
3.	COMP1	Soil	X		5/2/25	1:30PM	3								X	X	E
4.	SB91-3-4'	Soil		X	5/1/25	9:30AM	13	X	X	X	X	X	X	X	X		E
5.	SB2-MS	Soil		X	5/2/25	12:30PM	13	X	X	X	X	X	X	X			E
6.	SB2-MSD	Soil		X	5/2/25	12:40PM	13	X	X	X	X	X	X	X			E
7.	FB-05022025	Aqueous		X	5/2/25	11:00AM	9	X	X	X	X	X	X	X			E, A, B
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <i>[Signature]</i>	DATE/TIME: 1404 5.2.2025	RECEIVED BY: <i>[Signature]</i>	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☒ COOLER TEMP 3 °C
RELINQUISHED BY SAMPLER: 2. <i>[Signature]</i>	DATE/TIME:	RECEIVED BY:	Comments: IR Can #1 (Adjusted Factor + 1)
RELINQUISHED BY SAMPLER: 3. <i>[Signature]</i>	DATE/TIME: 1820 5.2.2025	RECEIVED BY:	

Page ____ of CLIENT: ☐ Hand Delivered ☐ Other Shipment Complete ☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1956	CAMP02	Order Date : 5/2/2025 3:14:35 PM	Project Mgr :
Client Name : CDM Smith		Project Name : Bergen Point Fueling System	Report Type : NYS ASP A
Client Contact : Marcie Ann Encinas		Receive DateTime : 5/2/2025 12:00:00 AM	EDD Type : EQUIS
Invoice Name : CDM Smith		Purchase Order : 18:26	Hard Copy Date :
Invoice Contact : Marcie Ann Encinas			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1956-01	SB1-3-4	Solid	05/01/2025	09:00	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q1956-02	SB2-4-5	Solid	05/02/2025	12:00	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q1956-03	Q1956-02MS	Solid	05/02/2025	12:30	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q1956-04	Q1956-02MSD	Solid	05/02/2025	12:40	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q1956-06	SB19-3-4 SB91	Solid	05/01/2025	09:30	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q1956-07	FB-05022025	Water	05/02/2025	11:00	VOC-TCLVOA-10		8260-Low	10 Bus. Days	

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1956	CAMP02	Order Date : 5/2/2025 3:14:35 PM	Project Mgr :
Client Name : CDM Smith		Project Name : Bergen Point Fueling System	Report Type : NYS ASP A
Client Contact : Marcie Ann Encinas		Receive DateTime : 5/2/2025 12:00:00 AM	EDD Type : EQUIS
Invoice Name : CDM Smith		Purchase Order : 18-26	Hard Copy Date :
Invoice Contact : Marcie Ann Encinas			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1956-09 TB		water	5/2/25	12:00	VOC-TCLVOA-10		8260-LOW		10 Bus Days.

Relinquished By : 

Date / Time : 5/5/25 0835

Received By : 

Date / Time : 5/5/25 08:35

Storage Area : VOA Refridgerator Room

Rep # 6
FZ-2