

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/07/2025

Lab File ID: BM050129.D Time Analyzed: 09:56

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	266306	7.746	985466	10.53	648787	14.39
UPPER LIMIT	532612	8.246	1970930	11.034	1297570	14.892
LOWER LIMIT	133153	7.246	492733	10.034	324394	13.892
EPA SAMPLE NO.						
01 PB167879BL	266570	7.75	933974	10.54	599928	14.39
02 PB167879BS	261504	7.75	899945	10.53	545059	14.39
03 FB-05022025	270666	7.75	990337	10.54	667274	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.