

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q1956

SDG NO.: Q1956

Lab File ID: BM050128.D

DFTPP Injection Date: 05/07/2025

Instrument ID: BNA_M

DFTPP Injection Time: 09:17

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	23
68	Less than 2.0% of mass 69	0.4 (1.4) 1
69	Mass 69 relative abundance	27.7
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	35
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	26.1
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	12
442	Greater than 50% of mass 198	74.9
443	15.0 - 24.0% of mass 442	14.6 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM050129.D	05/07/2025	09:56
PB167879BL	PB167879BL	BM050130.D	05/07/2025	10:35
PB167879BS	PB167879BS	BM050131.D	05/07/2025	11:14
FB-05022025	Q1956-07	BM050136.D	05/07/2025	15:44