

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1956

Client: CDM Smith

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1947-04MSD	Client Sample ID:	MH-PMSD					DataFile:	BP024565.D		
Pyridine	500	0	320	ug/L	64	3			10	109	20
1,4-Dichlorobenzene	500	0	400	ug/L	80	2			55	125	20
2-Methylphenol	500	0	430	ug/L	86	5			60	131	20
3+4-Methylphenols	500	0	420	ug/L	84	2			54	136	20
Hexachloroethane	500	0	410	ug/L	82	5			19	146	20
Nitrobenzene	500	0	420	ug/L	84	7			62	112	20
Hexachlorobutadiene	500	0	470	ug/L	94	4			52	125	20
2,4,6-Trichlorophenol	500	0	480	ug/L	96	6			78	112	20
2,4,5-Trichlorophenol	500	0	480	ug/L	96	4			71	111	20
2,4-Dinitrotoluene	500	0	510	ug/L	102	9			74	137	20
Hexachlorobenzene	500	0	510	ug/L	102	2			72	115	20
Pentachlorophenol	1000	0	990	ug/L	99	1			52	162	20