

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1956

Client: CDM Smith

Analytical Method: 8270E DataFile: BP024560.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167886BS	Pyridine	50	32.7	ug/L	65				29	97	
	1,4-Dichlorobenzene	50	43.5	ug/L	87				76	103	
	2-Methylphenol	50	48.9	ug/L	98				69	109	
	3+4-Methylphenols	50	47.5	ug/L	95				67	106	
	Hexachloroethane	50	45.3	ug/L	91				76	118	
	Nitrobenzene	50	43.1	ug/L	86				58	106	
	Hexachlorobutadiene	50	49.6	ug/L	99				69	101	
	2,4,6-Trichlorophenol	50	50.9	ug/L	102				61	110	
	2,4,5-Trichlorophenol	50	50.5	ug/L	101				70	106	
	2,4-Dinitrotoluene	50	49.1	ug/L	98				60	115	
	Hexachlorobenzene	50	51.8	ug/L	104				73	106	
	Pentachlorophenol	100	98.3	ug/L	98				47	114	