

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q1956

SDG NO.: Q1956

Lab File ID: BM050139.D

DFTPP Injection Date: 05/08/2025

Instrument ID: BNA\_M

DFTPP Injection Time: 09:59

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	21.7
68	Less than 2.0% of mass 69	0.4 ( 1.4 ) 1
69	Mass 69 relative abundance	25.8
70	Less than 2.0% of mass 69	0.1 ( 0.4 ) 1
127	10.0 - 80.0% of mass 198	33.5
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	27.3
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	13.6
442	Greater than 50% of mass 198	85.9
443	15.0 - 24.0% of mass 442	16.8 ( 19.5 ) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM050140.D	05/08/2025	10:41
PB167851TB	PB167851TB	BM050144.D	05/08/2025	13:18