

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG No.: Q1956

Instrument ID: MSVOA_N Calibration Date/Time: 05/06/2025 10:35

Lab File ID: VN086504.D Init. Calib. Date(s): 04/15/2025 04/15/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:29 14:29

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.822	0.657		-20.07	20
1,1-Dichloroethene	0.588	0.522		-11.22	20
2-Butanone	0.451	0.405		-10.2	20
Carbon Tetrachloride	0.441	0.447		1.36	20
Chloroform	1.180	1.187		0.59	20
Benzene	1.485	1.462		-1.55	20
1,2-Dichloroethane	0.474	0.511		7.81	20
Trichloroethene	0.352	0.379		7.67	20
Tetrachloroethene	0.385	0.415		7.79	20
Chlorobenzene	1.116	1.152	0.3	3.23	20
1,2-Dichloroethane-d4	0.725	0.698		-3.72	20
Dibromofluoromethane	0.232	0.250		7.76	20
Toluene-d8	1.240	1.178		-5	20
4-Bromofluorobenzene	0.452	0.460		1.77	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.