

Surrogate Summary

SDG No.: Q1956

Client: CDM Smith

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1956-05	COMP1	1,2-Dichloroethane-d4	50	47.6	95	74	125
		Dibromofluoromethane	50	56.2	112	75	124
		Toluene-d8	50	51.1	102	86	113
		4-Bromofluorobenzene	50	51.3	103	77	121
VN0506WBL01	VN0506WBL01	1,2-Dichloroethane-d4	50	46.0	92	74	125
		Dibromofluoromethane	50	59.9	120	75	124
		Toluene-d8	50	50.5	101	86	113
		4-Bromofluorobenzene	50	50.2	100	77	121
VN0506WBS01	VN0506WBS01	1,2-Dichloroethane-d4	50	46.3	93	74	125
		Dibromofluoromethane	50	57.8	116	75	124
		Toluene-d8	50	48.7	97	86	113
		4-Bromofluorobenzene	50	46.2	92	77	121
VN0506WBSD0	VN0506WBSD01	1,2-Dichloroethane-d4	50	48.4	97	74	125
		Dibromofluoromethane	50	58.0	116	75	124
		Toluene-d8	50	48.4	97	86	113
		4-Bromofluorobenzene	50	49.6	99	77	121