

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1956

Client: CDM Smith

Analytical Method: SW8260D

Datafile : VN086507.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0506WBS01	Vinyl chloride	20	15.5	ug/L	78			65	117	
	1,1-Dichloroethene	20	17.3	ug/L	86			74	110	
	2-Butanone	100	79.6	ug/L	80			65	122	
	Carbon Tetrachloride	20	20.2	ug/L	101			77	113	
	Chloroform	20	17.9	ug/L	90			79	113	
	Benzene	20	18.4	ug/L	92			82	109	
	1,2-Dichloroethane	20	19.2	ug/L	96			80	115	
	Trichloroethene	20	19.7	ug/L	99			77	113	
	Tetrachloroethene	20	20.4	ug/L	102			67	123	
	Chlorobenzene	20	19.4	ug/L	97			82	109	