

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1956

Client: CDM Smith

Analytical Method: SW8260D

Datafile : VN086508.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0506WBSD01	Vinyl chloride	20	15.7	ug/L	79	1		65	117	20
	1,1-Dichloroethene	20	17.3	ug/L	86	0		74	110	20
	2-Butanone	100	84.5	ug/L	85	6		65	122	20
	Carbon Tetrachloride	20	20.3	ug/L	102	1		77	113	20
	Chloroform	20	19.0	ug/L	95	5		79	113	20
	Benzene	20	18.9	ug/L	95	3		82	109	20
	1,2-Dichloroethane	20	20.5	ug/L	103	7		80	115	20
	Trichloroethene	20	20.3	ug/L	102	3		77	113	20
	Tetrachloroethene	20	21.1	ug/L	106	4		67	123	20
	Chlorobenzene	20	20.2	ug/L	101	4		82	109	20