

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02  
 Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956  
 Lab File ID: VN086504.D Date Analyzed: 05/06/2025  
 Instrument ID: MSVOA\_N Time Analyzed: 10:35  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 186971        | 8.22  | 338775        | 9.09  | 314072        | 11.86  |
| UPPER LIMIT    | 373942        | 8.724 | 677550        | 9.594 | 628144        | 12.359 |
| LOWER LIMIT    | 93485.5       | 7.724 | 169388        | 8.594 | 157036        | 11.359 |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| COMP1          | 173251        | 8.22  | 324303        | 9.10  | 314398        | 11.87  |
| VN0506WBL01    | 200028        | 8.22  | 363115        | 9.10  | 340109        | 11.86  |
| VN0506WBS01    | 202593        | 8.22  | 354153        | 9.10  | 311333        | 11.87  |
| VN0506WBSD01   | 187036        | 8.22  | 333546        | 9.10  | 299527        | 11.87  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.