

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG No.: Q1956

Instrument ID: MSVOA_Y Calibration Date/Time: 05/05/2025 08:46

Lab File ID: VY022145.D Init. Calib. Date(s): 04/22/2025 04/22/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:39 16:15

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Acetone	250.000	257.73	-3.091199	20