

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956
Lab File ID: VN086478.D Date Analyzed: 05/05/2025
Instrument ID: MSVOA_N Time Analyzed: 11:32
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	154484	8.22	297307	9.10	276417	11.87
UPPER LIMIT	308968	8.718	594614	9.6	552834	12.365
LOWER LIMIT	77242	7.718	148654	8.6	138209	11.365
EPA SAMPLE NO.						
FB-05022025	141270	8.22	272905	9.10	255761	11.87
VN0505WBL01	156438	8.22	296282	9.10	275019	11.87
VN0505WBS01	196497	8.22	366604	9.10	333505	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.