

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02  
 Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956  
 Lab File ID: VX046181.D Date Analyzed: 05/14/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 09:59  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 92522         | 5.54  | 155707        | 6.76  | 140519        | 10.05  |
| UPPER LIMIT    | 185044        | 6.043 | 311414        | 7.257 | 281038        | 10.549 |
| LOWER LIMIT    | 46261         | 5.043 | 77853.5       | 6.257 | 70259.5       | 9.549  |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| TB             | 65765         | 5.55  | 132071        | 6.76  | 124301        | 10.05  |
| VX0514WBL01    | 66143         | 5.55  | 133003        | 6.76  | 124085        | 10.05  |
| VX0514WBS01    | 87880         | 5.55  | 155746        | 6.76  | 138245        | 10.05  |
| VX0514WBSD01   | 86737         | 5.54  | 152178        | 6.76  | 134818        | 10.05  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.