

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956
 Lab File ID: VN086478.D Date Analyzed: 05/05/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:32
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	133852	13.788				
UPPER LIMIT	267704	14.288				
LOWER LIMIT	66926	13.288				
EPA SAMPLE NO.						
FB-05022025	111305	13.79				
VN0505WBL01	110905	13.79				
VN0505WBS01	147574	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.