

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q1956 SAS No.: Q1956 SDG NO.: Q1956
 Lab File ID: VY022145.D Date Analyzed: 05/05/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:46
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	221579	13.347				
UPPER LIMIT	443158	13.847				
LOWER LIMIT	110790	12.847				
EPA SAMPLE NO.						
SB1-3-4	144961	13.35				
SB2-4-5	140796	13.35				
SB2-4-5MS	125407	13.35				
SB2-4-5MSD	123121	13.35				
SB91-3-4	143859	13.35				
VY0505SBL01	191516	13.35				
VY0505SBS01	212383	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.