



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: Q1958

Project: Joint Base McGuire Dix Lakehurst JBMDL - DoD

RunNo.: LB135748

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: TOC	ICV1	mg/L	1030	1000	103	90-110	03/14/2025
Sample ID: TOC	CCV1	mg/L	1070	1000	107	90-110	05/13/2025
Sample ID: TOC	CCV2	mg/L	943	1000	94	90-110	05/13/2025
Sample ID: TOC	CCV3	mg/L	950	1000	95	90-110	05/13/2025