

Matrix Spike Summary

Client:	Weston Solutions	SDG No.:	Q1958
Project:	Joint Base McGuire Dix Lakehurst JBMDL - DoD	Sample ID:	Q1958-03
Client ID:	SS050P-SD30-050225-00MS	Percent Solids for Spike Sample:	79.1

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
TOC	mg/Kg	75-125	4880		3840		1000	1	104		05/13/2025

Matrix Spike Summary

Client:	Weston Solutions	SDG No.:	Q1958
Project:	Joint Base McGuire Dix Lakehurst JBMDL - DoD	Sample ID:	Q1958-03
Client ID:	SS050P-SD30-050225-00MSD	Percent Solids for Spike Sample:	79.1

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
TOC	mg/Kg	75-125	5010		3840		1000	1	117		05/13/2025