



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Parkway Generation Operating LLC **SDG No.:** Q1959
Contract: PSEG04 **Lab Code:** CHEM **Case No.:** Q1959 **SAS No.:** Q1959
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.20	0.2	100	40 - 160	CV	05/07/2025	09:31	LB135690
CRI01	Arsenic	19.7	20.0	98	40 - 160	P	05/07/2025	13:16	LB135699
	Barium	88.3	100	88	40 - 160	P	05/07/2025	13:16	LB135699
	Cadmium	6.09	6.0	102	40 - 160	P	05/07/2025	13:16	LB135699
	Chromium	9.49	10.0	95	40 - 160	P	05/07/2025	13:16	LB135699
	Lead	13.4	12.0	112	40 - 160	P	05/07/2025	13:16	LB135699
	Selenium	16.5	20.0	82	40 - 160	P	05/07/2025	13:16	LB135699
	Silver	10.2	10.0	102	40 - 160	P	05/07/2025	13:16	LB135699