

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Parkway Generation Operating LLC **SDG No.:** Q1959
Contract: PSEG04 **Lab Code:** CHEM **Case No.:** Q1959 **SAS No.:** Q1959
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	-1.60			-20	20	05/07/2025	13:21	LB135699
	Barium	-1.29	6.0	22	-94	106	05/07/2025	13:21	LB135699
	Cadmium	-2.63	1.0	263	-5	7	05/07/2025	13:21	LB135699
	Chromium	54.1	52.0	104	42	62	05/07/2025	13:21	LB135699
	Lead	-2.88			-12	12	05/07/2025	13:21	LB135699
	Selenium	-9.82			-20	20	05/07/2025	13:21	LB135699
	Silver	1.08			-10	10	05/07/2025	13:21	LB135699
ICSAB01	Arsenic	117	104	112	88.4	120	05/07/2025	13:25	LB135699
	Barium	502	537	94	437	637	05/07/2025	13:25	LB135699
	Cadmium	1070	972	110	826	1120	05/07/2025	13:25	LB135699
	Chromium	586	542	108	460	624	05/07/2025	13:25	LB135699
	Lead	55.5	49.0	113	37	61	05/07/2025	13:25	LB135699
	Selenium	41.2	46.0	90	26	66	05/07/2025	13:25	LB135699
	Silver	197	201	98	170	232	05/07/2025	13:25	LB135699
ICSA	Arsenic	-42.7			-20	20	05/07/2025	13:30	LB135699
	Barium	-89.7	6.0	1495	-94	106	05/07/2025	13:30	LB135699
	Cadmium	-4.12	1.0	412	-5	7	05/07/2025	13:30	LB135699
	Chromium	52.2	52.0	100	42	62	05/07/2025	13:30	LB135699
	Lead	20.5			-12	12	05/07/2025	13:30	LB135699
	Selenium	24.3			-20	20	05/07/2025	13:30	LB135699
	Silver	13.5			-10	10	05/07/2025	13:30	LB135699
ICSAB	Arsenic	159	104	153	88.4	120	05/07/2025	13:34	LB135699
	Barium	386	537	72	437	637	05/07/2025	13:34	LB135699
	Cadmium	1070	972	110	826	1120	05/07/2025	13:34	LB135699
	Chromium	589	542	109	460	624	05/07/2025	13:34	LB135699
	Lead	57.1	49.0	116	37	61	05/07/2025	13:34	LB135699
	Selenium	-24.2	46.0	53	26	66	05/07/2025	13:34	LB135699
	Silver	193	201	96	170	232	05/07/2025	13:34	LB135699