

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Parkway Generation Operating LLC

SDG No.: Q1959

Contract: PSEG04

Lab Code: CHEM

Case No.: Q1959

SAS No.: Q1959

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167882BS							
Arsenic	ug/L	4000	3680		92	80 - 120	P
Barium	ug/L	1000	828		83	80 - 120	P
Cadmium	ug/L	1000	940		94	80 - 120	P
Chromium	ug/L	2000	1900		95	80 - 120	P
Lead	ug/L	5000	4630		93	80 - 120	P
Selenium	ug/L	10000	9090		91	80 - 120	P
Silver	ug/L	380	349		92	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Parkway Generation Operating LLC

SDG No.: Q1959

Contract: PSEG04

Lab Code: CHEM

Case No.: Q1959

SAS No.: Q1959

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167883BS Mercury	ug/L	4.0	4.14		104	80 - 120	CV