



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Zimmer Biomet **SDG No.:** Q1967
Contract: ZIMM01 **Lab Code:** CHEM **Case No.:** Q1967 **SAS No.:** Q1967
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Zinc	42.5	40.0	106	40 - 160	P	05/12/2025	12:18	LB135744
CRI01	Zinc	42.6	40.0	107	65 - 135	P	05/13/2025	13:00	LB135756