

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Zimmer Biomet **SDG No.:** Q1967
Contract: ZIMM01 **Lab Code:** CHEM **Case No.:** Q1967 **SAS No.:** Q1967
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Zinc	4.74			-40	40	05/12/2025	12:22	LB135744
ICSAB01	Zinc	1090	952	114	809	1095	05/12/2025	12:26	LB135744
ICSA	Zinc	7.15			-40	40	05/12/2025	12:32	LB135744
ICSAB	Zinc	1100	952	116	809	1095	05/12/2025	12:36	LB135744
ICSA01	Zinc	3.75			-40	40	05/13/2025	13:04	LB135756
ICSAB01	Zinc	978	952	103	809	1095	05/13/2025	13:08	LB135756
ICSA	Zinc	6.18			-40	40	05/13/2025	13:12	LB135756
ICSAB	Zinc	1110	952	117	809	1095	05/13/2025	13:17	LB135756