



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: CDM Smith SDG No.: Q1982
Contract: CAMP02 Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.20	0.2	100	40 - 160	CV	05/12/2025	10:11	LB135733
CRI01	Aluminum	112	100	112	65 - 135	P	05/13/2025	13:00	LB135756
	Antimony	53.0	50.0	106	65 - 135	P	05/13/2025	13:00	LB135756
	Arsenic	22.6	20.0	113	65 - 135	P	05/13/2025	13:00	LB135756
	Barium	91.9	100	92	65 - 135	P	05/13/2025	13:00	LB135756
	Beryllium	6.28	6.0	105	65 - 135	P	05/13/2025	13:00	LB135756
	Cadmium	6.41	6.0	107	65 - 135	P	05/13/2025	13:00	LB135756
	Calcium	1990	2000	100	65 - 135	P	05/13/2025	13:00	LB135756
	Chromium	9.49	10.0	95	65 - 135	P	05/13/2025	13:00	LB135756
	Cobalt	29.7	30.0	99	65 - 135	P	05/13/2025	13:00	LB135756
	Copper	22.3	20.0	112	65 - 135	P	05/13/2025	13:00	LB135756
	Iron	96.7	100	97	65 - 135	P	05/13/2025	13:00	LB135756
	Lead	11.7	12.0	97	65 - 135	P	05/13/2025	13:00	LB135756
	Magnesium	2110	2000	106	65 - 135	P	05/13/2025	13:00	LB135756
	Manganese	20.2	20.0	101	65 - 135	P	05/13/2025	13:00	LB135756
	Nickel	40.2	40.0	100	65 - 135	P	05/13/2025	13:00	LB135756
	Potassium	1830	2000	91	65 - 135	P	05/13/2025	13:00	LB135756
	Selenium	24.0	20.0	120	65 - 135	P	05/13/2025	13:00	LB135756
	Silver	10.1	10.0	101	65 - 135	P	05/13/2025	13:00	LB135756
	Sodium	1700	2000	85	65 - 135	P	05/13/2025	13:00	LB135756
	Thallium	38.6	40.0	96	65 - 135	P	05/13/2025	13:00	LB135756
	Vanadium	38.1	40.0	95	65 - 135	P	05/13/2025	13:00	LB135756
	Zinc	42.6	40.0	107	65 - 135	P	05/13/2025	13:00	LB135756