



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	CDM Smith				SDG No.:	Q1982				
Contract:	CAMP02	Lab Code:	CHEM	Case No.:	Q1982	SAS No.:	Q1982			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB05	Mercury	0.20	+/-0.20	U	0.20	CV	05/12/2025	09:58	LB135733	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1982							
Contract: CAMP02		Lab Code: CHEM		Case No.: Q1982		SAS No.: Q1982			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB07	Mercury	0.20	+/-0.20	U	0.20	CV	05/12/2025	10:08	LB135733
CCB08	Mercury	0.20	+/-0.20	U	0.20	CV	05/12/2025	10:39	LB135733
CCB09	Mercury	0.20	+/-0.20	U	0.20	CV	05/12/2025	11:11	LB135733
CCB10	Mercury	0.20	+/-0.20	U	0.20	CV	05/12/2025	11:39	LB135733
CCB11	Mercury	0.20	+/-0.20	U	0.20	CV	05/12/2025	12:53	LB135733
CCB12	Mercury	0.20	+/-0.20	U	0.20	CV	05/12/2025	13:27	LB135733

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1982							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1982	SAS No.: Q1982						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	05/13/2025	12:55	LB135756
	Antimony	50.0	+/-50.0	U	50.0	P	05/13/2025	12:55	LB135756
	Arsenic	20.0	+/-20.0	U	20.0	P	05/13/2025	12:55	LB135756
	Barium	100	+/-100	U	100	P	05/13/2025	12:55	LB135756
	Beryllium	6.00	+/-6.00	U	6.00	P	05/13/2025	12:55	LB135756
	Cadmium	6.00	+/-6.00	U	6.00	P	05/13/2025	12:55	LB135756
	Calcium	2000	+/-2000	U	2000	P	05/13/2025	12:55	LB135756
	Chromium	10.0	+/-10.0	U	10.0	P	05/13/2025	12:55	LB135756
	Cobalt	30.0	+/-30.0	U	30.0	P	05/13/2025	12:55	LB135756
	Copper	20.0	+/-20.0	U	20.0	P	05/13/2025	12:55	LB135756
	Iron	100	+/-100	U	100	P	05/13/2025	12:55	LB135756
	Lead	12.0	+/-12.0	U	12.0	P	05/13/2025	12:55	LB135756
	Magnesium	2000	+/-2000	U	2000	P	05/13/2025	12:55	LB135756
	Manganese	20.0	+/-20.0	U	20.0	P	05/13/2025	12:55	LB135756
	Nickel	40.0	+/-40.0	U	40.0	P	05/13/2025	12:55	LB135756
	Potassium	2000	+/-2000	U	2000	P	05/13/2025	12:55	LB135756
	Selenium	20.0	+/-20.0	U	20.0	P	05/13/2025	12:55	LB135756
	Silver	10.0	+/-10.0	U	10.0	P	05/13/2025	12:55	LB135756
	Sodium	2000	+/-2000	U	2000	P	05/13/2025	12:55	LB135756
	Thallium	40.0	+/-40.0	U	40.0	P	05/13/2025	12:55	LB135756
	Vanadium	40.0	+/-40.0	U	40.0	P	05/13/2025	12:55	LB135756
	Zinc	40.0	+/-40.0	U	40.0	P	05/13/2025	12:55	LB135756

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1982							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1982	SAS No.: Q1982						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	05/13/2025	13:26	LB135756
	Antimony	50.0	+/-50.0	U	50.0	P	05/13/2025	13:26	LB135756
	Arsenic	20.0	+/-20.0	U	20.0	P	05/13/2025	13:26	LB135756
	Barium	100	+/-100	U	100	P	05/13/2025	13:26	LB135756
	Beryllium	6.00	+/-6.00	U	6.00	P	05/13/2025	13:26	LB135756
	Cadmium	6.00	+/-6.00	U	6.00	P	05/13/2025	13:26	LB135756
	Calcium	2000	+/-2000	U	2000	P	05/13/2025	13:26	LB135756
	Chromium	10.0	+/-10.0	U	10.0	P	05/13/2025	13:26	LB135756
	Cobalt	30.0	+/-30.0	U	30.0	P	05/13/2025	13:26	LB135756
	Copper	20.0	+/-20.0	U	20.0	P	05/13/2025	13:26	LB135756
	Iron	100	+/-100	U	100	P	05/13/2025	13:26	LB135756
	Lead	12.0	+/-12.0	U	12.0	P	05/13/2025	13:26	LB135756
	Magnesium	2000	+/-2000	U	2000	P	05/13/2025	13:26	LB135756
	Manganese	20.0	+/-20.0	U	20.0	P	05/13/2025	13:26	LB135756
	Nickel	40.0	+/-40.0	U	40.0	P	05/13/2025	13:26	LB135756
	Potassium	2000	+/-2000	U	2000	P	05/13/2025	13:26	LB135756
	Selenium	20.0	+/-20.0	U	20.0	P	05/13/2025	13:26	LB135756
	Silver	10.0	+/-10.0	U	10.0	P	05/13/2025	13:26	LB135756
	Sodium	2000	+/-2000	U	2000	P	05/13/2025	13:26	LB135756
	Thallium	40.0	+/-40.0	U	40.0	P	05/13/2025	13:26	LB135756
	Vanadium	40.0	+/-40.0	U	40.0	P	05/13/2025	13:26	LB135756
	Zinc	40.0	+/-40.0	U	40.0	P	05/13/2025	13:26	LB135756
CCB02	Aluminum	16.5	+/-100	J	100	P	05/13/2025	14:56	LB135756
	Antimony	50.0	+/-50.0	U	50.0	P	05/13/2025	14:56	LB135756
	Arsenic	20.0	+/-20.0	U	20.0	P	05/13/2025	14:56	LB135756
	Barium	100	+/-100	U	100	P	05/13/2025	14:56	LB135756
	Beryllium	6.00	+/-6.00	U	6.00	P	05/13/2025	14:56	LB135756
	Cadmium	6.00	+/-6.00	U	6.00	P	05/13/2025	14:56	LB135756
	Calcium	2000	+/-2000	U	2000	P	05/13/2025	14:56	LB135756
	Chromium	10.0	+/-10.0	U	10.0	P	05/13/2025	14:56	LB135756
	Cobalt	30.0	+/-30.0	U	30.0	P	05/13/2025	14:56	LB135756
	Copper	20.0	+/-20.0	U	20.0	P	05/13/2025	14:56	LB135756
	Iron	100	+/-100	U	100	P	05/13/2025	14:56	LB135756
	Lead	12.0	+/-12.0	U	12.0	P	05/13/2025	14:56	LB135756
	Magnesium	2000	+/-2000	U	2000	P	05/13/2025	14:56	LB135756
	Manganese	20.0	+/-20.0	U	20.0	P	05/13/2025	14:56	LB135756
	Nickel	40.0	+/-40.0	U	40.0	P	05/13/2025	14:56	LB135756
	Potassium	2000	+/-2000	U	2000	P	05/13/2025	14:56	LB135756
	Selenium	20.0	+/-20.0	U	20.0	P	05/13/2025	14:56	LB135756

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1982							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1982	SAS No.: Q1982						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	05/13/2025	14:56	LB135756
	Sodium	2000	+/-2000	U	2000	P	05/13/2025	14:56	LB135756
	Thallium	40.0	+/-40.0	U	40.0	P	05/13/2025	14:56	LB135756
	Vanadium	40.0	+/-40.0	U	40.0	P	05/13/2025	14:56	LB135756
	Zinc	40.0	+/-40.0	U	40.0	P	05/13/2025	14:56	LB135756
CCB03	Aluminum	100	+/-100	U	100	P	05/13/2025	16:47	LB135756
	Antimony	50.0	+/-50.0	U	50.0	P	05/13/2025	16:47	LB135756
	Arsenic	20.0	+/-20.0	U	20.0	P	05/13/2025	16:47	LB135756
	Barium	100	+/-100	U	100	P	05/13/2025	16:47	LB135756
	Beryllium	6.00	+/-6.00	U	6.00	P	05/13/2025	16:47	LB135756
	Cadmium	6.00	+/-6.00	U	6.00	P	05/13/2025	16:47	LB135756
	Calcium	2000	+/-2000	U	2000	P	05/13/2025	16:47	LB135756
	Chromium	10.0	+/-10.0	U	10.0	P	05/13/2025	16:47	LB135756
	Cobalt	30.0	+/-30.0	U	30.0	P	05/13/2025	16:47	LB135756
	Copper	20.0	+/-20.0	U	20.0	P	05/13/2025	16:47	LB135756
	Iron	100	+/-100	U	100	P	05/13/2025	16:47	LB135756
	Lead	12.0	+/-12.0	U	12.0	P	05/13/2025	16:47	LB135756
	Magnesium	2000	+/-2000	U	2000	P	05/13/2025	16:47	LB135756
	Manganese	20.0	+/-20.0	U	20.0	P	05/13/2025	16:47	LB135756
	Nickel	40.0	+/-40.0	U	40.0	P	05/13/2025	16:47	LB135756
	Potassium	2000	+/-2000	U	2000	P	05/13/2025	16:47	LB135756
	Selenium	20.0	+/-20.0	U	20.0	P	05/13/2025	16:47	LB135756
	Silver	10.0	+/-10.0	U	10.0	P	05/13/2025	16:47	LB135756
	Sodium	2000	+/-2000	U	2000	P	05/13/2025	16:47	LB135756
	Thallium	40.0	+/-40.0	U	40.0	P	05/13/2025	16:47	LB135756
	Vanadium	40.0	+/-40.0	U	40.0	P	05/13/2025	16:47	LB135756
	Zinc	40.0	+/-40.0	U	40.0	P	05/13/2025	16:47	LB135756
CCB04	Aluminum	11.3	+/-100	J	100	P	05/13/2025	17:46	LB135756
	Antimony	50.0	+/-50.0	U	50.0	P	05/13/2025	17:46	LB135756
	Arsenic	20.0	+/-20.0	U	20.0	P	05/13/2025	17:46	LB135756
	Barium	100	+/-100	U	100	P	05/13/2025	17:46	LB135756
	Beryllium	6.00	+/-6.00	U	6.00	P	05/13/2025	17:46	LB135756
	Cadmium	6.00	+/-6.00	U	6.00	P	05/13/2025	17:46	LB135756
	Calcium	2000	+/-2000	U	2000	P	05/13/2025	17:46	LB135756
	Chromium	10.0	+/-10.0	U	10.0	P	05/13/2025	17:46	LB135756
	Cobalt	30.0	+/-30.0	U	30.0	P	05/13/2025	17:46	LB135756
	Copper	20.0	+/-20.0	U	20.0	P	05/13/2025	17:46	LB135756
	Iron	100	+/-100	U	100	P	05/13/2025	17:46	LB135756
	Lead	12.0	+/-12.0	U	12.0	P	05/13/2025	17:46	LB135756

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1982							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1982	SAS No.: Q1982						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	05/13/2025	17:46	LB135756
	Manganese	20.0	+/-20.0	U	20.0	P	05/13/2025	17:46	LB135756
	Nickel	40.0	+/-40.0	U	40.0	P	05/13/2025	17:46	LB135756
	Potassium	2000	+/-2000	U	2000	P	05/13/2025	17:46	LB135756
	Selenium	20.0	+/-20.0	U	20.0	P	05/13/2025	17:46	LB135756
	Silver	10.0	+/-10.0	U	10.0	P	05/13/2025	17:46	LB135756
	Sodium	956	+/-2000	J	2000	P	05/13/2025	17:46	LB135756
	Thallium	40.0	+/-40.0	U	40.0	P	05/13/2025	17:46	LB135756
	Vanadium	40.0	+/-40.0	U	40.0	P	05/13/2025	17:46	LB135756
	Zinc	40.0	+/-40.0	U	40.0	P	05/13/2025	17:46	LB135756
CCB05	Aluminum	14.7	+/-100	J	100	P	05/13/2025	18:39	LB135756
	Antimony	50.0	+/-50.0	U	50.0	P	05/13/2025	18:39	LB135756
	Arsenic	20.0	+/-20.0	U	20.0	P	05/13/2025	18:39	LB135756
	Barium	100	+/-100	U	100	P	05/13/2025	18:39	LB135756
	Beryllium	6.00	+/-6.00	U	6.00	P	05/13/2025	18:39	LB135756
	Cadmium	6.00	+/-6.00	U	6.00	P	05/13/2025	18:39	LB135756
	Calcium	2000	+/-2000	U	2000	P	05/13/2025	18:39	LB135756
	Chromium	10.0	+/-10.0	U	10.0	P	05/13/2025	18:39	LB135756
	Cobalt	30.0	+/-30.0	U	30.0	P	05/13/2025	18:39	LB135756
	Copper	20.0	+/-20.0	U	20.0	P	05/13/2025	18:39	LB135756
	Iron	100	+/-100	U	100	P	05/13/2025	18:39	LB135756
	Lead	12.0	+/-12.0	U	12.0	P	05/13/2025	18:39	LB135756
	Magnesium	2000	+/-2000	U	2000	P	05/13/2025	18:39	LB135756
	Manganese	20.0	+/-20.0	U	20.0	P	05/13/2025	18:39	LB135756
	Nickel	40.0	+/-40.0	U	40.0	P	05/13/2025	18:39	LB135756
	Potassium	2000	+/-2000	U	2000	P	05/13/2025	18:39	LB135756
	Selenium	20.0	+/-20.0	U	20.0	P	05/13/2025	18:39	LB135756
	Silver	10.0	+/-10.0	U	10.0	P	05/13/2025	18:39	LB135756
	Sodium	2000	+/-2000	U	2000	P	05/13/2025	18:39	LB135756
	Thallium	40.0	+/-40.0	U	40.0	P	05/13/2025	18:39	LB135756
	Vanadium	40.0	+/-40.0	U	40.0	P	05/13/2025	18:39	LB135756
	Zinc	40.0	+/-40.0	U	40.0	P	05/13/2025	18:39	LB135756
CCB06	Aluminum	100	+/-100	U	100	P	05/13/2025	19:25	LB135756
	Antimony	50.0	+/-50.0	U	50.0	P	05/13/2025	19:25	LB135756
	Arsenic	20.0	+/-20.0	U	20.0	P	05/13/2025	19:25	LB135756
	Barium	100	+/-100	U	100	P	05/13/2025	19:25	LB135756
	Beryllium	6.00	+/-6.00	U	6.00	P	05/13/2025	19:25	LB135756
	Cadmium	6.00	+/-6.00	U	6.00	P	05/13/2025	19:25	LB135756
	Calcium	2000	+/-2000	U	2000	P	05/13/2025	19:25	LB135756

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1982							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1982	SAS No.: Q1982						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	05/13/2025	19:25	LB135756
	Cobalt	30.0	+/-30.0	U	30.0	P	05/13/2025	19:25	LB135756
	Copper	20.0	+/-20.0	U	20.0	P	05/13/2025	19:25	LB135756
	Iron	100	+/-100	U	100	P	05/13/2025	19:25	LB135756
	Lead	12.0	+/-12.0	U	12.0	P	05/13/2025	19:25	LB135756
	Magnesium	2000	+/-2000	U	2000	P	05/13/2025	19:25	LB135756
	Manganese	20.0	+/-20.0	U	20.0	P	05/13/2025	19:25	LB135756
	Nickel	40.0	+/-40.0	U	40.0	P	05/13/2025	19:25	LB135756
	Potassium	2000	+/-2000	U	2000	P	05/13/2025	19:25	LB135756
	Selenium	20.0	+/-20.0	U	20.0	P	05/13/2025	19:25	LB135756
	Silver	10.0	+/-10.0	U	10.0	P	05/13/2025	19:25	LB135756
	Sodium	2000	+/-2000	U	2000	P	05/13/2025	19:25	LB135756
	Thallium	40.0	+/-40.0	U	40.0	P	05/13/2025	19:25	LB135756
	Vanadium	40.0	+/-40.0	U	40.0	P	05/13/2025	19:25	LB135756
	Zinc	40.0	+/-40.0	U	40.0	P	05/13/2025	19:25	LB135756
CCB07	Aluminum	100	+/-100	U	100	P	05/13/2025	20:10	LB135756
	Antimony	50.0	+/-50.0	U	50.0	P	05/13/2025	20:10	LB135756
	Arsenic	20.0	+/-20.0	U	20.0	P	05/13/2025	20:10	LB135756
	Barium	100	+/-100	U	100	P	05/13/2025	20:10	LB135756
	Beryllium	6.00	+/-6.00	U	6.00	P	05/13/2025	20:10	LB135756
	Cadmium	6.00	+/-6.00	U	6.00	P	05/13/2025	20:10	LB135756
	Calcium	2000	+/-2000	U	2000	P	05/13/2025	20:10	LB135756
	Chromium	10.0	+/-10.0	U	10.0	P	05/13/2025	20:10	LB135756
	Cobalt	30.0	+/-30.0	U	30.0	P	05/13/2025	20:10	LB135756
	Copper	20.0	+/-20.0	U	20.0	P	05/13/2025	20:10	LB135756
	Iron	100	+/-100	U	100	P	05/13/2025	20:10	LB135756
	Lead	12.0	+/-12.0	U	12.0	P	05/13/2025	20:10	LB135756
	Magnesium	2000	+/-2000	U	2000	P	05/13/2025	20:10	LB135756
	Manganese	20.0	+/-20.0	U	20.0	P	05/13/2025	20:10	LB135756
	Nickel	40.0	+/-40.0	U	40.0	P	05/13/2025	20:10	LB135756
	Potassium	2000	+/-2000	U	2000	P	05/13/2025	20:10	LB135756
	Selenium	20.0	+/-20.0	U	20.0	P	05/13/2025	20:10	LB135756
	Silver	10.0	+/-10.0	U	10.0	P	05/13/2025	20:10	LB135756
	Sodium	2000	+/-2000	U	2000	P	05/13/2025	20:10	LB135756
	Thallium	40.0	+/-40.0	U	40.0	P	05/13/2025	20:10	LB135756
	Vanadium	40.0	+/-40.0	U	40.0	P	05/13/2025	20:10	LB135756
	Zinc	40.0	+/-40.0	U	40.0	P	05/13/2025	20:10	LB135756
CCB08	Aluminum	100	+/-100	U	100	P	05/13/2025	20:56	LB135756
	Antimony	50.0	+/-50.0	U	50.0	P	05/13/2025	20:56	LB135756

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1982							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1982	SAS No.: Q1982						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	05/13/2025	20:56	LB135756
	Barium	100	+/-100	U	100	P	05/13/2025	20:56	LB135756
	Beryllium	6.00	+/-6.00	U	6.00	P	05/13/2025	20:56	LB135756
	Cadmium	6.00	+/-6.00	U	6.00	P	05/13/2025	20:56	LB135756
	Calcium	2000	+/-2000	U	2000	P	05/13/2025	20:56	LB135756
	Chromium	10.0	+/-10.0	U	10.0	P	05/13/2025	20:56	LB135756
	Cobalt	30.0	+/-30.0	U	30.0	P	05/13/2025	20:56	LB135756
	Copper	20.0	+/-20.0	U	20.0	P	05/13/2025	20:56	LB135756
	Iron	100	+/-100	U	100	P	05/13/2025	20:56	LB135756
	Lead	12.0	+/-12.0	U	12.0	P	05/13/2025	20:56	LB135756
	Magnesium	2000	+/-2000	U	2000	P	05/13/2025	20:56	LB135756
	Manganese	20.0	+/-20.0	U	20.0	P	05/13/2025	20:56	LB135756
	Nickel	40.0	+/-40.0	U	40.0	P	05/13/2025	20:56	LB135756
	Potassium	2000	+/-2000	U	2000	P	05/13/2025	20:56	LB135756
	Selenium	20.0	+/-20.0	U	20.0	P	05/13/2025	20:56	LB135756
	Silver	10.0	+/-10.0	U	10.0	P	05/13/2025	20:56	LB135756
	Sodium	2000	+/-2000	U	2000	P	05/13/2025	20:56	LB135756
	Thallium	40.0	+/-40.0	U	40.0	P	05/13/2025	20:56	LB135756
	Vanadium	40.0	+/-40.0	U	40.0	P	05/13/2025	20:56	LB135756
	Zinc	40.0	+/-40.0	U	40.0	P	05/13/2025	20:56	LB135756
CCB09	Aluminum	13.1	+/-100	J	100	P	05/13/2025	21:42	LB135756
	Antimony	50.0	+/-50.0	U	50.0	P	05/13/2025	21:42	LB135756
	Arsenic	20.0	+/-20.0	U	20.0	P	05/13/2025	21:42	LB135756
	Barium	100	+/-100	U	100	P	05/13/2025	21:42	LB135756
	Beryllium	6.00	+/-6.00	U	6.00	P	05/13/2025	21:42	LB135756
	Cadmium	6.00	+/-6.00	U	6.00	P	05/13/2025	21:42	LB135756
	Calcium	2000	+/-2000	U	2000	P	05/13/2025	21:42	LB135756
	Chromium	10.0	+/-10.0	U	10.0	P	05/13/2025	21:42	LB135756
	Cobalt	30.0	+/-30.0	U	30.0	P	05/13/2025	21:42	LB135756
	Copper	20.0	+/-20.0	U	20.0	P	05/13/2025	21:42	LB135756
	Iron	100	+/-100	U	100	P	05/13/2025	21:42	LB135756
	Lead	12.0	+/-12.0	U	12.0	P	05/13/2025	21:42	LB135756
	Magnesium	2000	+/-2000	U	2000	P	05/13/2025	21:42	LB135756
	Manganese	20.0	+/-20.0	U	20.0	P	05/13/2025	21:42	LB135756
	Nickel	40.0	+/-40.0	U	40.0	P	05/13/2025	21:42	LB135756
	Potassium	2000	+/-2000	U	2000	P	05/13/2025	21:42	LB135756
	Selenium	20.0	+/-20.0	U	20.0	P	05/13/2025	21:42	LB135756
	Silver	10.0	+/-10.0	U	10.0	P	05/13/2025	21:42	LB135756
	Sodium	2000	+/-2000	U	2000	P	05/13/2025	21:42	LB135756

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1982							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1982	SAS No.: Q1982						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	40.0	+/-40.0	U	40.0	P	05/13/2025	21:42	LB135756
	Vanadium	40.0	+/-40.0	U	40.0	P	05/13/2025	21:42	LB135756
	Zinc	40.0	+/-40.0	U	40.0	P	05/13/2025	21:42	LB135756
CCB10	Aluminum	100	+/-100	U	100	P	05/13/2025	22:29	LB135756
	Antimony	50.0	+/-50.0	U	50.0	P	05/13/2025	22:29	LB135756
	Arsenic	20.0	+/-20.0	U	20.0	P	05/13/2025	22:29	LB135756
	Barium	100	+/-100	U	100	P	05/13/2025	22:29	LB135756
	Beryllium	6.00	+/-6.00	U	6.00	P	05/13/2025	22:29	LB135756
	Cadmium	6.00	+/-6.00	U	6.00	P	05/13/2025	22:29	LB135756
	Calcium	2000	+/-2000	U	2000	P	05/13/2025	22:29	LB135756
	Chromium	10.0	+/-10.0	U	10.0	P	05/13/2025	22:29	LB135756
	Cobalt	30.0	+/-30.0	U	30.0	P	05/13/2025	22:29	LB135756
	Copper	20.0	+/-20.0	U	20.0	P	05/13/2025	22:29	LB135756
	Iron	100	+/-100	U	100	P	05/13/2025	22:29	LB135756
	Lead	12.0	+/-12.0	U	12.0	P	05/13/2025	22:29	LB135756
	Magnesium	2000	+/-2000	U	2000	P	05/13/2025	22:29	LB135756
	Manganese	20.0	+/-20.0	U	20.0	P	05/13/2025	22:29	LB135756
	Nickel	40.0	+/-40.0	U	40.0	P	05/13/2025	22:29	LB135756
	Potassium	2000	+/-2000	U	2000	P	05/13/2025	22:29	LB135756
	Selenium	20.0	+/-20.0	U	20.0	P	05/13/2025	22:29	LB135756
	Silver	10.0	+/-10.0	U	10.0	P	05/13/2025	22:29	LB135756
	Sodium	2000	+/-2000	U	2000	P	05/13/2025	22:29	LB135756
	Thallium	40.0	+/-40.0	U	40.0	P	05/13/2025	22:29	LB135756
	Vanadium	40.0	+/-40.0	U	40.0	P	05/13/2025	22:29	LB135756
	Zinc	40.0	+/-40.0	U	40.0	P	05/13/2025	22:29	LB135756