

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith **SDG No.:** Q1982
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1982 **SAS No.:** Q1982

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167930BS							
Aluminum	mg/Kg	93.5	85.9		92	80 - 120	P
Antimony	mg/Kg	37.4	36.7		98	80 - 120	P
Arsenic	mg/Kg	37.4	35.6		95	80 - 120	P
Barium	mg/Kg	9.3	8.32		90	80 - 120	P
Beryllium	mg/Kg	9.3	8.73		94	80 - 120	P
Cadmium	mg/Kg	9.3	8.94		96	80 - 120	P
Calcium	mg/Kg	46.7	44.2	J	95	80 - 120	P
Chromium	mg/Kg	18.7	18.0		96	80 - 120	P
Cobalt	mg/Kg	9.3	8.68		93	80 - 120	P
Copper	mg/Kg	14.0	13.7		98	80 - 120	P
Iron	mg/Kg	140	133		95	80 - 120	P
Lead	mg/Kg	46.7	43.8		94	80 - 120	P
Magnesium	mg/Kg	93.5	85.3	J	91	80 - 120	P
Manganese	mg/Kg	9.3	8.76		94	80 - 120	P
Nickel	mg/Kg	23.4	21.9		94	80 - 120	P
Potassium	mg/Kg	470	434		92	80 - 120	P
Selenium	mg/Kg	93.5	89.6		96	80 - 120	P
Silver	mg/Kg	3.5	3.31		95	80 - 120	P
Sodium	mg/Kg	140	137		98	80 - 120	P
Thallium	mg/Kg	93.5	94.0		100	80 - 120	P
Vanadium	mg/Kg	14.0	12.8		91	80 - 120	P
Zinc	mg/Kg	9.3	9.06		97	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith **SDG No.:** Q1982
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1982 **SAS No.:** Q1982

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167937BS Mercury	mg/Kg	0.26	0.26		99	80 - 120	CV